

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051322\
 Data File : BM035080.D
 Acq On : 15 May 2022 00:31
 Operator : CG/JU
 Sample : PB144662BL
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK662

Quant Time: May 15 02:32:29 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM051222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 12 18:15:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.917	152	3785	0.400	ng/ul	0.00
4) Naphthalene-d8	10.716	136	12052	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.552	164	8348	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.292	188	18542	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.474	240	16849	0.400	ng/ul	0.00
23) Perylene-d12	23.859	264	16555	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	22079	4.870	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.310	152	5838	0.304	ng/ul	0.00
18) Fluoranthene-d10	19.322	212	15458	0.268	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051322\
Data File : BM035080.D
Acq On : 15 May 2022 00:31
Operator : CG/JU
Sample : PB144662BL
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK662

Quant Time: May 15 02:32:29 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM051222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 12 18:15:39 2022
Response via : Initial Calibration

