

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051324\
 Data File : BM045724.D
 Acq On : 13 May 2024 22:35
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM051324

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/14/2024
 Supervised By :mohammad ahmed 05/17/2024

Quant Time: May 14 02:44:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM051324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 14 02:40:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.581	152	97459	20.000	ng	0.00
21) Naphthalene-d8	10.351	136	443718	20.000	ng	0.00
39) Acenaphthene-d10	14.210	164	322592	20.000	ng	0.00
64) Phenanthrene-d10	16.957	188	772238	20.000	ng	0.00
76) Chrysene-d12	21.174	240	510522	20.000	ng	0.00
86) Perylene-d12	23.974	264	486066	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.269	112	466884	82.211	ng	0.00
7) Phenol-d6	6.857	99	577413	83.378	ng	-0.01
23) Nitrobenzene-d5	8.745	82	769249	79.623	ng	0.00
42) 2,4,6-Tribromophenol	15.715	330	363023	82.417	ng	0.00
45) 2-Fluorobiphenyl	12.839	172	1784729	76.253	ng	0.00
79) Terphenyl-d14	19.598	244	3723940	71.187	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.210	88	69993	45.213	ng	98
3) Pyridine	3.640	79	257822m	42.163	ng	
4) n-Nitrosodimethylamine	3.522	42	137700	42.758	ng	95
6) Aniline	6.963	93	241005	42.202	ng	99
8) 2-Chlorophenol	7.187	128	216802	41.351	ng	99
9) Benzaldehyde	6.757	77	192793	38.900	ng	98
10) Phenol	6.887	94	267276	41.406	ng	99
11) bis(2-Chloroethyl)ether	7.040	93	230735	40.140	ng	95
12) 1,3-Dichlorobenzene	7.469	146	280952	39.536	ng	98
13) 1,4-Dichlorobenzene	7.616	146	289441	39.590	ng	97
14) 1,2-Dichlorobenzene	7.934	146	259438	39.766	ng	98
15) Benzyl Alcohol	7.875	79	257049	40.226	ng	98
16) 2,2'-oxybis(1-Chloropr...	8.128	45	227890	40.208	ng	90
17) 2-Methylphenol	8.098	107	190610	41.321	ng	98
18) Hexachloroethane	8.645	117	125697	39.953	ng	99
19) n-Nitroso-di-n-propyla...	8.422	70	216545	41.627	ng	98
20) 3+4-Methylphenols	8.428	107	269948	41.413	ng	# 88
22) Acetophenone	8.428	105	400052	40.019	ng	98
24) Nitrobenzene	8.792	77	366057	39.156	ng	100
25) Isophorone	9.339	82	562996	40.308	ng	99
26) 2-Nitrophenol	9.492	139	134435	42.350	ng	96
27) 2,4-Dimethylphenol	9.604	122	163925	41.106	ng	95
28) bis(2-Chloroethoxy)met...	9.798	93	322304	39.877	ng	99
29) 2,4-Dichlorophenol	10.045	162	238147	40.582	ng	99
30) 1,2,4-Trichlorobenzene	10.210	180	280214	39.460	ng	99
31) Naphthalene	10.398	128	870242	39.641	ng	100
32) Benzoic acid	9.851	122	166732m	42.561	ng	
33) 4-Chloroaniline	10.569	127	320923	40.725	ng	99
34) Hexachlorobutadiene	10.686	225	179181	39.588	ng	97
35) Caprolactam	11.480	113	76830	40.233	ng	98
36) 4-Chloro-3-methylphenol	11.727	107	301254	41.819	ng	98
37) 2-Methylnaphthalene	12.016	142	640332	39.882	ng	99
38) 1-Methylnaphthalene	12.233	142	628459	39.752	ng	100
40) 1,2,4,5-Tetrachloroben...	12.380	216	322594	37.675	ng	98
41) Hexachlorocyclopentadiene	12.369	237	137694	36.677	ng	98
43) 2,4,6-Trichlorophenol	12.663	196	216606	38.566	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051324\
 Data File : BM045724.D
 Acq On : 13 May 2024 22:35
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM051324

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/14/2024
 Supervised By :mohammad ahmed 05/17/2024

Quant Time: May 14 02:44:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM051324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 14 02:40:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.751	196	247510	39.248	ng	99
46) 1,1'-Biphenyl	13.045	154	878832	37.927	ng	99
47) 2-Chloronaphthalene	13.080	162	683017	38.107	ng	99
48) 2-Nitroaniline	13.333	65	235757	40.068	ng	98
49) Acenaphthylene	13.933	152	1108189	39.775	ng	100
50) Dimethylphthalate	13.716	163	911358	38.763	ng	100
51) 2,6-Dinitrotoluene	13.816	165	202398	39.805	ng	98
52) Acenaphthene	14.274	154	678865	39.290	ng	98
53) 3-Nitroaniline	14.174	138	203276	40.673	ng	100
54) 2,4-Dinitrophenol	14.351	184	111418	37.965	ng	98
55) Dibenzofuran	14.616	168	1137559	39.693	ng	99
56) 4-Nitrophenol	14.557	139	181695	42.243	ng	98
57) 2,4-Dinitrotoluene	14.610	165	315456	40.951	ng	97
58) Fluorene	15.263	166	1020943	39.634	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.868	232	237007	40.939	ng	99
60) Diethylphthalate	15.080	149	1064824	39.835	ng	100
61) 4-Chlorophenyl-phenyle...	15.268	204	493586	39.513	ng	99
62) 4-Nitroaniline	15.351	138	222675	40.901	ng	97
63) Azobenzene	15.563	77	1031460	39.334	ng	98
65) 4,6-Dinitro-2-methylph...	15.363	198	167817	37.913	ng	98
66) n-Nitrosodiphenylamine	15.492	169	787218	39.288	ng	99
67) 4-Bromophenyl-phenylether	16.157	248	294211	38.467	ng	96
68) Hexachlorobenzene	16.257	284	353182	38.413	ng	98
69) Atrazine	16.474	200	251771	39.878	ng	97
70) Pentachlorophenol	16.633	266	238284	40.913	ng	96
71) Phenanthrene	16.998	178	1532972	39.369	ng	100
72) Anthracene	17.092	178	1488779	39.690	ng	99
73) Carbazole	17.386	167	1463501	40.227	ng	99
74) Di-n-butylphthalate	17.951	149	1884368	39.843	ng	100
75) Fluoranthene	19.015	202	1933561	40.807	ng	100
77) Benzidine	19.239	184	477187	42.866	ng	100
78) Pyrene	19.380	202	1892066	36.331	ng	99
80) Butylbenzylphthalate	20.303	149	798061	34.977	ng	99
81) Benzo(a)anthracene	21.156	228	1379474	39.484	ng	100
82) 3,3'-Dichlorobenzidine	21.103	252	565834	43.464	ng	100
83) Chrysene	21.215	228	1263744	39.443	ng	99
84) Bis(2-ethylhexyl)phtha...	21.103	149	1114340	37.897	ng	100
85) Di-n-octyl phthalate	22.168	149	1372361	38.972	ng	100
87) Indeno(1,2,3-cd)pyrene	27.097	276	1117519	36.536	ng	# 100
88) Benzo(b)fluoranthene	23.097	252	1291005	41.091	ng	99
89) Benzo(k)fluoranthene	23.162	252	1159412	40.286	ng	99
90) Benzo(a)pyrene	23.844	252	1068766	39.217	ng	# 99
91) Dibenzo(a,h)anthracene	27.144	278	917113	35.587	ng	98
92) Benzo(g,h,i)perylene	28.062	276	1043708	39.376	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051324\
Data File : BM045724.D
Acq On : 13 May 2024 22:35
Operator : MA/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
ICVBM051324

Quant Time: May 14 02:44:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM051324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 14 02:40:40 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.581 min Scan# 70
Delta R.T. -0.000 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Instrument :
BNA_M
ClientSampleId :
ICVBM051324

Tgt Ion:152 Resp: 97459
Ion Ratio Lower Upper
152 100
150 154.7 123.2 184.8
115 64.4 52.1 78.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#2
1,4-Dioxane
Concen: 45.213 ng
RT: 3.210 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35


Tgt Ion: 88 Resp: 69993
Ion Ratio Lower Upper
88 100
58 66.1 54.2 81.2
43 34.5 28.8 43.2





#3
Pyridine
Concen: 42.163 ng m
RT: 3.640 min Scan# 94
Delta R.T. -0.012 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324




Tgt Ion: 79 Resp: 25782
Ion Ratio Lower Upper
79 100
52 56.3 46.5 69.7
51 34.9 29.6 44.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024




#4
n-Nitrosodimethylamine
Concen: 42.758 ng
RT: 3.522 min Scan# 74
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Tgt Ion: 42 Resp: 137700
Ion Ratio Lower Upper
42 100
74 104.2 79.4 119.2
44 4.6 3.6 5.4





#5
2-Fluorophenol
Concen: 82.211 ng
RT: 5.269 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324



Tgt Ion:112 Resp: 466884
Ion Ratio Lower Upper
112 100
64 58.3 45.8 68.8
63 36.9 29.9 44.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#6
Aniline
Concen: 42.202 ng
RT: 6.963 min Scan# 659
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Tgt Ion: 93 Resp: 241005
Ion Ratio Lower Upper
93 100
66 40.4 32.7 49.1
65 22.2 17.5 26.3






#7
Phenol-d6
Concen: 83.378 ng
RT: 6.857 min Scan# 641
Delta R.T. -0.012 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :

ICVBM051324






Tgt Ion: 99 Resp: 57741
Ion Ratio Lower Upper
99 100
42 22.6 18.2 27.2
71 48.8 39.1 58.7

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#8
2-Chlorophenol
Concen: 41.351 ng
RT: 7.187 min Scan# 697
Delta R.T. -0.005 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Tgt Ion:128 Resp: 216802
Ion Ratio Lower Upper
128 100
130 32.1 12.6 52.6
64 46.0 25.8 65.8









#9
Benzaldehyde
Concen: 38.900 ng
RT: 6.757 min Scan# 61
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Instrument :
BNA_M
ClientSampleId :
ICVBM051324

Tgt Ion: 77 Resp: 19279
Ion Ratio Lower Upper
77 100
105 88.6 71.1 111.1
106 87.0 65.3 105.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#10
Phenol
Concen: 41.406 ng
RT: 6.887 min Scan# 646
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Tgt Ion: 94 Resp: 267276
Ion Ratio Lower Upper
94 100
65 36.4 16.5 56.5
66 54.4 35.7 75.7






#11
bis(2-Chloroethyl)ether
Concen: 40.140 ng
RT: 7.040 min Scan# 61
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :

ICVBM051324






Tgt Ion: 93 Resp: 230735
Ion Ratio Lower Upper
93 100
63 75.1 59.7 99.7
95 31.4 13.6 53.6

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#12
1,3-Dichlorobenzene
Concen: 39.536 ng
RT: 7.469 min Scan# 745
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Tgt Ion:146 Resp: 280952
Ion Ratio Lower Upper
146 100
148 63.1 51.8 77.6
75 42.5 32.6 48.8





#13

1,4-Dichlorobenzene

Concen: 39.590 ng

RT: 7.616 min Scan# 770

Delta R.T. 0.000 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :

ICVBM051324





Tgt Ion:146 Resp: 28944

Ion	Ratio	Lower	Upper
146	100		
148	66.7	51.7	77.5
111	45.1	34.9	52.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#14

1,2-Dichlorobenzene

Concen: 39.766 ng

RT: 7.934 min Scan# 824




Delta R.T. 0.000 min


Lab File: BM045724.D

Acq: 13 May 2024 22:35

Tgt Ion:146 Resp: 259438


Ion	Ratio	Lower	Upper
146	100		
148	62.5	51.4	77.2
111	46.1	36.6	55.0





#15
Benzyl Alcohol
Concen: 40.226 ng
RT: 7.875 min Scan# 814
Delta R.T. -0.012 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324



Tgt Ion: 79 Resp: 257049
Ion Ratio Lower Upper
79 100
108 64.3 52.8 79.2
77 70.0 56.8 85.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.208 ng
RT: 8.128 min Scan# 857
Delta R.T. 0.000 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35


Tgt Ion: 45 Resp: 227890
Ion Ratio Lower Upper
45 100
77 23.1 8.5 48.5
79 17.7 2.4 42.4





#17
2-Methylphenol
Concen: 41.321 ng
RT: 8.098 min Scan# 852
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324



Tgt Ion:	107	Resp:	190610
Ion Ratio	Lower	Upper	
107	100		
108	116.8	95.3	142.9
77	68.3	55.7	83.5
79	63.1	51.7	77.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#18
Hexachloroethane
Concen: 39.953 ng
RT: 8.645 min Scan# 945
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35


Tgt Ion:	117	Resp:	125697
Ion Ratio	Lower	Upper	
117	100		
119	96.3	77.3	115.9
201	83.8	69.0	103.4





#19
n-Nitroso-di-n-propylamine
Concen: 41.627 ng
RT: 8.422 min Scan# 907
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324



Tgt Ion: 70 Resp: 21654
Ion Ratio Lower Upper
70 100
42 56.4 46.8 70.2
101 9.6 7.4 11.0
130 19.4 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#20
3+4-Methylphenols
Concen: 41.413 ng
RT: 8.428 min Scan# 908
Delta R.T. -0.012 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Tgt Ion:107 Resp: 269948
Ion Ratio Lower Upper
107 100
108 84.5 70.3 110.3
77 181.4 140.1 180.1#
79 35.0 19.5 59.5






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.351 min Scan# 1035
Delta R.T. 0.000 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :

ICVBM051324






Tgt Ion:136 Resp: 443718
Ion Ratio Lower Upper
136 100
137 10.7 8.8 13.2
54 6.9 5.8 8.6
68 4.6 4.0 6.0

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#22
Acetophenone
Concen: 40.019 ng
RT: 8.428 min Scan# 908
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Tgt Ion:105 Resp: 400052
Ion Ratio Lower Upper
105 100
71 7.2 6.4 9.6
51 39.4 30.5 45.7
120 21.9 16.8 25.2





#23
Nitrobenzene-d5
Concen: 79.623 ng
RT: 8.745 min Scan# 962
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :





ICVBM051324

Tgt Ion: 82 Resp: 769249
Ion Ratio Lower Upper
82 100
128 35.7 29.4 44.2
54 47.9 38.8 58.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#24
Nitrobenzene
Concen: 39.156 ng
RT: 8.792 min Scan# 970
Delta R.T. 0.000 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35


Tgt Ion: 77 Resp: 366057
Ion Ratio Lower Upper
77 100
123 39.9 31.8 47.6
65 14.6 11.8 17.6





#25
Isophorone
Concen: 40.308 ng
RT: 9.339 min Scan# 1063
Delta R.T. -0.012 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324



Tgt Ion: 82 Resp: 562990
Ion Ratio Lower Upper
82 100
95 8.8 6.9 10.3
138 16.5 13.0 19.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#26
2-Nitrophenol
Concen: 42.350 ng
RT: 9.492 min Scan# 1089
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35


Tgt Ion:139 Resp: 134435
Ion Ratio Lower Upper
139 100
109 48.3 37.8 56.8
65 67.2 49.9 74.9





#27
2,4-Dimethylphenol
Concen: 41.106 ng
RT: 9.604 min Scan# 1108
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324



Tgt Ion:122 Resp: 163925
Ion Ratio Lower Upper
122 100
107 127.2 106.1 159.1
121 54.7 47.3 70.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#28
bis(2-Chloroethoxy)methane
Concen: 39.877 ng
RT: 9.798 min Scan# 1141
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Tgt Ion: 93 Resp: 322304
Ion Ratio Lower Upper
93 100
95 32.9 26.2 39.2
123 11.7 9.8 14.6






#29
2,4-Dichlorophenol
Concen: 40.582 ng
RT: 10.045 min Scan# 1183
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :

ICVBM051324






Tgt Ion:162 Resp: 23814
Ion Ratio Lower Upper
162 100
164 64.9 45.5 85.5
98 38.7 18.2 58.2


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024




#30
1,2,4-Trichlorobenzene
Concen: 39.460 ng
RT: 10.210 min Scan# 1211
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Tgt Ion:180 Resp: 280214
Ion Ratio Lower Upper
180 100
182 95.7 77.6 116.4
145 33.7 26.0 39.0





#31
Naphthalene
Concen: 39.641 ng
RT: 10.398 min Scan# 1100
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324




Tgt Ion:128 Resp: 87024
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 12.9 10.5 15.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024




#32
Benzoic acid
Concen: 42.561 ng m
RT: 9.851 min Scan# 1150
Delta R.T. -0.018 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Tgt Ion:122 Resp: 166732
Ion Ratio Lower Upper
122 100
105 122.8 111.1 151.1
77 107.7 92.6 132.6





#33
4-Chloroaniline
Concen: 40.725 ng
RT: 10.569 min Scan# 1272
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324




Tgt Ion	Ratio	Lower	Upper
127	100		
129	32.2	25.0	37.6
65	34.4	27.4	41.0
92	19.6	15.8	23.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024




#34
Hexachlorobutadiene
Concen: 39.588 ng
RT: 10.686 min Scan# 1292
Delta R.T. 0.000 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Tgt Ion	Ratio	Lower	Upper
225	100		
223	63.1	50.0	75.0
227	66.6	50.6	75.8





#35
Caprolactam
Concen: 40.233 ng
RT: 11.480 min Scan# 1431
Delta R.T. -0.024 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324




Tgt Ion:113 Resp: 76830
Ion Ratio Lower Upper
113 100
55 145.5 125.2 165.2
56 117.5 101.8 141.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024




#36
4-Chloro-3-methylphenol
Concen: 41.819 ng
RT: 11.727 min Scan# 1469
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Tgt Ion:107 Resp: 301254
Ion Ratio Lower Upper
107 100
144 24.7 19.6 29.4
142 77.0 60.2 90.4





#37
2-Methylnaphthalene
Concen: 39.882 ng
RT: 12.016 min Scan# 1518
Delta R.T. 0.000 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324




Tgt Ion:142 Resp: 640332
Ion Ratio Lower Upper
142 100
141 88.8 71.0 106.6
115 45.5 35.0 52.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024




#38
1-Methylnaphthalene
Concen: 39.752 ng
RT: 12.233 min Scan# 1555
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Tgt Ion:142 Resp: 628459
Ion Ratio Lower Upper
142 100
141 92.0 73.5 110.3
116 4.7 3.9 5.9





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.210 min Scan# 1892
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324




Tgt Ion:164 Resp: 322592
Ion Ratio Lower Upper
164 100
162 103.5 81.8 122.8
160 46.2 36.4 54.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024




#40
1,2,4,5-Tetrachlorobenzene
Concen: 37.675 ng
RT: 12.380 min Scan# 1580
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35



Tgt Ion:216 Resp: 322594
Ion Ratio Lower Upper
216 100
214 77.3 63.4 95.0
179 23.2 18.6 27.8
108 23.7 19.8 29.8





#41

Hexachlorocyclopentadiene

Concen: 36.677 ng

RT: 12.369 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :

ICVBM051324

Tgt Ion:237 Resp: 137694

Ion Ratio Lower Upper

237 100





235 63.8 45.2 85.2

272 12.7 0.0 31.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#42

2,4,6-Tribromophenol

Concen: 82.417 ng

RT: 15.715 min Scan# 2147

Delta R.T. -0.006 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35




Tgt Ion:330 Resp: 363023


Ion Ratio Lower Upper

330 100

332 96.6 77.8 116.6


141 46.8 37.2 55.8





#43
2,4,6-Trichlorophenol
Concen: 38.566 ng
RT: 12.663 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324




Tgt Ion:196 Resp: 21660
Ion Ratio Lower Upper
196 100
198 97.6 78.2 117.4
200 31.7 25.1 37.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024




#44
2,4,5-Trichlorophenol
Concen: 39.248 ng
RT: 12.751 min Scan# 1643
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35



Tgt Ion:196 Resp: 247510
Ion Ratio Lower Upper
196 100
198 97.3 77.8 116.8
97 65.9 54.5 81.7
132 30.7 24.6 36.8





#45

2-Fluorobiphenyl

Concen: 76.253 ng

RT: 12.839 min Scan# 1658

Delta R.T. 0.000 min

Lab File: BM045724.D


Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :

ICVBM051324



Tgt Ion:172 Resp: 1784729

Ion Ratio Lower Upper

172 100

171 34.7 27.8 41.8




170 23.2 18.9 28.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/14/2024

Supervised By :mohammad ahmed 05/17/2024



#46

1,1'-Biphenyl

Concen: 37.927 ng

RT: 13.045 min Scan# 1693

Delta R.T. -0.006 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35




Tgt Ion:154 Resp: 878832


Ion Ratio Lower Upper

154 100

153 39.9 20.7 60.7


76 11.3 0.0 31.7





#47
2-Chloronaphthalene
Concen: 38.107 ng
RT: 13.080 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324



Tgt Ion:162 Resp: 68301
Ion Ratio Lower Upper
162 100
127 42.4 34.5 51.7
164 32.8 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#48
2-Nitroaniline
Concen: 40.068 ng
RT: 13.333 min Scan# 1742
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Tgt Ion: 65 Resp: 235757
Ion Ratio Lower Upper
65 100
92 66.3 51.5 77.3
138 92.6 75.3 112.9





#49

Acenaphthylene

Concen: 39.775 ng

RT: 13.933 min Scan# 1845

Delta R.T. -0.006 min

Lab File: BM045724.D


Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :

ICVBM051324



Tgt Ion:152 Resp: 1108189

Ion Ratio Lower Upper

152 100

151 20.0 16.1 24.1




153 13.0 10.4 15.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/14/2024

Supervised By :mohammad ahmed 05/17/2024



#50

Dimethylphthalate

Concen: 38.763 ng

RT: 13.716 min Scan# 1807

Delta R.T. -0.006 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35




Tgt Ion:163 Resp: 911358


Ion Ratio Lower Upper

163 100

194 4.0 3.1 4.7


164 10.1 8.1 12.1





#51
2,6-Dinitrotoluene
Concen: 39.805 ng
RT: 13.816 min Scan# 1825
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324



Tgt Ion:165 Resp: 202398
Ion Ratio Lower Upper
165 100
63 81.8 63.1 94.7
89 80.9 65.2 97.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#52
Acenaphthene
Concen: 39.290 ng
RT: 14.274 min Scan# 1902
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35


Tgt Ion:154 Resp: 678865
Ion Ratio Lower Upper
154 100
153 111.8 91.5 137.3
152 53.9 42.8 64.2





#53
3-Nitroaniline
Concen: 40.673 ng
RT: 14.174 min Scan# 1885
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324



Tgt Ion:138 Resp: 203270
Ion Ratio Lower Upper
138 100
108 10.0 8.4 12.6
92 124.4 99.4 149.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#54
2,4-Dinitrophenol
Concen: 37.965 ng
RT: 14.351 min Scan# 1915
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Tgt Ion:184 Resp: 111418
Ion Ratio Lower Upper
184 100
63 90.8 74.3 111.5
154 78.5 63.8 95.6





#55

Dibenzenofuran

Concen: 39.693 ng

RT: 14.616 min Scan# 1960

Delta R.T. -0.006 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :

ICVBM051324

Tgt Ion:168 Resp: 113755

Ion Ratio Lower Upper

168 100

139 44.6 35.1 52.7





169 13.0 10.8 16.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/14/2024

Supervised By :mohammad ahmed 05/17/2024



#56

4-Nitrophenol

Concen: 42.243 ng

RT: 14.557 min Scan# 1950

Delta R.T. -0.006 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35




Tgt Ion:139 Resp: 181695


Ion Ratio Lower Upper

139 100

109 113.8 91.8 131.8

65 119.6 100.9 140.9





#57
2,4-Dinitrotoluene
Concen: 40.951 ng
RT: 14.610 min Scan# 1959
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :





ICVBM051324

Tgt Ion:165 Resp: 315450
Ion Ratio Lower Upper
165 100
63 71.8 59.3 88.9
89 94.0 78.1 117.1
182 2.4 2.2 3.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#58
Fluorene
Concen: 39.634 ng
RT: 15.263 min Scan# 2070
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35


Tgt Ion:166 Resp: 1020943
Ion Ratio Lower Upper
166 100
165 99.0 79.0 118.4
167 13.7 10.9 16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 40.939 ng
RT: 14.868 min Scan# 2003
Delta R.T. 0.000 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324



Tgt Ion:232 Resp: 237001
Ion Ratio Lower Upper
232 100
131 59.4 47.0 70.4
130 3.6 2.7 4.1
166 34.5 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#60
Diethylphthalate
Concen: 39.835 ng
RT: 15.080 min Scan# 2039
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35


Tgt Ion:149 Resp: 1064824
Ion Ratio Lower Upper
149 100
177 20.0 16.1 24.1
150 12.5 9.8 14.6





#61
4-Chlorophenyl-phenylether
Concen: 39.513 ng
RT: 15.268 min Scan# 2071
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324



Tgt Ion:204 Resp: 493580
Ion Ratio Lower Upper
204 100
206 33.1 27.0 40.6
141 70.1 55.1 82.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#62
4-Nitroaniline
Concen: 40.901 ng
RT: 15.351 min Scan# 2085
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35


Tgt Ion:138 Resp: 222675
Ion Ratio Lower Upper
138 100
92 59.0 38.8 78.8
108 125.1 99.7 139.7





#63
Azobenzene
Concen: 39.334 ng
RT: 15.563 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324



Tgt Ion: 77 Resp: 1031460
Ion Ratio Lower Upper
77 100
182 21.9 0.6 40.6
105 17.0 0.0 36.2
51 31.5 10.8 50.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.957 min Scan# 2358
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35


Tgt Ion:188 Resp: 772238
Ion Ratio Lower Upper
188 100
94 7.4 5.8 8.6
80 9.1 6.9 10.3





#65
4,6-Dinitro-2-methylphenol
Concen: 37.913 ng
RT: 15.363 min Scan# 2109
Delta R.T. -0.005 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324



Tgt Ion:198 Resp: 16781
Ion Ratio Lower Upper
198 100
51 55.5 33.8 73.8
105 50.9 29.9 69.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#66
n-Nitrosodiphenylamine
Concen: 39.288 ng
RT: 15.492 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Tgt Ion:169 Resp: 787218
Ion Ratio Lower Upper
169 100
168 67.2 54.6 81.8
167 37.1 30.2 45.4






#67
4-Bromophenyl-phenylether
Concen: 38.467 ng
RT: 16.157 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :

ICVBM051324






Tgt Ion:248 Resp: 29421
Ion Ratio Lower Upper
248 100
250 100.2 78.5 117.7
141 93.2 70.8 106.2


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024




#68
Hexachlorobenzene
Concen: 38.413 ng
RT: 16.257 min Scan# 2239
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35



Tgt Ion:284 Resp: 353182
Ion Ratio Lower Upper
284 100
142 43.3 32.9 49.3
249 31.3 25.2 37.8





#69

Atrazine

Concen: 39.878 ng

RT: 16.474 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :

ICVBM051324

Tgt Ion:200 Resp: 25177

Ion Ratio Lower Upper

200 100





173 25.3 4.0 44.0

215 46.9 24.9 64.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#70

Pentachlorophenol

Concen: 40.913 ng

RT: 16.633 min Scan# 2303

Delta R.T. -0.006 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35




Tgt Ion:266 Resp: 238284


Ion Ratio Lower Upper

266 100

268 64.9 50.3 75.5


264 63.6 47.8 71.6





#71
Phenanthrene
Concen: 39.369 ng
RT: 16.998 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324



Tgt Ion:178 Resp: 1532972
Ion Ratio Lower Upper
178 100
176 19.3 15.3 22.9
179 15.2 12.0 18.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#72
Anthracene
Concen: 39.690 ng
RT: 17.092 min Scan# 2381
Delta R.T. 0.000 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35


Tgt Ion:178 Resp: 1488779
Ion Ratio Lower Upper
178 100
176 18.9 14.7 22.1
179 15.2 12.2 18.2





#73
 Carbazole
 Concen: 40.227 ng
 RT: 17.386 min Scan# 2432
 Delta R.T. -0.006 min
 Lab File: BM045724.D
 Acq: 13 May 2024 22:35




Instrument :
 BNA_M
 ClientSampleId :
 ICVBM051324




Tgt Ion:167 Resp: 146350
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.0 25.4
 139 14.1 11.0 16.4

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 05/14/2024
 Supervised By :mohammad ahmed 05/17/2024




#74
 Di-n-butylphthalate
 Concen: 39.843 ng
 RT: 17.951 min Scan# 2527
 Delta R.T. 0.000 min
 Lab File: BM045724.D
 Acq: 13 May 2024 22:35



Tgt Ion:149 Resp: 1884368
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.3 10.9
 104 6.4 5.2 7.8






#75
Fluoranthene
Concen: 40.807 ng
RT: 19.015 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :

ICVBM051324






Tgt Ion:202 Resp: 193356
Ion Ratio Lower Upper
202 100
101 8.8 0.0 28.7
203 17.4 0.0 37.4


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.174 min Scan# 3075
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35



Tgt Ion:240 Resp: 510522
Ion Ratio Lower Upper
240 100
120 7.8 5.8 8.6
236 25.6 20.6 31.0





#77
Benzidine
Concen: 42.866 ng
RT: 19.239 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :





ICVBM051324

Tgt Ion:184 Resp: 47718
Ion Ratio Lower Upper
184 100
185 14.0 11.0 16.4
183 11.6 9.3 13.9

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#78
Pyrene
Concen: 36.331 ng
RT: 19.380 min Scan# 2770
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Tgt Ion:202 Resp: 1892066
Ion Ratio Lower Upper
202 100
200 20.9 16.5 24.7
203 17.4 13.6 20.4





#79

Terphenyl-d14

Concen: 71.187 ng

RT: 19.598 min Scan# 2807

Delta R.T. -0.005 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :

ICVBM051324

Tgt Ion:244 Resp: 3723940

Ion Ratio Lower Upper

244 100





212 7.5 6.1 9.1

122 9.5 7.4 11.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#80

Butylbenzylphthalate

Concen: 34.977 ng

RT: 20.303 min Scan# 2927

Delta R.T. -0.006 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35




Tgt Ion:149 Resp: 798061


Ion Ratio Lower Upper

149 100

91 81.1 65.8 98.6


206 18.0 13.8 20.8





#81
Benzo(a)anthracene
Concen: 39.484 ng
RT: 21.156 min Scan# 3072
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324



Tgt Ion:228 Resp: 1379474
Ion Ratio Lower Upper
228 100
226 26.8 21.4 32.0
229 19.3 15.6 23.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#82
3,3'-Dichlorobenzidine
Concen: 43.464 ng
RT: 21.103 min Scan# 3063
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35

Tgt Ion:252 Resp: 565834
Ion Ratio Lower Upper
252 100
254 64.2 51.5 77.3
126 11.4 9.1 13.7





#83

Chrysene

Concen: 39.443 ng

RT: 21.215 min Scan# 3083

Delta R.T. -0.006 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :

ICVBM051324

Tgt Ion:228 Resp: 126374

Ion Ratio Lower Upper

228 100

226 29.5 23.9 35.9





229 20.2 15.5 23.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/14/2024

Supervised By :mohammad ahmed 05/17/2024



#84

Bis(2-ethylhexyl)phthalate

Concen: 37.897 ng

RT: 21.103 min Scan# 3063

Delta R.T. -0.006 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35




Tgt Ion:149 Resp: 1114340


Ion Ratio Lower Upper

149 100

167 25.3 20.4 30.6


279 3.9 3.2 4.8





#85
Di-n-octyl phthalate
Concen: 38.972 ng
RT: 22.168 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324




Tgt Ion:149 Resp: 137236
Ion Ratio Lower Upper
149 100
167 1.5 1.1 1.7
43 9.8 7.8 11.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024




#86
Perylene-d12
Concen: 20.000 ng
RT: 23.974 min Scan# 3551
Delta R.T. -0.012 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35



Tgt Ion:264 Resp: 486066
Ion Ratio Lower Upper
264 100
260 23.3 18.9 28.3
265 22.7 18.2 27.4





#87

Indeno(1,2,3-cd)pyrene

Concen: 36.536 ng

RT: 27.097 min Scan# 4083

Delta R.T. -0.006 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :

ICVBM051324

Tgt Ion:276 Resp: 1117519

Ion Ratio Lower Upper

276 100





138 14.8 0.0 0.0

277 25.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#88

Benzo(b)fluoranthene

Concen: 41.091 ng

RT: 23.097 min Scan# 3402

Delta R.T. -0.012 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35




Tgt Ion:252 Resp: 1291005


Ion Ratio Lower Upper

252 100

253 22.1 17.6 26.4


125 7.6 5.2 7.8





#89
Benzo(k)fluoranthene
Concen: 40.286 ng
RT: 23.162 min Scan# 3414
Delta R.T. -0.006 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35




Instrument :
BNA_M
ClientSampleId :
ICVBM051324




Tgt Ion:252 Resp: 115941
Ion Ratio Lower Upper
252 100
253 21.9 17.7 26.5
125 7.0 5.2 7.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024




#90
Benzo(a)pyrene
Concen: 39.217 ng
RT: 23.844 min Scan# 3529
Delta R.T. -0.012 min
Lab File: BM045724.D
Acq: 13 May 2024 22:35



Tgt Ion:252 Resp: 1068766
Ion Ratio Lower Upper
252 100
253 22.2 17.4 26.2
125 7.7 5.1 7.7#





#91

Dibenzo(a,h)anthracene

Concen: 35.587 ng

RT: 27.144 min Scan# 4092

Delta R.T. -0.012 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35

Instrument :

BNA_M

ClientSampleId :





ICVBM051324

Tgt Ion:	278	Resp:	91711
Ion Ratio	Lower	Upper	
278	100		
139	9.6	7.9	11.9
279	24.3	18.7	28.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/14/2024
Supervised By :mohammad ahmed 05/17/2024



#92

Benzo(g,h,i)perylene

Concen: 39.376 ng

RT: 28.062 min Scan# 4246

Delta R.T. -0.018 min

Lab File: BM045724.D

Acq: 13 May 2024 22:35

Tgt Ion:	276	Resp:	1043708
Ion Ratio	Lower	Upper	
276	100		
277	24.1	19.1	28.7
138	14.5	10.8	16.2

