

Data Path : Z:\HPCHEM1\BNA_M\Data\BM051415\
 Data File : BM001276.D
 Acq On : 13 May 2015 15:28
 Operator : TP/IZ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.471

Quant Time: May 14 01:27:49 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM051115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 12 03:03:15 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
3	Naphthalene	1.012	1.035	-2.3	128	0.00
4 SURR	2-Methylnaphthalene-d10	0.508	0.524	-3.1	132	0.00
5	2-Methylnaphthalene	0.664	0.684	-3.0	133	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	133	0.00
7	Acenaphthylene	1.766	1.780	-0.8	135	0.00
8 C	Acenaphthene	1.392	1.374	1.3	131	0.00
9	Fluorene	1.552	1.531	1.4	131	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	127	0.00
11	Pentachlorophenol	0.087	0.066	24.1#	106	0.00
12	Phenanthrene	1.137	1.127	0.9	127	0.00
13	Anthracene	1.043	1.048	-0.5	129	0.00
14 SURR	Fluoranthene-d10	0.873	0.826	5.4	121	0.00
15 C	Fluoranthene	1.262	1.219	3.4	122	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	120	0.00
17	Pyrene	1.557	1.610	-3.4	123	0.00
18	Benzo(a)anthracene	1.222	1.277	-4.5	127	0.00
19	Chrysene	1.393	1.432	-2.8	119	0.00
20 I	Perylene-d12	1.000	1.000	0.0	121	0.00
21	Benzo(b)fluoranthene	1.442	1.406	2.5	118	0.00
22	Benzo(k)fluoranthene	1.485	1.548	-4.2	124	0.00
23 C	Benzo(a)pyrene	1.381	1.397	-1.2	125	0.00
24	Indeno(1,2,3-cd)pyrene	1.522	1.600	-5.1	125	0.00
25	Dibenzo(a,h)anthracene	1.213	1.279	-5.4	125	0.00
26	Benzo(g,h,i)perylene	1.358	1.413	-4.1	123	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0