


Data Path : Z:\HPCHEM1\BNA_M\Data\BM051415\
 Data File : BM001275.D
 Acq On : 13 May 2015 08:49
 Operator : TP/IZ
 Sample : DFTPP57
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP57

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM051115.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu May 14 01:20:45 2015



AutoFind: Scans 2489, 2490, 2491; Background Corrected with Scan 2482

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	48.9	239957	PASS
68	69	0.00	2	1.6	3450	PASS
69	198	0.00	100	43.2	211797	PASS
70	69	0.00	2	0.6	1171	PASS
127	198	10	80	51.3	251946	PASS
197	198	0.00	2	0.5	2241	PASS
198	198	100	100	100.0	490816	PASS
199	198	5	9	6.7	32818	PASS
275	198	10	60	23.7	116538	PASS
365	198	1	100	2.7	13289	PASS
441	443	0.01	100	73.7	58616	PASS
442	198	50	100	81.6	400725	PASS
443	442	15	24	19.9	79568	PASS

m/z	Abundance
36.10	445.0
37.10	1158.0
38.00	2621.0
39.10	17960.0
40.00	1601.0
41.10	374.0
44.00	3466.0
49.00	1668.0
50.10	56864.0
51.10	221632.0
52.10	11211.0
52.90	364.0
53.20	333.0
55.00	1204.0
56.00	6243.0
57.00	14380.0
58.00	671.0
59.10	385.0
61.00	3292.0
62.00	3075.0
63.00	8391.0
64.00	1087.0
65.00	3931.0
66.10	326.0
68.00	3681.0
69.00	193216.0
70.00	1312.0
73.00	1795.0
74.00	19936.0
75.00	31232.0
76.00	10876.0
77.10	217600.0
78.10	15274.0
79.00	13952.0
80.00	11286.0
81.00	16359.0
82.10	4054.0
83.00	3764.0
83.90	520.0
85.00	3249.0
86.00	4551.0
87.00	2184.0
88.00	823.0
88.90	338.0
91.00	3574.0
92.00	4481.0
93.00	25168.0
94.20	1853.0
96.00	778.0
97.10	561.0
98.00	20048.0
99.00	16808.0
100.10	1332.0
101.00	9205.0
101.90	404.0
103.00	2374.0
104.00	5134.0
105.00	5134.0
106.10	1627.0
107.00	67344.0
108.00	9616.0
109.00	2240.0
110.00	112528.0
111.00	17472.0
112.00	2041.0
113.10	784.0
116.00	3508.0
117.00	55808.0
118.00	3895.0
119.00	377.0
120.10	483.0
121.00	533.0
122.00	4053.0
123.00	5914.0
124.00	3277.0
125.00	2638.0
126.10	1034.0
127.00	217600.0
128.00	16400.0
129.00	85768.0
130.00	7851.0
131.00	1376.0
132.00	1004.0
133.00	341.0
134.00	2424.0
135.00	6355.0
136.00	2579.0
137.00	4021.0
137.80	946.0
139.10	558.0
140.00	998.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

141.00	10748.0
142.00	3474.0
143.00	2535.0
143.90	545.0
144.90	415.0
146.00	1694.0
147.00	5424.0
148.00	11302.0
149.00	3360.0
149.90	615.0
150.80	1255.0
151.40	1592.0
152.10	924.0
153.10	2968.0
154.00	2401.0
155.00	5671.0
156.00	8368.0
157.00	2386.0
157.90	2008.0
158.90	1451.0
159.90	3575.0
161.00	4806.0
162.00	1727.0
163.10	520.0
164.20	550.0
165.00	3502.0
166.00	3038.0
167.00	21776.0
168.00	12335.0
169.00	1661.0
170.10	588.0
171.10	1130.0
172.00	1886.0
173.00	3161.0
174.10	4114.0
175.10	8981.0
175.90	2090.0
176.90	3539.0
178.00	1369.0
179.00	14881.0
180.00	12238.0
181.00	5588.0
181.90	919.0
184.00	1016.0
185.00	6702.0
186.00	55608.0
187.00	14268.0
188.00	2440.0
188.90	3174.0
190.00	665.0
191.00	1964.0
192.00	5459.0
193.00	5209.0
194.00	1312.0
195.10	758.0
196.00	10443.0
196.80	3245.0
198.00	386240.0
199.00	25648.0
200.00	2114.0
201.60	2328.0
203.00	2811.0
204.00	13715.0
205.00	24544.0
206.00	102208.0
207.00	14192.0
208.00	3482.0
208.90	1449.0
210.30	2132.0
211.10	3501.0
214.80	981.0
216.10	2331.0
217.00	25944.0
218.00	3441.0
219.00	435.0
219.90	554.0
221.00	13999.0
222.00	3296.0
223.00	5843.0
224.00	54112.0
225.00	14543.0
226.10	2343.0
227.00	22024.0
228.10	3665.0
229.00	4343.0
230.10	839.0
231.10	1779.0
232.10	457.0
232.80	342.0
234.10	1652.0
235.00	1725.0
236.00	1023.0
236.90	2436.0
237.80	516.0
239.10	716.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

240.00	790.0
241.00	1122.0
242.00	2256.0
243.10	2533.0
244.10	38808.0
245.00	4824.0
246.00	8331.0
247.00	1833.0
248.10	532.0
248.90	1684.0
252.00	705.0
253.00	611.0
254.00	1417.0
255.00	198592.0
256.00	30280.0
257.10	3395.0
258.00	11227.0
258.90	1821.0
263.90	325.0
265.00	4591.0
266.10	817.0
268.00	312.0
271.00	402.0
272.00	621.0
273.00	5551.0
274.00	15233.0
275.00	86048.0
276.00	10969.0
277.00	7009.0
278.00	1170.0
279.00	324.0
283.10	914.0
284.10	556.0
284.90	1155.0
285.20	1221.0
288.90	401.0
290.00	316.0
293.00	1827.0
293.90	656.0
295.00	592.0
296.00	19872.0
297.00	3175.0
301.00	395.0
302.10	452.0
303.00	2908.0
304.00	1060.0
310.10	375.0
313.90	850.0
315.00	2772.0
316.00	1551.0
320.90	623.0
322.00	383.0
323.00	6545.0
324.10	1294.0
326.90	1589.0
327.90	741.0
331.80	339.0
333.00	673.0
334.00	4696.0
335.00	1319.0
340.80	689.0
341.80	333.0
346.00	1396.0
352.00	2104.0
353.10	1517.0
354.00	2211.0
359.00	331.0
365.00	9886.0
366.00	1186.0
370.00	305.0
371.00	551.0
371.90	2862.0
372.90	1029.0
383.00	1050.0
390.00	600.0
402.10	1212.0
403.00	1705.0
404.00	920.0
421.00	2065.0
422.00	1525.0
423.00	12779.0
424.00	5669.0
425.00	1013.0
441.00	32648.0
442.00	233152.0
443.00	50944.0
444.10	4465.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

m/z	Abundance
36.00	526.0
37.00	1590.0
38.10	3095.0
39.10	22344.0
40.10	2134.0
41.10	647.0
44.00	3483.0
45.00	696.0
49.00	1904.0
50.00	76872.0
51.10	291648.0
52.10	15112.0
53.10	710.0
54.90	1188.0
56.00	8014.0
57.00	17144.0
58.00	885.0
61.10	2942.0
62.00	4100.0
63.10	11198.0
64.00	1516.0
65.00	5381.0
66.00	445.0
68.10	3391.0
69.00	257472.0
70.00	1124.0
73.00	2184.0
74.00	28208.0
75.00	42880.0
76.10	14484.0
77.10	291968.0
78.10	19768.0
79.00	18992.0
80.00	15930.0
81.00	22248.0
82.00	5051.0
83.10	4387.0
84.00	707.0
85.00	4327.0
86.00	5671.0
87.00	2488.0
87.90	945.0
88.90	491.0
91.00	4696.0
92.00	5149.0
93.00	33808.0
94.00	2205.0
95.00	371.0
96.10	1515.0
97.00	722.0
98.00	29664.0
99.00	20976.0
99.90	1418.0
101.00	11193.0
102.00	882.0
103.00	4601.0
104.00	7694.0
105.00	7367.0
106.00	1816.0
107.00	88704.0
108.00	14697.0
109.00	2365.0
110.00	154176.0
111.00	22336.0
112.00	2405.0
113.10	1075.0
114.80	324.0
116.00	5204.0
117.00	76536.0
118.00	5065.0
119.00	750.0
119.90	800.0
120.80	363.0
122.10	6246.0
123.00	8318.0
124.00	3668.0
125.00	4192.0
126.10	1155.0
127.00	302912.0
128.00	24208.0
129.00	118744.0
130.00	10049.0
131.00	1764.0
131.90	1318.0
132.90	538.0
134.00	3232.0
135.00	9895.0
136.00	3889.0
137.10	4830.0
138.00	1015.0
139.10	665.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

139.90	1299.0
141.00	15567.0
142.00	4336.0
143.00	3274.0
144.00	816.0
145.10	992.0
146.00	2138.0
147.00	7633.0
148.00	17848.0
149.10	3871.0
150.00	914.0
151.10	1962.0
152.00	1365.0
152.90	5179.0
154.00	3537.0
155.00	7325.0
156.10	11445.0
156.90	2576.0
158.00	2890.0
159.00	1694.0
160.00	5063.0
161.00	7602.0
162.00	2039.0
163.00	676.0
164.10	1015.0
165.00	7014.0
166.00	4675.0
167.00	31344.0
168.00	15708.0
169.10	3114.0
170.00	941.0
171.00	1591.0
172.00	2995.0
173.00	4097.0
174.10	5662.0
175.00	10970.0
176.10	2787.0
177.00	4734.0
178.10	1549.0
179.00	22232.0
180.00	15386.0
181.00	7907.0
182.10	1388.0
182.80	413.0
184.00	2004.0
185.00	9825.0
186.00	82616.0
187.00	23224.0
188.00	2598.0
189.00	6188.0
190.00	1437.0
191.10	2631.0
192.00	6942.0
193.10	8159.0
194.00	1339.0
195.00	780.0
196.10	17744.0
196.80	3479.0
197.00	3388.0
198.00	593536.0
199.00	40664.0
200.00	3034.0
201.50	2288.0
203.00	3967.0
204.00	21976.0
205.00	36592.0
206.00	155520.0
207.00	21056.0
208.00	5418.0
209.00	1825.0
210.10	2027.0
211.00	5187.0
211.70	1754.0
212.90	460.0
215.00	1654.0
216.00	3857.0
217.00	39448.0
218.00	5270.0
219.00	422.0
221.00	26760.0
223.00	8603.0
224.00	87992.0
225.00	18776.0
225.90	2341.0
227.00	33704.0
228.00	4584.0
229.00	7246.0
230.00	1147.0
231.00	3706.0
232.90	413.0
234.10	2615.0
235.00	2259.0
236.00	1806.0
237.00	4208.0
238.00	577.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

238.90	1386.0
240.00	1143.0
241.00	1926.0
242.00	4204.0
243.10	4728.0
244.10	62880.0
245.00	9041.0
246.00	11282.0
247.00	2402.0
247.90	516.0
249.10	2075.0
249.90	723.0
251.10	548.0
251.90	407.0
253.00	1368.0
254.00	1560.0
255.00	329920.0
256.00	45920.0
257.00	4197.0
258.00	17344.0
259.00	2955.0
260.00	560.0
261.00	663.0
264.00	748.0
265.00	6892.0
265.90	1220.0
270.00	313.0
270.90	619.0
272.00	1045.0
273.00	9724.0
274.00	23024.0
275.00	134656.0
276.00	18648.0
277.00	11883.0
278.00	2078.0
279.00	544.0
281.90	353.0
283.00	1186.0
284.00	668.0
285.00	1862.0
285.90	681.0
289.10	559.0
290.90	404.0
292.10	338.0
293.00	3177.0
294.10	649.0
294.90	590.0
296.00	36592.0
297.00	5213.0
297.90	305.0
301.00	764.0
302.00	726.0
303.00	3912.0
304.00	908.0
308.10	443.0
308.90	393.0
310.00	557.0
314.00	2107.0
315.00	3398.0
316.00	2752.0
317.00	387.0
321.00	878.0
322.10	541.0
323.10	11938.0
324.00	2605.0
326.00	373.0
327.00	1989.0
328.00	893.0
331.90	990.0
333.00	955.0
334.00	7484.0
335.10	2353.0
336.10	406.0
341.00	1272.0
342.10	492.0
346.00	2335.0
346.80	620.0
352.00	3015.0
353.00	2825.0
354.00	3365.0
354.90	617.0
365.00	14025.0
366.10	2188.0
370.00	312.0
371.10	701.0
372.00	5274.0
373.00	1420.0
383.00	1512.0
383.90	640.0
390.00	978.0
391.00	450.0
392.00	422.0
402.00	2335.0
403.00	3977.0
404.00	1348.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

405.00	504.0
421.00	3163.0
422.00	3213.0
423.00	21488.0
424.00	8371.0
424.80	657.0
441.00	67176.0
442.00	448896.0
443.00	90016.0
444.00	8182.0
445.00	448.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

m/z	Abundance
37.00	1101.0
38.10	2357.0
39.10	16140.0
40.00	1806.0
41.10	482.0
44.00	3842.0
45.00	300.0
49.10	1396.0
50.00	50128.0
51.10	206592.0
52.00	11226.0
53.00	379.0
55.10	992.0
56.00	5824.0
57.00	13801.0
58.00	750.0
61.00	2608.0
62.10	2822.0
63.00	8056.0
64.10	1237.0
65.10	3686.0
66.00	317.0
68.10	3280.0
69.00	184704.0
70.00	1079.0
73.00	1890.0
74.00	19984.0
75.00	31392.0
76.00	10871.0
77.10	213184.0
78.10	13526.0
79.00	14717.0
80.00	11180.0
81.00	16544.0
82.00	3519.0
83.00	3168.0
84.00	510.0
85.00	3599.0
86.00	4388.0
87.00	2016.0
88.00	603.0
88.90	373.0
91.00	2478.0
92.10	3752.0
93.00	27400.0
93.90	1833.0
95.00	449.0
96.00	1219.0
97.20	404.0
98.00	19976.0
99.00	15037.0
99.90	1399.0
101.00	8842.0
102.00	542.0
103.00	2893.0
104.00	5155.0
105.00	4982.0
106.00	1953.0
107.00	67168.0
108.00	9543.0
108.90	1802.0
110.00	119648.0
111.00	18672.0
112.10	1944.0
112.90	531.0
114.90	305.0
116.00	3084.0
117.00	58480.0
118.00	4007.0
119.00	448.0
120.10	719.0
121.20	414.0
122.00	4717.0
123.00	7643.0
124.00	2943.0
125.00	3214.0
126.10	1201.0
127.00	235328.0
128.00	16187.0
129.00	90368.0
130.00	7308.0
131.00	1323.0
131.90	727.0
133.00	470.0
134.00	3039.0
135.00	7604.0
136.00	3200.0
137.10	3732.0
138.00	636.0
139.00	493.0
140.00	1165.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

141.00	11925.0
142.00	4334.0
143.00	2613.0
144.10	849.0
145.00	650.0
146.10	1837.0
147.00	5970.0
148.00	14314.0
149.00	2808.0
149.90	1140.0
151.20	870.0
152.00	1173.0
153.00	3603.0
154.00	2809.0
155.00	7252.0
156.00	10268.0
157.10	1957.0
158.00	1970.0
159.00	1511.0
160.00	3599.0
161.00	4780.0
161.90	1748.0
163.00	503.0
164.10	503.0
165.00	5032.0
166.10	3254.0
167.00	24576.0
168.00	12890.0
169.10	2184.0
170.20	655.0
170.90	963.0
171.90	2071.0
173.00	3612.0
174.00	4954.0
175.00	10372.0
176.10	2877.0
176.90	4067.0
178.10	1335.0
179.00	18376.0
180.00	14033.0
181.00	5854.0
182.00	1213.0
182.90	580.0
184.00	1655.0
185.00	8229.0
186.00	67480.0
187.00	18216.0
188.00	2340.0
189.00	4797.0
190.10	563.0
191.00	1950.0
192.00	6059.0
193.00	6668.0
193.90	1001.0
194.90	912.0
196.00	16041.0
197.10	2284.0
198.00	492672.0
199.00	32144.0
200.00	2970.0
200.80	830.0
201.50	2062.0
203.10	3824.0
204.00	19928.0
205.00	30856.0
206.10	129728.0
207.10	17936.0
208.10	4645.0
209.00	1696.0
210.00	1838.0
211.00	4646.0
211.70	1645.0
213.00	441.0
215.00	1362.0
215.90	2133.0
217.00	34840.0
218.00	4459.0
219.10	449.0
221.00	22864.0
223.00	7539.0
224.00	71248.0
225.10	18264.0
226.00	2417.0
227.00	30888.0
228.00	3928.0
229.00	6376.0
229.80	552.0
231.00	2809.0
232.00	589.0
232.90	491.0
234.00	1772.0
235.00	2738.0
236.00	1182.0
237.00	2965.0
237.90	408.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

238.90	1136.0
240.00	1330.0
241.00	2080.0
242.10	3403.0
243.00	4181.0
244.00	56824.0
245.00	8190.0
246.00	12210.0
246.90	2874.0
247.90	487.0
249.00	2062.0
249.90	323.0
250.90	393.0
252.20	826.0
252.90	980.0
254.10	1788.0
255.00	288704.0
256.00	41128.0
257.10	3903.0
258.00	15818.0
259.00	2666.0
260.10	497.0
261.00	468.0
265.00	6982.0
266.00	921.0
270.00	466.0
270.90	504.0
271.90	919.0
273.00	8553.0
274.00	20304.0
275.00	128912.0
276.00	17736.0
277.00	10853.0
278.10	1841.0
281.20	459.0
281.90	382.0
283.00	1100.0
283.90	1178.0
284.90	1554.0
288.90	411.0
289.90	345.0
291.90	336.0
293.00	2124.0
294.00	748.0
295.10	524.0
296.00	35888.0
297.00	4710.0
297.80	323.0
300.80	355.0
302.00	920.0
303.00	3191.0
304.00	1516.0
308.10	809.0
308.80	327.0
310.10	543.0
314.10	1897.0
315.00	3509.0
316.10	1799.0
317.10	728.0
320.90	1343.0
322.00	562.0
323.00	12109.0
324.10	1601.0
327.00	1841.0
327.90	1083.0
332.00	866.0
333.10	1000.0
334.10	7337.0
335.10	1659.0
338.90	464.0
340.90	1341.0
341.90	404.0
345.90	2167.0
351.10	336.0
352.00	3729.0
353.00	2629.0
354.00	3294.0
355.10	769.0
359.00	345.0
365.00	15958.0
365.90	2275.0
370.00	363.0
370.90	818.0
372.10	6621.0
373.20	1150.0
376.90	315.0
383.00	2282.0
384.00	352.0
389.90	778.0
391.00	605.0
400.90	474.0
402.00	2170.0
403.10	3367.0
404.00	1321.0
420.90	3668.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

422.00	3873.0
423.00	25376.0
424.00	6456.0
425.00	903.0
441.00	76024.0
442.00	520128.0
443.00	97744.0
444.00	9039.0
445.10	486.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57

Scan 2482 (17.286 min): BM001275.D
DFTPP57

m/z	Abundance
40.00	1050.0
44.00	3221.0
206.80	470.0

Instrument :
BNA_M
ClientSampleId :
DFTPP57