

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051421\
 Data File : BM030017.D
 Acq On : 14 May 2021 13:33
 Operator : CG/JU
 Sample : SST00.479
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST00.4079

Quant Time: May 14 14:15:56 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 14:15:56 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	1652	0.40	ng/ul	0.00
4) Naphthalene-d8	10.33	136	9101	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.19	164	5312	0.40	ng/ul	0.00
13) Phenanthrene-d10	16.95	188	9516	0.40	ng/ul	0.00
17) Chrysene-d12	21.13	240	8045	0.40	ng/ul	0.00
23) Perylene-d12	23.29	264	7915	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.08	96	1134	0.58	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.93	152	5249	0.38	ng/ul	0.00
18) Fluoranthene-d10	18.97	212	10171	0.43	ng/ul	0.00
Target Compounds				Qvalue		
2) 1,4-Dioxane	3.11	88	1073	0.557	ng/ul	100
5) Naphthalene	10.38	128	10048	0.399	ng/ul	100
7) 2-Methylnaphthalene	12.01	142	6603	0.383	ng/ul	100
8) 1-Methylnaphthalene	12.22	142	6533	0.397	ng/ul	100
10) Acenaphthylene	13.92	152	8557	0.363	ng/ul	100
11) Acenaphthene	14.26	153	7697	0.416	ng/ul	100
12) Fluorene	15.26	166	8031	0.381	ng/ul	100
14) Pentachlorophenol	16.63	266	794	0.304	ng/ul	100
15) Phenanthrene	16.99	178	11517	0.396	ng/ul	100
16) Anthracene	17.09	178	9684	0.360	ng/ul	100
19) Fluoranthene	19.00	202	13026	0.434	ng/ul	100
20) Pyrene	19.37	202	13783	0.449	ng/ul	100
21) Benzo(a)anthracene	21.12	228	9559	0.334	ng/ul	100
22) Chrysene	21.17	228	12765	0.432	ng/ul	100
24) Benzo(b)fluoranthene	22.65	252	11526	0.369	ng/ul	100
25) Benzo(k)fluoranthene	22.69	252	13841	0.450	ng/ul	100
26) Benzo(a)pyrene	23.20	252	11107	0.413	ng/ul	100
27) Indeno(1,2,3-cd)pyrene	25.44	276	14513	0.409	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.45	278	11749	0.396	ng/ul	100
29) Benzo(g,h,i)perylene	26.09	276	13004	0.435	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051421\
Data File : BM030017.D
Acq On : 14 May 2021 13:33
Operator : CG/JU
Sample : SST0.479
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4079

Quant Time: May 14 14:15:56 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM051421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 14 14:15:56 2021
Response via : Initial Calibration

