

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051515\
 Data File : BM001291.D
 Acq On : 14 May 2015 13:38
 Operator : TP/IZ
 Sample : PB83265BL
 Misc : SBLK60
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK60

Quant Time: May 14 17:37:21 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM051515.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 14 05:09:17 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.44	152	215898	20.00	ng/ul	0.00
18) Naphthalene-d8	10.22	136	894242	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.12	164	488626	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.87	188	1001543	20.00	ng/ul	0.00
75) Chrysene-d12	21.07	240	988412	20.00	ng/ul	0.00
83) Perylene-d12	23.19	264	971102	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	2.87	96	25401	5.60	ng/uL	0.00
5) Phenol-d5	6.63	99	543581	33.40	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.78	67	345925	34.24	ng/ul	0.00
9) 2-Chlorophenol-d4	6.97	132	450846	34.13	ng/ul	0.00
13) 4-Methylphenol-d8	8.16	113	465293	34.98	ng/ul	0.00
19) Nitrobenzene-d5	8.60	128	216484	36.78	ng/ul	0.00
22) 2-Nitrophenol-d4	9.32	143	252001	38.36	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.85	165	445900	34.17	ng/ul	0.00
29) 4-Chloroaniline-d4	10.37	131	661518	41.15	ng/ul	0.00
43) Dimethylphthalate-d6	13.54	166	1339506	41.56	ng/ul	0.00
46) Acenaphthylene-d8	13.80	160	1756435	37.19	ng/ul	0.00
51) 4-Nitrophenol-d4	14.34	143	217860	33.62	ng/ul	0.00
57) Fluorene-d10	15.12	176	1205008	37.22	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.25	200	206540	34.53	ng/ul	0.00
70) Anthracene-d10	16.97	188	1786971	38.68	ng/ul	0.00
76) Pyrene-d10	19.27	212	1873170	41.89	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.06	264	1690678	39.73	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051515\
Data File : BM001291.D
Acq On : 14 May 2015 13:38
Operator : TP/IZ
Sample : PB83265BL
Misc : SBLK60
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK60

Quant Time: May 14 17:37:21 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM051515.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 14 05:09:17 2015
Response via : Initial Calibration

