

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM051521\
 Data File : BM030053.D
 Acq On : 15 May 2021 17:52
 Operator : CG/JU
 Sample : SP5563
 Misc : SFAM SURROGATE80/16/0.8
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SP5563

Quant Time: May 16 01:57:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat May 15 09:59:28 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.53	152	125195	20.00	ng/ul	0.00
20) Naphthalene-d8	10.30	136	551944	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.18	164	356766	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.93	188	755800	20.00	ng/ul	0.00
79) Chrysene-d12	21.12	240	773686	20.00	ng/ul	0.00
88) Perylene-d12	23.26	264	822181	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.05	96	49229	15.70	ng/uL	0.00
4) Pyridine-d5	3.45	84	609206	77.62	ng/ul	-0.01
7) Phenol-d5	6.72	99	802128	68.90	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	6.87	67	513278	71.10	ng/ul	0.00
11) 2-Chlorophenol-d4	7.06	132	653420	72.72	ng/ul	0.00
15) 4-Methylphenol-d8	8.25	113	658874	68.43	ng/ul	0.00
21) Nitrobenzene-d5	8.69	128	314465	82.68	ng/ul	0.00
24) 2-Nitrophenol-d4	9.40	143	354912	82.67	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.93	165	585660	68.03	ng/ul	0.00
31) 4-Chloroaniline-d4	10.45	131	902164	68.86	ng/ul	0.00
46) Dimethylphthalate-d6	13.60	166	2011609	72.74	ng/ul	0.00
49) Acenaphthylene-d8	13.87	160	2544346	72.60	ng/ul	0.00
54) 4-Nitrophenol-d4	14.41	143	288007	63.94	ng/ul	0.00
60) Fluorene-d10	15.18	176	1729959	73.69	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.32	200	335003	68.24	ng/ul	0.00
73) Anthracene-d10	17.03	188	2695638	70.83	ng/ul	0.00
81) Pyrene-d10	19.33	212	3112239	78.41	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.13	264	3274261	70.25	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM051521\
Data File : BM030053.D
Acq On : 15 May 2021 17:52
Operator : CG/JU
Sample : SP5563
Misc : SFAM SURROGATE80/16/0.8
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SP5563

Quant Time: May 16 01:57:42 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050721.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat May 15 09:59:28 2021
Response via : Initial Calibration

