

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051615\
 Data File : BM001315.D
 Acq On : 15 May 2015 05:06
 Operator : TP/IZ
 Sample : G2228-02
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 051115-P03-E-U-M

Quant Time: May 15 06:19:57 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 13 05:24:15 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.45	152	14211	5.00	ng	0.00
6) Naphthalene-d8	10.22	136	50012	5.00	ng	0.00
12) Acenaphthene-d10	14.12	164	25896	5.00	ng	0.00
18) Phenanthrene-d10	16.85	188	76202	5.00	ng	0.00
24) Chrysene-d12	21.07	240	46140	5.00	ng	-0.01
30) Perylene-d12	23.18	264	44191	5.00	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.02	112	12722	4.60	ng	-0.02
5) Phenol-d6	6.63	99	9253	2.96	ng	0.00
7) Nitrobenzene-d5	8.60	82	19974	6.30	ng	0.00
13) 2,4,6-Tribromophenol	15.63	330	7645	8.61	ng	0.00
14) 2-Fluorobiphenyl	12.73	172	61264	7.62	ng	0.00
26) Terphenyl-d14	19.51	244	52715	8.50	ng	-0.01


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) n-Nitrosodimethylamine	3.18	42	30	0.02	ng	# 1
8) Nitrobenzene	8.60	77	76	0.15	ng	# 31
9) Naphthalene	10.27	128	86	0.01	ng	# 1
10) Hexachlorobutadiene	10.68	225	5	0.00	ng	# 8
11) 2-Methylnaphthalene	11.90	142	24	0.00	ng	# 68
15) Acenaphthylene	13.83	152	47	0.00	ng	# 62
16) Acenaphthene	14.18	154	22	0.00	ng	# 94
17) Fluorene	14.74	166	391	0.06	ng	# 14
19) Hexachlorobenzene	16.18	284	9	0.00	ng	# 46
20) Pentachlorophenol	16.52	266	31	0.43	ng	# 71
21) Phenanthrene	16.88	178	323	0.03	ng	# 97
22) Anthracene	16.98	178	121	0.01	ng	# 89
23) Fluoranthene	18.93	202	551	0.03	ng	98
25) Pyrene	19.30	202	537	0.04	ng	95
27) Benzo(a)anthracene	21.06	228	442	0.04	ng	# 70
28) Chrysene	21.10	228	305	0.02	ng	# 67
29) Indeno(1,2,3-cd)pyrene	25.26	276	352	0.03	ng	97
31) Benzo(b)fluoranthene	22.56	252	607	0.05	ng	# 9
32) Benzo(k)fluoranthene	22.56	252	607	0.05	ng	# 10
33) Benzo(a)pyrene	23.08	252	261	0.02	ng	# 1
34) Dibenzo(a,h)anthracene	25.28	278	221	0.02	ng	# 1
35) Benzo(g,h,i)perylene	25.89	276	326	0.03	ng	# 48


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051615\
Data File : BM001315.D
Acq On : 15 May 2015 05:06
Operator : TP/IZ
Sample : G2228-02
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
051115-P03-E-U-M

Quant Time: May 15 06:19:57 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM051315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 13 05:24:15 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.45 min Scan# 309

Delta R.T. -0.00 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06

Instrument :

BNA_M

ClientSampleId :

051115-P03-E-U-M





Tgt Ion:152 Resp: 14211

Ion Ratio Lower Upper

152 100

150 132.4 114.6 172.0

115 42.6 39.3 58.9



#3

n-Nitrosodimethylamine

Concen: 0.02 ng

RT: 3.18 min Scan# 33

Delta R.T. -0.06 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06



Tgt Ion: 42 Resp: 30


Ion Ratio Lower Upper

42 100

74 0.0 82.1 123.1#

44 200.0 3.0 4.4#





#4

2-Fluorophenol

Concen: 4.60 ng

RT: 5.02 min Scan# 152

Delta R.T. -0.02 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06

Instrument :

BNA_M

ClientSampleId :

051115-P03-E-U-M


Tgt Ion: 112 Resp: 12722

Ion Ratio Lower Upper

112 100

64 63.6 50.8 76.2




63 35.6 28.2 42.4



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BM0

Ion 63.00 (62.70 to 63.70): BM0



#5

Phenol-d6

Concen: 2.96 ng

RT: 6.63 min Scan# 256

Delta R.T. 0.00 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06


Tgt Ion: 99 Resp: 9253

Ion Ratio Lower Upper

99 100

42 24.1 19.9 29.9



71 34.7 28.0 42.0




Abundance Ion 99.00 (98.70 to 99.70): BM0

Ion 42.00 (41.70 to 42.70): BM0

Ion 71.00 (70.70 to 71.70): BM0









#6
Naphthalene-d8
Concen: 5.00 ng
RT: 10.22 min Scan# 537
Delta R.T. -0.00 min
Lab File: BM001315.D
Acq: 15 May 2015 05:06




Instrument :
BNA_M
ClientSampleId :
051115-P03-E-U-M


Tgt Ion:	136	Resp:	50012
Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.2	12.4
54	8.5	9.5	14.3#
68	5.3	5.5	8.3#



#7
Nitrobenzene-d5
Concen: 6.30 ng
RT: 8.60 min Scan# 397
Delta R.T. 0.00 min
Lab File: BM001315.D
Acq: 15 May 2015 05:06

Tgt Ion:	82	Resp:	19974
Ion	Ratio	Lower	Upper
82	100		
128	40.4	28.0	42.0
54	53.7	57.3	85.9#





#8

Nitrobenzene

Concen: 0.15 ng

RT: 8.60 min Scan# 397

Delta R.T. -0.04 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06

Instrument :

BNA_M

ClientSampleId :

051115-P03-E-U-M


Tgt Ion: 77 Resp: 76

Ion Ratio Lower Upper

77 100

123 0.0 33.4 50.2#


65 43.4 10.6 15.8#





Abundance Ion 77.00 (76.70 to 77.70): BM001262.D (-394) (-)

Ion 123.00 (122.70 to 123.70): BM001262.D (-394) (-)

Ion 65.00 (64.70 to 65.70): BM001262.D (-394) (-)



Time-->



#9

Naphthalene

Concen: 0.01 ng

RT: 10.27 min Scan# 540

Delta R.T. 0.00 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06


Tgt Ion: 128 Resp: 86

Ion Ratio Lower Upper

128 100

129 59.1 9.1 13.7#


127 59.1 10.6 16.0#




Abundance Ion 128.00 (127.70 to 128.70): BM001262.D (-536) (-)


Ion 129.00 (128.70 to 129.70): BM001262.D (-536) (-)

Ion 127.00 (126.70 to 127.70): BM001262.D (-536) (-)



Time-->









#10
Hexachlorobutadiene
Concen: 0.00 ng
RT: 10.68 min Scan# 564
Delta R.T. 0.12 min
Lab File: BM001315.D
Acq: 15 May 2015 05:06




Instrument :
BNA_M
ClientSampleId :
051115-P03-E-U-M


Tgt Ion: 225 Resp: 5
Ion Ratio Lower Upper
225 100
223 140.0 49.6 74.4#
227 0.0 51.4 77.2#



#11
2-Methylnaphthalene
Concen: 0.00 ng
RT: 11.90 min Scan# 657
Delta R.T. -0.00 min
Lab File: BM001315.D
Acq: 15 May 2015 05:06

Tgt Ion: 142 Resp: 24
Ion Ratio Lower Upper
142 100
141 100.0 76.1 114.1
115 100.0 33.1 49.7#









#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.12 min Scan# 861
Delta R.T. -0.00 min
Lab File: BM001315.D
Acq: 15 May 2015 05:06




Instrument :
BNA_M
ClientSampleId :
051115-P03-E-U-M


Tgt Ion:164 Resp: 25896
Ion Ratio Lower Upper
164 100
162 99.3 87.9 131.9
160 42.6 40.9 61.3



#13
2,4,6-Tribromophenol
Concen: 8.61 ng
RT: 15.63 min Scan# 940
Delta R.T. 0.00 min
Lab File: BM001315.D
Acq: 15 May 2015 05:06

Tgt Ion:330 Resp: 7645
Ion Ratio Lower Upper
330 100
332 99.9 89.0 133.4
141 0.0 0.0 0.0









#14
2-Fluorobiphenyl
Concen: 7.62 ng
RT: 12.73 min Scan# 737
Delta R.T. -0.00 min
Lab File: BM001315.D
Acq: 15 May 2015 05:06




Instrument :
BNA_M
ClientSampleId :
051115-P03-E-U-M


Tgt Ion:172 Resp: 61264
Ion Ratio Lower Upper
172 100
171 31.9 28.7 43.1
170 19.4 19.5 29.3#



#15
Acenaphthylene
Concen: 0.00 ng
RT: 13.83 min Scan# 837
Delta R.T. 0.00 min
Lab File: BM001315.D
Acq: 15 May 2015 05:06

Tgt Ion:152 Resp: 47
Ion Ratio Lower Upper
152 100
151 0.0 15.0 22.4#
153 0.0 10.1 15.1#





#16

Acenaphthene

Concen: 0.00 ng

RT: 14.18 min Scan# 866

Delta R.T. 0.00 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06

Instrument :

BNA_M

ClientSampleId :

051115-P03-E-U-M


Tgt Ion:154 Resp: 22

Ion Ratio Lower Upper




154 100

153 113.6 91.4 137.2

152 68.2 44.3 66.5#



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E



#17

Fluorene

Concen: 0.06 ng

RT: 14.74 min Scan# 911

Delta R.T. -0.43 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06


Tgt Ion:166 Resp: 391

Ion Ratio Lower Upper



166 100


165 0.0 70.6 106.0#

167 0.0 11.3 16.9#



Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E





#18

Phenanthrene-d10

Concen: 5.00 ng

RT: 16.85 min Scan# 980

Delta R.T. -0.00 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06

Instrument :

BNA_M

ClientSampleId :

051115-P03-E-U-M


Tgt Ion:188 Resp: 76202

Ion Ratio Lower Upper

188 100

94 3.3 7.8 11.8#




80 2.9 7.5 11.3#



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BM

Ion 80.00 (79.70 to 80.70): BM



#19

Hexachlorobenzene

Concen: 0.00 ng

RT: 16.18 min Scan# 958

Delta R.T. 0.00 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06


Tgt Ion:284 Resp: 9

Ion Ratio Lower Upper

284 100

142 0.0 0.0 0.0



249 0.0 22.7 34.1#




Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E





#20

Pentachlorophenol

Concen: 0.43 ng

RT: 16.52 min Scan# 969

Delta R.T. 0.00 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06

Instrument :

BNA_M

ClientSampleId :

051115-P03-E-U-M





Tgt Ion:266 Resp: 31

Ion Ratio Lower Upper

266 100

268 101.9 49.4 74.0#

264 96.2 67.8 101.6



#21

Phenanthrene

Concen: 0.03 ng

RT: 16.88 min Scan# 981

Delta R.T. 0.00 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06




Tgt Ion:178 Resp: 323


Ion Ratio Lower Upper

178 100

176 21.1 15.9 23.9

179 0.0 0.0 0.0





#22

Anthracene

Concen: 0.01 ng

RT: 16.98 min Scan# 984

Delta R.T. 0.00 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06

Instrument :

BNA_M

ClientSampleId :

051115-P03-E-U-M


Tgt Ion:178 Resp: 121

Ion Ratio Lower Upper

178 100

176 14.0 15.3 22.9#

179 0.0 0.0 0.0



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E

250

200

150

100

50



0

16.80

17.00

17.20

Time-->



#23

Fluoranthene

Concen: 0.03 ng

RT: 18.93 min Scan# 1083

Delta R.T. -0.00 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06


Tgt Ion:202 Resp: 551

Ion Ratio Lower Upper

202 100

101 12.7 9.1 13.7

203 17.6 13.8 20.6



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E

600

500

400

300

200

100

0


18.80


18.90

19.00

19.10

Time-->





#24

Chrysene-d12

Concen: 5.00 ng

RT: 21.07 min Scan# 1266

Delta R.T. -0.01 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06

Instrument :

BNA_M

ClientSampleId :

051115-P03-E-U-M





Tgt Ion:240 Resp: 46140

Ion Ratio Lower Upper

240 100

120 11.7 5.5 8.3#

236 28.4 20.4 30.6



#25

Pyrene

Concen: 0.04 ng

RT: 19.30 min Scan# 1113

Delta R.T. 0.00 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06




Tgt Ion:202 Resp: 537


Ion Ratio Lower Upper

202 100

200 21.6 16.4 24.6

203 20.9 13.9 20.9









#26
 Terphenyl-d14
 Concen: 8.50 ng
 RT: 19.51 min Scan# 1131
 Delta R.T. -0.01 min
 Lab File: BM001315.D
 Acq: 15 May 2015 05:06




Instrument :
 BNA_M
 ClientSampleId :
 051115-P03-E-U-M


Tgt Ion: 244 Resp: 52715
 Ion Ratio Lower Upper
 244 100
 212 9.4 7.2 10.8
 122 10.3 5.6 8.4#



#27
 Benzo(a)anthracene
 Concen: 0.04 ng
 RT: 21.06 min Scan# 1265
 Delta R.T. 0.00 min
 Lab File: BM001315.D
 Acq: 15 May 2015 05:06

Tgt Ion: 228 Resp: 442
 Ion Ratio Lower Upper
 228 100
 226 36.7 20.5 30.7#
 229 38.7 16.2 24.4#





#28

Chrysene

Concen: 0.02 ng

RT: 21.10 min Scan# 1269

Delta R.T. -0.01 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06

Instrument :

BNA_M

ClientSampleId :

051115-P03-E-U-M


Tgt Ion:228 Resp: 305

Ion Ratio Lower Upper

228 100

226 44.1 22.2 33.2#




229 36.8 16.5 24.7#



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#29

Indeno(1,2,3-cd)pyrene

Concen: 0.03 ng

RT: 25.26 min Scan# 1668

Delta R.T. -0.01 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06


Tgt Ion:276 Resp: 352

Ion Ratio Lower Upper

276 100

138 29.5 24.9 37.3



277 22.4 19.5 29.3




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E





#30

Perylene-d12

Concen: 5.00 ng

RT: 23.18 min Scan# 1469

Delta R.T. -0.01 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06

Instrument :

BNA_M

ClientSampleId :

051115-P03-E-U-M


Tgt Ion: 264 Resp: 44191

Ion Ratio Lower Upper

264 100

260 24.0 18.9 28.3




265 22.3 18.1 27.1



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E



#31

Benzo(b)fluoranthene

Concen: 0.05 ng

RT: 22.56 min Scan# 1409

Delta R.T. 0.00 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06


Tgt Ion: 252 Resp: 607

Ion Ratio Lower Upper

252 100

253 69.2 17.8 26.6#



125 44.6 10.4 15.6#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#32

Benzo(k)fluoranthene

Concen: 0.05 ng

RT: 22.56 min Scan# 1409

Delta R.T. -0.04 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06

Instrument :

BNA_M

ClientSampleId :

051115-P03-E-U-M


Tgt Ion: 252 Resp: 607

Ion Ratio Lower Upper

252 100

253 69.2 18.6 27.8#




125 44.6 10.2 15.4#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#33

Benzo(a)pyrene

Concen: 0.02 ng

RT: 23.08 min Scan# 1459

Delta R.T. -0.01 min

Lab File: BM001315.D

Acq: 15 May 2015 05:06


Tgt Ion: 252 Resp: 261

Ion Ratio Lower Upper

252 100

253 78.2 18.3 27.5#



125 50.5 12.2 18.2#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#34
Dibenzo(a,h)anthracene
Concen: 0.02 ng
RT: 25.28 min Scan# 1670
Delta R.T. -0.01 min
Lab File: BM001315.D
Acq: 15 May 2015 05:06

Instrument :
BNA_M
ClientSampleId :
051115-P03-E-U-M

Tgt Ion: 278 Resp: 221

Ion	Ratio	Lower	Upper
278	100		
139	68.8	16.6	24.8#
279	77.1	20.0	30.0#



#35
Benzo(g,h,i)perylene
Concen: 0.03 ng
RT: 25.89 min Scan# 1729
Delta R.T. -0.02 min
Lab File: BM001315.D
Acq: 15 May 2015 05:06

Tgt Ion: 276 Resp: 326

Ion	Ratio	Lower	Upper
276	100		
277	49.7	19.9	29.9#
138	54.5	21.0	31.4#

