

Data Path : Z:\HPCHEM1\BNA M\Data\BM051615\
 Data File : BM001294.D
 Acq On : 14 May 2015 16:06
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 15 05:26:32 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 13 05:24:15 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.582	0.567	2.6	88	0.00
3	n-Nitrosodimethylamine	0.566	0.561	0.9	110	0.00
4 S	2-Fluorophenol	0.974	0.919	5.6	97	0.00
5 S	Phenol-d6	1.099	1.034	5.9	99	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
7 S	Nitrobenzene-d5	0.263	0.253	3.8	98	0.00
8	Nitrobenzene	0.276	0.258	6.5	97	0.00
9	Naphthalene	1.064	1.041	2.2	97	0.00
10	Hexachlorobutadiene	0.208	0.200	3.8	98	0.00
11	2-Methylnaphthalene	0.692	0.656	5.2	97	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
13 S	2,4,6-Tribromophenol	0.149	0.123	17.4	95	0.00
14 S	2-Fluorobiphenyl	1.552	1.473	5.1	95	0.00
15	Acenaphthylene	2.058	1.942	5.6	97	0.00
16	Acenaphthene	1.318	1.245	5.5	97	0.00
17	Fluorene	1.194	1.018	14.7	92	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
19	Hexachlorobenzene	0.322	0.276	14.3	100	0.00
20	Pentachlorophenol	0.067	0.043	35.8#	97	0.00
21	Phenanthrene	0.760	0.650	14.5	100	0.00
22	Anthracene	0.741	0.616	16.9	98	0.00
23	Fluoranthene	1.193	0.971	18.6	97	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
25	Pyrene	1.545	1.421	8.0	98	0.00
26 S	Terphenyl-d14	0.672	0.605	10.0	93	-0.01
27	Benzo(a)anthracene	1.357	1.228	9.5	99	0.00
28	Chrysene	1.376	1.307	5.0	98	0.00
29	Indeno(1,2,3-cd)pyrene	1.398	1.427	-2.1	107	0.00
30 I	Perylene-d12	1.000	1.000	0.0	101	-0.01
31	Benzo(b)fluoranthene	1.465	1.360	7.2	102	0.00
32	Benzo(k)fluoranthene	1.394	1.294	7.2	100	0.00
33 C	Benzo(a)pyrene	1.269	1.190	6.2	104	0.00
34	Dibenzo(a,h)anthracene	1.193	1.184	0.8	106	0.00
35	Benzo(q,h,i)perylene	1.309	1.297	0.9	107	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0