

Data Path : Z:\HPCHEM1\BNA M\Data\BM051615\
 Data File : BM001309.D
 Acq On : 15 May 2015 01:32
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 15 05:34:12 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 15 05:14:53 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	98	0.00
2	1,4-Dioxane	1.000	0.668	33.2#	80	0.00
3	n-Nitrosodimethylamine	1.000	0.984	1.6	111	0.00
4 S	2-Fluorophenol	1.000	0.946	5.4	98	-0.02
5 S	Phenol-d6	1.000	0.948	5.2	101	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	99	0.00
7 S	Nitrobenzene-d5	1.000	0.874	12.6	98	0.00
8	Nitrobenzene	1.000	0.848	15.2	97	0.00
9	Naphthalene	1.000	0.963	3.7	99	0.00
10	Hexachlorobutadiene	1.000	0.934	6.6	99	0.00
11	2-Methylnaphthalene	1.000	0.936	6.4	99	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	97	0.00
13 S	2,4,6-Tribromophenol	1.000	0.720	28.0#	85	0.00
14 S	2-Fluorobiphenyl	1.000	0.951	4.9	97	0.00
15	Acenaphthylene	1.000	0.933	6.7	97	0.00
16	Acenaphthene	1.000	0.938	6.2	98	0.00
17	Fluorene	1.000	0.916	8.4	100	-0.03
18 I	Phenanthrene-d10	5.000	5.000	0.0	132	0.00
19	Hexachlorobenzene	1.000	0.659	34.1#	94	0.00
20	Pentachlorophenol	1.000	0.916	8.4	138	0.00
21	Phenanthrene	1.000	0.977	2.3	140	0.00
22	Anthracene	1.000	0.937	6.3	135	0.00
23	Fluoranthene	1.000	0.680	32.0#	99	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	100	-0.01
25	Pyrene	1.000	0.917	8.3	100	0.00
26 S	Terphenyl-d14	1.000	0.904	9.6	97	-0.01
27	Benzo(a)anthracene	1.000	0.934	6.6	105	-0.01
28	Chrysene	1.000	0.934	6.6	99	-0.01
29	Indeno(1,2,3-cd)pyrene	1.000	1.034	-3.4	111	-0.01
30 I	Perylene-d12	5.000	5.000	0.0	104	-0.02
31	Benzo(b)fluoranthene	1.000	0.930	7.0	106	-0.01
32	Benzo(k)fluoranthene	1.000	0.932	6.8	104	-0.01
33 C	Benzo(a)pyrene	1.000	0.938	6.2	108	-0.01
34	Dibenzo(a,h)anthracene	1.000	0.996	0.4	111	-0.02
35	Benzo(q,h,i)perylene	1.000	0.987	1.3	111	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0