

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051622\
 Data File : BM035121.D
 Acq On : 17 May 2022 02:09
 Operator : CG/JU
 Sample : N2801-06
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EXLB9

Quant Time: May 17 04:04:59 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM051222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 12 18:15:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.913	152	2713	0.400	ng/ul	0.00
4) Naphthalene-d8	10.716	136	9338	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.548	164	6798	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.289	188	14996	0.400	ng/ul	0.00
17) Chrysene-d12	21.473	240	12917	0.400	ng/ul	0.00
23) Perylene-d12	23.858	264	11741	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	14757	4.541	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.305	152	5073	0.340	ng/ul	0.00
18) Fluoranthene-d10	19.318	212	16036	0.362	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.759	128	8078	0.262	ng/ul	95
7) 2-Methylnaphthalene	12.376	142	1098	0.054	ng/ul	100
8) 1-Methylnaphthalene	12.591	142	5658	0.300	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051622\
Data File : BM035121.D
Acq On : 17 May 2022 02:09
Operator : CG/JU
Sample : N2801-06
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EXLB9

Quant Time: May 17 04:04:59 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM051222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 12 18:15:39 2022
Response via : Initial Calibration

