

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051715\
 Data File : BM001325.D
 Acq On : 15 May 2015 12:41
 Operator : TP/IZ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.482

Quant Time: May 16 02:27:57 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM051115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 16 02:13:44 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.43	152	1103	0.40	ng/ul	0.00
2) Naphthalene-d8	10.23	136	3834	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.12	164	1944	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.86	188	4051	0.40	ng/ul	0.00
16) Chrysene-d12	21.08	240	3418	0.40	ng/ul	0.00
20) Perylene-d12	23.18	264	3236	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	11.83	152	1964	0.40	ng/ul	0.00
14) Fluoranthene-d10	18.90	212	3500	0.40	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	10.26	128	3845	0.40	ng/ul	97
5) 2-Methylnaphthalene	11.90	142	2555	0.40	ng/ul	99
7) Acenaphthylene	13.82	152	3481	0.41	ng/ul	99
8) Acenaphthene	14.17	153	2721	0.40	ng/ul	96
9) Fluorene	15.18	166	3036	0.40	ng/ul	94
11) Pentachlorophenol	16.53	266	234	0.27	ng/ul	92
12) Phenanthrene	16.91	178	4621	0.40	ng/ul	98
13) Anthracene	17.00	178	4269	0.40	ng/ul	99
15) Fluoranthene	18.94	202	5146	0.40	ng/ul	99
17) Pyrene	19.30	202	5348	0.40	ng/ul	99
18) Benzo(a)anthracene	21.06	228	4250	0.41	ng/ul	99
19) Chrysene	21.11	228	5032	0.42	ng/ul	98
21) Benzo(b)fluoranthene	22.56	252	4870	0.42	ng/ul	99
22) Benzo(k)fluoranthene	22.60	252	4630	0.39	ng/ul	99
23) Benzo(a)pyrene	23.09	252	4572	0.41	ng/ul	97
24) Indeno(1,2,3-cd)pyrene	25.27	276	5143	0.42	ng/ul	98
25) Dibenzo(a,h)anthracene	25.28	278	4081	0.42	ng/ul	97
26) Benzo(g,h,i)perylene	25.90	276	4567	0.42	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051715\
Data File : BM001325.D
Acq On : 15 May 2015 12:41
Operator : TP/IZ
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.482

Quant Time: May 16 02:27:57 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM051115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 16 02:13:44 2015
Response via : Initial Calibration

