

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051815\
 Data File : BM001328.D
 Acq On : 15 May 2015 15:19
 Operator : TP/IZ
 Sample : G2182-02
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 S32-0483

Quant Time: May 18 11:19:18 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 16 01:36:33 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.45	152	14104	5.00	ng	0.02
6) Naphthalene-d8	10.22	136	48974	5.00	ng	0.00
12) Acenaphthene-d10	14.12	164	25198	5.00	ng	0.00
18) Phenanthrene-d10	16.85	188	73067	5.00	ng	0.00
24) Chrysene-d12	21.08	240	47192	5.00	ng	0.01
30) Perylene-d12	23.18	264	45502	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.02	112	10990	4.00	ng	0.00
5) Phenol-d6	6.63	99	7511	2.42	ng	0.00
7) Nitrobenzene-d5	8.60	82	18575	5.99	ng	0.00
13) 2,4,6-Tribromophenol	15.63	330	7373	8.53	ng	0.03
14) 2-Fluorobiphenyl	12.73	172	56526	7.23	ng	0.00
26) Terphenyl-d14	19.51	244	54085	8.53	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051815\
Data File : BM001328.D
Acq On : 15 May 2015 15:19
Operator : TP/IZ
Sample : G2182-02
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
S32-0483

Quant Time: May 18 11:19:18 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM051315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 16 01:36:33 2015
Response via : Initial Calibration

