

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051815\
 Data File : BM001327.D
 Acq On : 15 May 2015 14:26
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 16 01:49:33 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 13 05:24:15 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	94	-0.02
2	1,4-Dioxane	1.000	0.925	7.5	90	0.00
3	n-Nitrosodimethylamine	1.000	1.004	-0.4	108	0.00
4 S	2-Fluorophenol	1.000	0.962	3.8	95	-0.02
5 S	Phenol-d6	1.000	0.948	5.2	96	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	93	0.00
7 S	Nitrobenzene-d5	1.000	0.876	12.4	92	0.00
8	Nitrobenzene	1.000	0.866	13.4	93	0.00
9	Naphthalene	1.000	0.969	3.1	93	0.00
10	Hexachlorobutadiene	1.000	0.970	3.0	96	0.00
11	2-Methylnaphthalene	1.000	0.928	7.2	92	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	90	0.00
13 S	2,4,6-Tribromophenol	1.000	0.685	31.5#	74	-0.03
14 S	2-Fluorobiphenyl	1.000	0.967	3.3	91	0.00
15	Acenaphthylene	1.000	0.949	5.1	92	0.00
16	Acenaphthene	1.000	0.952	4.8	92	0.00
17	Fluorene	1.000	0.963	3.7	97	-0.03
18 I	Phenanthrene-d10	5.000	5.000	0.0	121	0.00
19	Hexachlorobenzene	1.000	0.664	33.6#	87	0.00
20	Pentachlorophenol	1.000	0.914	8.6	126	0.00
21	Phenanthrene	1.000	1.119	-11.9	148	0.00
22	Anthracene	1.000	0.993	0.7	131	0.00
23	Fluoranthene	1.000	0.701	29.9#	94	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	96	-0.01
25	Pyrene	1.000	0.923	7.7	97	0.00
26 S	Terphenyl-d14	1.000	0.918	8.2	94	-0.01
27	Benzo(a)anthracene	1.000	0.926	7.4	100	-0.01
28	Chrysene	1.000	0.947	5.3	96	-0.01
29	Indeno(1,2,3-cd)pyrene	1.000	1.026	-2.6	106	-0.01
30 I	Perylene-d12	5.000	5.000	0.0	100	-0.01
31	Benzo(b)fluoranthene	1.000	0.901	9.9	99	-0.01
32	Benzo(k)fluoranthene	1.000	0.979	2.1	105	-0.01
33 C	Benzo(a)pyrene	1.000	0.944	5.6	104	-0.01
34	Dibenzo(a,h)anthracene	1.000	0.992	0.8	106	-0.01
35	Benzo(g,h,i)perylene	1.000	0.989	1.1	106	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0