

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051815\
 Data File : BM001329.D
 Acq On : 15 May 2015 15:56
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 16 01:50:29 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 16 01:36:33 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	98	-0.02
2	1,4-Dioxane	1.000	0.647	35.3#	78	0.00
3	n-Nitrosodimethylamine	1.000	0.894	10.6	101	0.00
4 S	2-Fluorophenol	1.000	0.979	2.1	102	-0.02
5 S	Phenol-d6	1.000	0.955	4.5	102	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	97	0.00
7 S	Nitrobenzene-d5	1.000	0.899	10.1	99	0.00
8	Nitrobenzene	1.000	0.871	12.9	98	0.00
9	Naphthalene	1.000	0.957	4.3	96	0.00
10	Hexachlorobutadiene	1.000	0.967	3.3	100	0.00
11	2-Methylnaphthalene	1.000	0.942	5.8	97	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	94	0.00
13 S	2,4,6-Tribromophenol	1.000	0.715	28.5#	82	0.00
14 S	2-Fluorobiphenyl	1.000	0.969	3.1	96	0.00
15	Acenaphthylene	1.000	0.937	6.3	95	0.00
16	Acenaphthene	1.000	0.947	5.3	96	0.00
17	Fluorene	1.000	0.882	11.8	94	-0.03
18 I	Phenanthrene-d10	5.000	5.000	0.0	126	0.00
19	Hexachlorobenzene	1.000	0.686	31.4#	93	0.00
20	Pentachlorophenol	1.000	0.874	12.6	120	0.00
21	Phenanthrene	1.000	0.968	3.2	133	0.00
22	Anthracene	1.000	0.923	7.7	127	0.00
23	Fluoranthene	1.000	0.692	30.8#	96	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	96	-0.01
25	Pyrene	1.000	0.933	6.7	98	0.00
26 S	Terphenyl-d14	1.000	0.930	7.0	96	-0.01
27	Benzo(a)anthracene	1.000	0.932	6.8	101	0.00
28	Chrysene	1.000	0.965	3.5	99	-0.01
29	Indeno(1,2,3-cd)pyrene	1.000	1.025	-2.5	106	0.00
30 I	Perylene-d12	5.000	5.000	0.0	101	-0.01
31	Benzo(b)fluoranthene	1.000	0.987	1.3	109	0.00
32	Benzo(k)fluoranthene	1.000	0.893	10.7	96	0.00
33 C	Benzo(a)pyrene	1.000	0.942	5.8	105	0.00
34	Dibenzo(a,h)anthracene	1.000	0.986	1.4	106	-0.01
35	Benzo(g,h,i)perylene	1.000	0.982	1.8	106	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0