

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051815\
 Data File : BM001329.D
 Acq On : 15 May 2015 15:56
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 16 01:50:29 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 16 01:36:33 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	-0.02
2	1,4-Dioxane	0.582	0.500	14.1	78	0.00
3	n-Nitrosodimethylamine	0.566	0.506	10.6	101	0.00
4 S	2-Fluorophenol	0.974	0.954	2.1	102	-0.02
5 S	Phenol-d6	1.099	1.049	4.5	102	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
7 S	Nitrobenzene-d5	0.263	0.252	4.2	99	0.00
8	Nitrobenzene	0.276	0.257	6.9	98	0.00
9	Naphthalene	1.064	1.018	4.3	96	0.00
10	Hexachlorobutadiene	0.208	0.201	3.4	100	0.00
11	2-Methylnaphthalene	0.692	0.652	5.8	97	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
13 S	2,4,6-Tribromophenol	0.149	0.108	27.5#	82	0.00
14 S	2-Fluorobiphenyl	1.552	1.504	3.1	96	0.00
15	Acenaphthylene	2.058	1.930	6.2	95	0.00
16	Acenaphthene	1.318	1.247	5.4	96	0.00
17	Fluorene	1.194	1.053	11.8	94	-0.03
18 I	Phenanthrene-d10	1.000	1.000	0.0	126	0.00
19	Hexachlorobenzene	0.322	0.221	31.4#	93	0.00
20	Pentachlorophenol	0.067	0.046	31.3#	120	0.00
21	Phenanthrene	0.760	0.736	3.2	133	0.00
22	Anthracene	0.741	0.683	7.8	127	0.00
23	Fluoranthene	1.193	0.825	30.8#	96	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	96	-0.01
25	Pyrene	1.545	1.441	6.7	98	0.00
26 S	Terphenyl-d14	0.672	0.625	7.0	96	-0.01
27	Benzo(a)anthracene	1.357	1.265	6.8	101	0.00
28	Chrysene	1.376	1.328	3.5	99	-0.01
29	Indeno(1,2,3-cd)pyrene	1.398	1.433	-2.5	106	0.00
30 I	Perylene-d12	1.000	1.000	0.0	101	-0.01
31	Benzo(b)fluoranthene	1.465	1.447	1.2	109	0.00
32	Benzo(k)fluoranthene	1.394	1.245	10.7	96	0.00
33 C	Benzo(a)pyrene	1.269	1.195	5.8	105	0.00
34	Dibenzo(a,h)anthracene	1.193	1.176	1.4	106	-0.01
35	Benzo(g,h,i)perylene	1.309	1.286	1.8	106	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0