

Data Path : Z:\HPCHEM1\BNA_M\Data\BM051815\
 Data File : BM001331.D
 Acq On : 18 May 2015 15:19
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 19 02:22:21 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 13 05:24:15 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	118	-0.02
2	1,4-Dioxane	0.582	0.537	7.7	101	-0.01
3	n-Nitrosodimethylamine	0.566	0.594	-4.9	142	-0.02
4 S	2-Fluorophenol	0.974	0.904	7.2	116	-0.02
5 S	Phenol-d6	1.099	1.039	5.5	121	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	124	0.00
7 S	Nitrobenzene-d5	0.263	0.254	3.4	127	0.00
8	Nitrobenzene	0.276	0.264	4.3	128	0.00
9	Naphthalene	1.064	1.024	3.8	123	0.00
10	Hexachlorobutadiene	0.208	0.203	2.4	129	0.00
11	2-Methylnaphthalene	0.692	0.688	0.6	131	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	133	0.00
13 S	2,4,6-Tribromophenol	0.149	0.114	23.5	122	0.00
14 S	2-Fluorobiphenyl	1.552	1.465	5.6	132	0.00
15	Acenaphthylene	2.058	1.991	3.3	139	0.00
16	Acenaphthene	1.318	1.246	5.5	135	0.00
17	Fluorene	1.194	1.070	10.4	135	-0.03
18 I	Phenanthrene-d10	1.000	1.000	0.0	183#	0.00
19	Hexachlorobenzene	0.322	0.225	30.1#	138	0.00
20	Pentachlorophenol	0.067	0.055	17.9	211#	0.00
21	Phenanthrene	0.760	0.755	0.7	198#	0.00
22	Anthracene	0.741	0.698	5.8	189#	0.00
23	Fluoranthene	1.193	0.840	29.6#	143	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	134	0.00
25	Pyrene	1.545	1.536	0.6	146	0.00
26 S	Terphenyl-d14	0.672	0.666	0.9	141	-0.01
27	Benzo(a)anthracene	1.357	1.253	7.7	139	0.00
28	Chrysene	1.376	1.304	5.2	134	0.00
29	Indeno(1,2,3-cd)pyrene	1.398	1.138	18.6	117	0.00
30 I	Perylene-d12	1.000	1.000	0.0	122	-0.01
31	Benzo(b)fluoranthene	1.465	1.470	-0.3	134	0.00
32	Benzo(k)fluoranthene	1.394	1.301	6.7	122	0.00
33 C	Benzo(a)pyrene	1.269	1.192	6.1	126	0.00
34	Dibenzo(a,h)anthracene	1.193	1.048	12.2	114	-0.01
35	Benzo(g,h,i)perylene	1.309	1.161	11.3	116	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0