

Data Path : Z:\HPCHEM1\BNA_M\Data\BM051815\
 Data File : BM001340.D
 Acq On : 18 May 2015 20:59
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 19 02:32:58 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 19 02:11:49 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	101	0.00
2	1,4-Dioxane	1.000	0.403	59.7#	66	-0.01
3	n-Nitrosodimethylamine	1.000	0.927	7.3	107	0.00
4 S	2-Fluorophenol	1.000	0.937	6.3	100	-0.02
5 S	Phenol-d6	1.000	0.969	3.1	106	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	102	0.00
7 S	Nitrobenzene-d5	1.000	0.913	8.7	106	0.00
8	Nitrobenzene	1.000	0.880	12.0	105	0.00
9	Naphthalene	1.000	0.964	3.6	102	0.00
10	Hexachlorobutadiene	1.000	0.954	4.6	104	0.00
11	2-Methylnaphthalene	1.000	0.956	4.4	104	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	101	0.00
13 S	2,4,6-Tribromophenol	1.000	0.700	30.0#	86	0.00
14 S	2-Fluorobiphenyl	1.000	0.962	3.8	103	0.00
15	Acenaphthylene	1.000	0.947	5.3	103	0.00
16	Acenaphthene	1.000	0.947	5.3	103	0.00
17	Fluorene	1.000	0.847	15.3	97	-0.03
18 I	Phenanthrene-d10	5.000	5.000	0.0	130	0.00
19	Hexachlorobenzene	1.000	0.714	28.6#	100	0.00
20	Pentachlorophenol	1.000	0.909	9.1	134	0.00
21	Phenanthrene	1.000	0.959	4.1	136	0.00
22	Anthracene	1.000	0.927	7.3	132	0.00
23	Fluoranthene	1.000	0.704	29.6#	101	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	101	-0.01
25	Pyrene	1.000	0.928	7.2	103	0.00
26 S	Terphenyl-d14	1.000	0.955	4.5	103	-0.01
27	Benzo(a)anthracene	1.000	0.923	7.7	105	-0.01
28	Chrysene	1.000	0.947	5.3	102	-0.01
29	Indeno(1,2,3-cd)pyrene	1.000	0.941	5.9	102	-0.01
30 I	Perylene-d12	5.000	5.000	0.0	101	-0.01
31	Benzo(b)fluoranthene	1.000	0.906	9.4	100	-0.01
32	Benzo(k)fluoranthene	1.000	0.976	2.4	105	-0.01
33 C	Benzo(a)pyrene	1.000	0.930	7.0	103	-0.01
34	Dibenzo(a,h)anthracene	1.000	0.949	5.1	102	-0.01
35	Benzo(g,h,i)perylene	1.000	0.942	5.8	102	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0