

Data Path : Z:\HPCHEM1\BNA_M\Data\BM051815\
 Data File : BM001340.D
 Acq On : 18 May 2015 20:59
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 19 02:32:58 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 19 02:11:49 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.582	0.411	29.4#	66	-0.01
3	n-Nitrosodimethylamine	0.566	0.525	7.2	107	0.00
4 S	2-Fluorophenol	0.974	0.913	6.3	100	-0.02
5 S	Phenol-d6	1.099	1.065	3.1	106	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
7 S	Nitrobenzene-d5	0.263	0.257	2.3	106	0.00
8	Nitrobenzene	0.276	0.260	5.8	105	0.00
9	Naphthalene	1.064	1.026	3.6	102	0.00
10	Hexachlorobutadiene	0.208	0.198	4.8	104	0.00
11	2-Methylnaphthalene	0.692	0.662	4.3	104	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
13 S	2,4,6-Tribromophenol	0.149	0.105	29.5#	86	0.00
14 S	2-Fluorobiphenyl	1.552	1.494	3.7	103	0.00
15	Acenaphthylene	2.058	1.950	5.2	103	0.00
16	Acenaphthene	1.318	1.248	5.3	103	0.00
17	Fluorene	1.194	1.011	15.3	97	-0.03
18 I	Phenanthrene-d10	1.000	1.000	0.0	130	0.00
19	Hexachlorobenzene	0.322	0.230	28.6#	100	0.00
20	Pentachlorophenol	0.067	0.049	26.9#	134	0.00
21	Phenanthrene	0.760	0.729	4.1	136	0.00
22	Anthracene	0.741	0.687	7.3	132	0.00
23	Fluoranthene	1.193	0.840	29.6#	101	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	101	-0.01
25	Pyrene	1.545	1.435	7.1	103	0.00
26 S	Terphenyl-d14	0.672	0.642	4.5	103	-0.01
27	Benzo(a)anthracene	1.357	1.253	7.7	105	-0.01
28	Chrysene	1.376	1.303	5.3	102	-0.01
29	Indeno(1,2,3-cd)pyrene	1.398	1.317	5.8	102	-0.01
30 I	Perylene-d12	1.000	1.000	0.0	101	-0.01
31	Benzo(b)fluoranthene	1.465	1.327	9.4	100	-0.01
32	Benzo(k)fluoranthene	1.394	1.361	2.4	105	-0.01
33 C	Benzo(a)pyrene	1.269	1.180	7.0	103	-0.01
34	Dibenzo(a,h)anthracene	1.193	1.132	5.1	102	-0.01
35	Benzo(g,h,i)perylene	1.309	1.233	5.8	102	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0