

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051816\
 Data File : BM005473.D
 Acq On : 18 May 2016 17:17
 Operator : UM/SJ
 Sample : PB90367BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK67

Quant Time: May 19 01:44:17 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM051816.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 18 14:47:27 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	33650	20.00	ng/ul	0.00
18) Naphthalene-d8	10.63	136	173153	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.47	164	118939	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.20	188	318436	20.00	ng/ul	0.00
75) Chrysene-d12	21.38	240	478749	20.00	ng/ul	0.00
83) Perylene-d12	23.67	264	534727	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.32	96	4212	4.82	ng/uL	0.00
5) Phenol-d5	6.99	99	86692	26.82	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.17	67	55809	31.52	ng/ul	0.00
9) 2-Chlorophenol-d4	7.37	132	69795	28.11	ng/ul	0.00
13) 4-Methylphenol-d8	8.53	113	75692	27.49	ng/ul	0.00
19) Nitrobenzene-d5	8.99	128	36099	28.96	ng/ul	0.00
22) 2-Nitrophenol-d4	9.71	143	39531	29.07	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.24	165	73518	27.20	ng/ul	0.00
29) 4-Chloroaniline-d4	10.76	131	106087	34.13	ng/ul	0.00
43) Dimethylphthalate-d6	13.87	166	331527	33.91	ng/ul	0.00
46) Acenaphthylene-d8	14.16	160	363060	31.45	ng/ul	0.00
51) 4-Nitrophenol-d4	14.64	143	60399	30.11	ng/ul	0.00
57) Fluorene-d10	15.46	176	287796	33.36	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.57	200	42642	25.40	ng/ul	0.00
70) Anthracene-d10	17.30	188	492830	33.65	ng/ul	0.00
76) Pyrene-d10	19.59	212	623017	32.36	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.52	264	821914	34.38	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051816\
Data File : BM005473.D
Acq On : 18 May 2016 17:17
Operator : UM/SJ
Sample : PB90367BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK67

Quant Time: May 19 01:44:17 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM051816.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed May 18 14:47:27 2016
Response via : Initial Calibration

