


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM051821\
 Data File : BM030061.D
 Acq On : 18 May 2021 10:20
 Operator : CG/JU
 Sample : DFTPP38
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP038

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050721.M
 Title : SVOA CALIBRATION
 Last Update : Tue May 18 16:38:21 2021



AutoFind: Scans 2462, 2463, 2464; Background Corrected with Scan 2456

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.0	89529	PASS
68	69	0.00	2	1.9	1731	PASS
69	198	0.00	100	46.8	91072	PASS
70	69	0.00	2	0.5	495	PASS
127	198	10	80	52.4	101981	PASS
197	198	0.00	2	0.7	1322	PASS
198	198	100	100	100.0	194602	PASS
199	198	5	9	6.8	13249	PASS
275	198	10	60	30.5	59429	PASS
365	198	1	100	4.4	8552	PASS
441	198	0.01	100	13.6	26394	PASS
442	198	50	100	89.6	174314	PASS
443	442	15	24	19.6	34090	PASS

m/z	Abundance
38.90	159.0
40.00	890.0
41.00	303.0
43.00	553.0
44.00	1805.0
44.80	217.0
49.00	228.0
51.00	188.0
55.10	184.0
73.00	823.0
83.90	244.0
133.00	277.0
147.00	411.0
190.90	230.0
193.00	286.0
207.00	2136.0
208.00	323.0
209.00	234.0
281.10	590.0
283.10	156.0

Instrument :
BNA_M
ClientSampleId :
DFTPP038

m/z	Abundance
37.00	755.0
38.00	1592.0
39.10	9164.0
40.00	1364.0
41.00	413.0
42.00	205.0
43.00	525.0
44.00	2051.0
45.00	548.0
48.10	157.0
49.00	1173.0
50.00	29672.0
51.00	114872.0
52.00	6195.0
53.10	217.0
55.00	821.0
56.00	3503.0
57.00	8068.0
58.00	403.0
59.00	159.0
60.00	335.0
61.00	1350.0
62.00	1721.0
63.00	4275.0
63.90	584.0
65.00	2072.0
66.10	278.0
66.90	217.0
68.00	2203.0
69.00	115384.0
70.00	778.0
73.00	2072.0
74.00	11083.0
75.00	18144.0
76.10	5773.0
77.00	121480.0
78.00	8624.0
79.00	8892.0
80.00	6472.0
81.00	8667.0
82.00	1996.0
83.00	1892.0
83.90	538.0
85.00	1459.0
86.00	2685.0
87.00	1495.0
88.00	493.0
91.00	2190.0
92.00	2477.0
93.00	13908.0
94.00	884.0
95.10	381.0
96.00	717.0
97.10	300.0
98.00	11308.0
99.00	9040.0
99.90	672.0
101.00	4884.0
101.80	225.0
103.00	1643.0
104.00	3258.0
105.00	2836.0
106.00	1037.0
107.00	36680.0
108.00	5543.0
109.10	1103.0
110.00	61552.0
111.00	9504.0
112.00	1456.0
113.00	531.0
115.00	162.0
116.00	2099.0
117.00	30784.0
118.00	2487.0
119.00	459.0
119.90	477.0
120.90	167.0
122.00	2656.0
123.00	3622.0
124.00	1735.0
125.00	1379.0
127.00	124528.0
128.00	9650.0
129.00	46824.0
130.00	4349.0
131.00	963.0
132.00	499.0
133.10	401.0
133.90	1539.0
135.00	3888.0
136.00	1446.0
137.00	2352.0

Instrument :
BNA_M
ClientSampleId :
DFTPP038

137.90	554.0
138.90	197.0
140.00	653.0
141.00	6192.0
142.00	2399.0
143.00	1642.0
143.80	417.0
144.80	385.0
146.00	969.0
147.00	3698.0
148.00	7364.0
149.00	1445.0
150.00	392.0
151.00	865.0
152.00	542.0
153.00	1824.0
153.90	1521.0
155.00	3213.0
156.00	4936.0
157.00	1043.0
158.00	998.0
159.00	911.0
160.00	1928.0
161.00	2898.0
162.00	840.0
163.00	271.0
163.90	447.0
165.00	2330.0
166.00	1751.0
167.00	12017.0
168.00	7043.0
169.10	1175.0
170.00	340.0
171.00	494.0
171.90	1066.0
172.90	1491.0
174.00	2686.0
175.00	4844.0
176.00	1416.0
177.00	1963.0
178.00	1028.0
179.00	7874.0
180.00	7738.0
181.00	2893.0
181.90	491.0
182.90	274.0
184.00	977.0
185.00	4738.0
186.00	31936.0
187.00	9130.0
188.00	1088.0
188.90	2241.0
190.00	460.0
191.00	1045.0
192.00	2737.0
193.00	3154.0
193.90	637.0
194.90	320.0
196.00	6657.0
197.10	1568.0
198.00	221568.0
199.00	15353.0
199.90	1179.0
201.40	1234.0
203.00	1641.0
204.00	9321.0
205.00	14891.0
206.00	62264.0
207.00	10324.0
208.00	2366.0
209.00	1119.0
210.30	1115.0
211.00	2393.0
213.10	266.0
215.00	687.0
216.00	1362.0
217.00	17792.0
218.00	2122.0
219.00	260.0
221.00	9362.0
222.00	2016.0
223.00	3706.0
224.00	35552.0
225.00	8334.0
226.10	1192.0
227.00	15200.0
228.00	2150.0
229.00	3118.0
229.90	462.0
231.00	1204.0
232.10	306.0
233.00	274.0
234.00	772.0
235.00	981.0
236.00	983.0
237.00	2099.0

Instrument :
BNA_M
ClientSampleId :
DFTPP038

238.00	329.0
238.90	568.0
240.00	857.0
241.00	570.0
242.00	2061.0
243.00	1904.0
244.00	25312.0
245.00	3852.0
246.00	5017.0
246.90	1129.0
248.00	219.0
248.90	885.0
249.90	223.0
251.00	335.0
252.00	304.0
253.00	707.0
254.20	1207.0
255.00	140416.0
256.00	20208.0
257.00	1950.0
258.00	8053.0
259.00	1330.0
259.90	321.0
264.10	248.0
265.00	3546.0
266.00	737.0
270.00	201.0
271.00	370.0
271.90	459.0
273.00	4058.0
274.00	10444.0
275.00	62984.0
276.00	8381.0
277.00	5844.0
278.00	1199.0
279.00	190.0
281.00	432.0
281.90	418.0
283.00	706.0
284.00	565.0
285.00	1042.0
286.00	150.0
288.90	254.0
290.10	167.0
292.90	1408.0
293.90	451.0
295.10	434.0
296.00	18960.0
297.00	2611.0
298.00	218.0
301.00	219.0
301.90	269.0
303.00	2253.0
304.00	590.0
307.90	260.0
309.00	206.0
309.90	285.0
313.10	152.0
314.00	770.0
315.00	2128.0
316.00	1126.0
316.90	270.0
321.00	457.0
322.00	237.0
323.10	4905.0
324.10	934.0
326.90	1102.0
327.90	496.0
329.00	214.0
331.90	442.0
333.00	511.0
334.00	3731.0
335.00	887.0
336.10	164.0
341.10	626.0
341.90	160.0
346.00	1149.0
347.00	338.0
352.00	1659.0
353.00	1140.0
354.00	1978.0
354.90	420.0
358.90	155.0
364.20	279.0
365.00	8587.0
366.00	1245.0
370.90	658.0
372.00	3023.0
373.10	724.0
383.00	797.0
384.00	239.0
389.90	388.0
390.90	312.0
392.20	157.0
400.90	150.0
402.00	1288.0

Instrument :
BNA_M
ClientSampleId :
DFTPP038

403.00	1751.0
404.00	707.0
405.10	227.0
421.00	1239.0
422.00	1297.0
423.00	9835.0
424.00	3664.0
424.90	522.0
441.10	24160.0
442.10	154752.0
443.00	32040.0
444.00	2909.0
445.00	207.0

Instrument :
BNA_M
ClientSampleId :
DFTPP038

m/z	Abundance
36.00	234.0
37.00	429.0
38.00	1332.0
39.10	7988.0
40.00	1238.0
41.00	347.0
43.00	416.0
44.00	2072.0
45.00	622.0
46.90	222.0
49.00	853.0
50.00	26688.0
51.00	103064.0
52.00	4518.0
53.00	252.0
55.00	583.0
56.00	3002.0
57.00	7050.0
58.00	358.0
59.00	238.0
60.10	187.0
61.00	1438.0
62.00	1506.0
63.00	4028.0
63.90	552.0
65.10	1937.0
66.00	150.0
66.90	215.0
68.10	1972.0
69.00	104648.0
70.00	496.0
73.00	1888.0
74.00	10186.0
75.00	16148.0
76.10	5334.0
77.10	112272.0
78.00	7787.0
79.00	7713.0
80.00	5361.0
81.00	8614.0
82.00	2161.0
83.00	1871.0
83.90	490.0
85.00	1268.0
86.00	2693.0
87.00	1205.0
87.90	378.0
89.00	209.0
91.00	2037.0
92.00	2115.0
93.00	12745.0
94.00	993.0
95.00	262.0
96.00	597.0
98.00	11021.0
99.00	8529.0
99.90	831.0
101.00	4527.0
101.80	270.0
103.00	1635.0
104.00	3019.0
105.00	3174.0
106.20	861.0
107.00	34344.0
108.00	5501.0
109.10	942.0
110.00	58976.0
111.00	8559.0
112.00	1160.0
113.00	293.0
114.70	192.0
115.10	199.0
116.00	2011.0
117.00	29176.0
118.00	2370.0
119.00	447.0
120.00	417.0
121.00	209.0
122.00	2197.0
123.00	3629.0
124.00	1746.0
125.00	1398.0
127.00	119208.0
128.00	8318.0
129.00	48280.0
130.00	4362.0
130.90	938.0
131.90	354.0
132.90	367.0
133.90	1455.0
135.00	4112.0
135.90	1563.0

Instrument :
BNA_M
ClientSampleId :
DFTPP038

137.00	1965.0
137.80	383.0
139.00	340.0
140.00	533.0
141.00	5550.0
142.00	2142.0
142.90	1443.0
144.00	343.0
145.00	355.0
146.00	1376.0
147.00	3497.0
148.00	7415.0
149.00	1885.0
149.90	390.0
151.10	722.0
151.90	362.0
153.00	2098.0
153.90	1439.0
155.00	3700.0
156.00	4478.0
157.00	828.0
157.90	1143.0
158.90	959.0
160.00	2210.0
161.00	2942.0
162.00	1116.0
162.90	336.0
164.10	271.0
164.90	2704.0
165.90	1727.0
167.00	12458.0
168.00	7349.0
169.00	1054.0
169.90	357.0
171.00	534.0
172.00	1206.0
173.00	1570.0
174.00	2822.0
175.00	4318.0
175.90	1012.0
177.00	2478.0
177.90	658.0
179.00	9100.0
180.00	7229.0
181.00	3519.0
182.10	399.0
182.80	262.0
184.00	608.0
185.00	5191.0
186.00	32232.0
187.00	10529.0
188.10	1230.0
189.00	2275.0
189.80	396.0
191.00	1329.0
192.00	3289.0
193.00	3756.0
194.00	797.0
195.00	426.0
196.00	7689.0
197.10	1614.0
197.90	230720.0
199.00	15997.0
200.00	1207.0
200.80	291.0
201.50	1047.0
203.00	1786.0
204.00	9549.0
205.00	15193.0
206.00	66552.0
207.00	10216.0
208.00	2701.0
209.00	951.0
210.00	957.0
211.00	2594.0
211.70	605.0
212.70	158.0
214.90	716.0
216.00	1678.0
217.00	19848.0
218.00	2271.0
219.00	254.0
221.00	9317.0
221.80	2507.0
223.00	3861.0
224.00	38384.0
225.00	9447.0
226.00	975.0
227.00	16311.0
228.00	2162.0
229.00	3199.0
229.90	339.0
231.00	1347.0
231.90	228.0
232.90	324.0
233.90	709.0

Instrument :
BNA_M
ClientSampleId :
DFTPP038

235.00	1176.0
235.90	840.0
237.00	1814.0
239.00	519.0
240.00	742.0
240.90	935.0
242.00	2171.0
243.10	2023.0
244.00	28464.0
245.00	3356.0
246.00	5606.0
247.10	1157.0
248.00	407.0
249.00	1143.0
250.00	189.0
250.90	389.0
251.90	290.0
252.90	564.0
254.10	1277.0
255.00	154944.0
256.00	22328.0
257.00	1695.0
258.00	9582.0
259.00	1506.0
259.80	251.0
261.10	157.0
264.00	321.0
265.00	3754.0
265.90	467.0
266.80	185.0
267.90	180.0
269.90	207.0
271.00	409.0
272.00	450.0
272.90	4682.0
274.00	12098.0
275.00	71016.0
276.00	9359.0
277.00	5613.0
278.00	1460.0
279.00	183.0
280.90	503.0
282.00	380.0
282.90	652.0
284.00	528.0
284.90	858.0
285.80	187.0
288.90	203.0
289.90	239.0
292.00	246.0
293.00	1251.0
294.00	547.0
294.90	425.0
296.00	21600.0
297.00	2678.0
298.00	300.0
301.10	280.0
302.10	312.0
303.00	2483.0
304.00	646.0
307.90	280.0
308.90	186.0
310.00	294.0
312.90	185.0
314.00	1149.0
315.00	2287.0
316.10	1204.0
316.90	285.0
321.10	591.0
322.00	361.0
323.00	6627.0
324.00	1291.0
324.80	167.0
326.00	189.0
327.00	1539.0
327.90	794.0
332.00	609.0
332.90	559.0
334.00	4818.0
335.00	932.0
340.00	214.0
341.00	809.0
342.10	179.0
346.00	1749.0
347.00	459.0
352.00	2082.0
353.00	1169.0
354.00	2303.0
355.00	444.0
355.90	165.0
359.00	156.0
365.00	10384.0
365.90	1332.0
370.00	248.0
371.10	657.0
372.00	3500.0

Instrument :
BNA_M
ClientSampleId :
DFTPP038

373.00	947.0
383.00	904.0
384.00	189.0
389.90	621.0
390.90	289.0
391.80	258.0
401.00	217.0
402.00	1576.0
403.00	1882.0
404.00	956.0
405.00	217.0
421.00	1926.0
422.00	2111.0
423.00	12229.0
424.00	4163.0
425.00	469.0
441.00	31088.0
442.10	209600.0
443.10	41104.0
444.00	3710.0
445.00	266.0

Instrument :
BNA_M
ClientSampleId :
DFTPP038

m/z	Abundance
37.00	421.0
38.00	758.0
39.10	4384.0
40.00	904.0
41.00	366.0
41.90	158.0
43.00	680.0
44.00	1954.0
45.00	488.0
49.00	706.0
50.10	13461.0
51.00	51216.0
52.00	2520.0
55.10	513.0
56.00	1459.0
57.00	3629.0
58.00	234.0
59.90	202.0
61.00	612.0
62.00	656.0
63.00	2001.0
63.90	313.0
65.00	968.0
68.00	1018.0
69.00	53184.0
70.10	211.0
73.00	1409.0
74.00	5388.0
75.00	8358.0
76.00	2665.0
77.00	56296.0
78.00	4039.0
79.00	4232.0
80.00	2911.0
81.00	4421.0
82.00	958.0
83.00	975.0
84.00	281.0
85.00	720.0
86.00	1292.0
87.10	587.0
87.90	311.0
88.90	200.0
91.00	985.0
92.00	1034.0
93.00	6965.0
94.00	512.0
94.90	239.0
95.80	551.0
96.90	322.0
98.00	5910.0
99.00	4531.0
99.90	393.0
101.00	2396.0
102.00	178.0
102.90	770.0
104.00	1699.0
105.00	1430.0
106.00	559.0
107.00	18000.0
108.00	3027.0
109.00	503.0
110.00	30520.0
111.00	4924.0
111.90	532.0
113.00	255.0
115.90	1134.0
117.00	15742.0
118.00	1063.0
119.90	228.0
120.90	167.0
122.00	1237.0
123.00	1844.0
124.00	868.0
125.00	876.0
127.00	62208.0
128.00	4101.0
129.00	25808.0
130.00	2276.0
131.00	502.0
132.00	221.0
132.90	396.0
133.90	627.0
135.00	2154.0
136.00	627.0
137.00	1128.0
137.80	201.0
140.00	401.0
141.00	3525.0
141.90	885.0
142.90	860.0
144.00	235.0

Instrument :
BNA_M
ClientSampleId :
DFTPP038

145.10	163.0
146.00	508.0
147.00	1961.0
148.00	4056.0
149.00	944.0
149.80	233.0
150.20	230.0
151.00	487.0
151.70	458.0
153.00	1073.0
153.90	1061.0
155.10	1833.0
156.00	3150.0
157.00	445.0
157.90	690.0
159.00	529.0
160.00	1044.0
161.00	1318.0
162.00	303.0
163.00	278.0
164.10	202.0
165.00	1275.0
166.10	1212.0
167.00	6531.0
168.00	3956.0
168.90	622.0
169.80	276.0
171.00	257.0
171.90	596.0
172.90	831.0
174.00	1491.0
175.00	2313.0
176.00	764.0
176.90	1208.0
177.80	496.0
178.90	5283.0
180.00	3954.0
181.00	1794.0
181.90	305.0
182.90	215.0
183.90	508.0
185.00	2412.0
186.00	18752.0
187.00	5242.0
188.00	672.0
189.00	1492.0
189.80	170.0
191.00	904.0
192.00	2130.0
193.00	2535.0
194.00	343.0
194.90	277.0
196.00	3786.0
197.10	786.0
197.90	131520.0
199.00	8397.0
200.10	728.0
201.40	539.0
203.00	1038.0
204.00	5091.0
205.00	9079.0
206.00	38704.0
207.00	6971.0
208.00	1914.0
209.00	630.0
210.20	621.0
211.00	1549.0
214.90	412.0
216.00	675.0
217.00	10756.0
217.90	1323.0
221.00	5830.0
222.00	1675.0
223.00	2571.0
224.00	21600.0
225.00	5400.0
226.00	686.0
227.00	8880.0
228.00	1141.0
229.00	1808.0
229.90	292.0
230.90	1033.0
233.00	173.0
233.90	639.0
235.00	800.0
236.00	431.0
237.00	1120.0
238.10	220.0
238.90	364.0
239.90	257.0
240.90	542.0
242.00	1065.0
243.00	1322.0
244.00	17208.0
245.00	2090.0
246.00	3577.0

Instrument :
BNA_M
ClientSampleId :
DFTPP038

247.00	626.0
249.00	727.0
251.00	183.0
252.10	184.0
253.00	394.0
254.10	775.0
255.00	98056.0
256.00	12923.0
257.00	1072.0
258.00	5095.0
259.00	900.0
259.80	178.0
261.00	211.0
265.00	2529.0
265.90	364.0
267.00	210.0
271.00	216.0
272.00	280.0
273.00	3218.0
274.00	7840.0
275.00	44288.0
276.00	5668.0
277.00	3979.0
278.00	753.0
278.90	153.0
281.10	610.0
283.00	611.0
284.00	373.0
284.90	541.0
288.80	166.0
292.90	895.0
294.00	248.0
295.00	365.0
296.00	13804.0
297.00	2019.0
302.00	193.0
303.00	1521.0
304.00	414.0
307.70	210.0
309.00	150.0
310.00	260.0
312.90	294.0
314.00	637.0
315.00	1485.0
316.00	859.0
317.10	174.0
320.90	634.0
322.00	323.0
323.00	4543.0
324.00	858.0
327.00	931.0
328.00	489.0
332.00	227.0
333.00	438.0
334.00	2553.0
335.00	756.0
341.10	692.0
342.00	224.0
346.00	1150.0
347.00	378.0
352.00	1584.0
353.00	899.0
354.00	1686.0
354.80	230.0
365.00	6685.0
366.00	1035.0
369.90	251.0
371.10	352.0
372.00	2526.0
372.90	538.0
383.00	694.0
384.10	182.0
390.00	275.0
392.00	197.0
401.00	189.0
402.00	940.0
403.00	1552.0
403.90	549.0
421.00	1308.0
422.00	1474.0
423.00	9775.0
424.00	2797.0
425.00	420.0
441.00	23936.0
442.10	158592.0
443.10	29128.0
444.00	2428.0
445.00	154.0

Instrument :
BNA_M
ClientSampleId :
DFTPP038