

Data Path : Z:\HPCHEM1\BNA_M\Data\BM051915\
 Data File : BM001344.D
 Acq On : 19 May 2015 12:26
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 20 04:10:26 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 20 03:51:55 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	-0.02
2	1,4-Dioxane	0.582	0.623	-7.0	95	0.00
3	n-Nitrosodimethylamine	0.566	0.572	-1.1	111	0.00
4 S	2-Fluorophenol	0.974	0.962	1.2	100	-0.02
5 S	Phenol-d6	1.099	1.107	-0.7	105	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
7 S	Nitrobenzene-d5	0.263	0.265	-0.8	106	0.00
8	Nitrobenzene	0.276	0.274	0.7	107	0.00
9	Naphthalene	1.064	1.078	-1.3	104	0.00
10	Hexachlorobutadiene	0.208	0.210	-1.0	106	0.00
11	2-Methylnaphthalene	0.692	0.716	-3.5	109	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
13 S	2,4,6-Tribromophenol	0.149	0.108	27.5#	90	0.00
14 S	2-Fluorobiphenyl	1.552	1.588	-2.3	111	0.00
15	Acenaphthylene	2.058	2.030	1.4	110	0.00
16	Acenaphthene	1.318	1.302	1.2	110	0.00
17	Fluorene	1.194	1.106	7.4	108	-0.03
18 I	Phenanthrene-d10	1.000	1.000	0.0	129	0.00
19	Hexachlorobenzene	0.322	0.234	27.3#	102	0.00
20	Pentachlorophenol	0.067	0.052	22.4	140	0.00
21	Phenanthrene	0.760	0.848	-11.6	157#	0.00
22	Anthracene	0.741	0.761	-2.7	145	0.00
23	Fluoranthene	1.193	0.820	31.3#	98	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	88	-0.01
25	Pyrene	1.545	1.575	-1.9	99	0.00
26 S	Terphenyl-d14	0.672	0.702	-4.5	99	-0.01
27	Benzo(a)anthracene	1.357	1.279	5.7	94	0.00
28	Chrysene	1.376	1.355	1.5	92	-0.01
29	Indeno(1,2,3-cd)pyrene	1.398	1.349	3.5	92	-0.01
30 I	Perylene-d12	1.000	1.000	0.0	88	-0.01
31	Benzo(b)fluoranthene	1.465	1.535	-4.8	100	0.00
32	Benzo(k)fluoranthene	1.394	1.235	11.4	83	-0.01
33 C	Benzo(a)pyrene	1.269	1.197	5.7	91	0.00
34	Dibenzo(a,h)anthracene	1.193	1.167	2.2	91	-0.01
35	Benzo(g,h,i)perylene	1.309	1.295	1.1	93	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0