

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM051920\
 Data File : BM026020.D
 Acq On : 19 May 2020 14:48
 Operator : MA/SJ
 Sample : L2681-04
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C5CA6

Quant Time: May 19 15:28:25 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM051320MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 19 10:05:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	156434	20.00	ng/ul	0.00
18) Naphthalene-d8	10.26	136	641946	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.15	164	366120	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.89	188	782285	20.00	ng/ul	0.00
78) Chrysene-d12	21.09	240	828000	20.00	ng/ul	0.00
86) Perylene-d12	23.22	264	866738	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.08	96	27785	5.36	ng/uL	0.00
5) Phenol-d5	6.69	99	93029	6.56	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.85	67	288452	29.38	ng/ul	0.00
9) 2-Chlorophenol-d4	7.04	132	251434	22.64	ng/ul	0.00
13) 4-Methylphenol-d8	8.21	113	175844	16.44	ng/ul	0.00
19) Nitrobenzene-d5	8.65	128	153796	29.39	ng/ul	0.00
22) 2-Nitrophenol-d4	9.36	143	157295	30.38	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.89	165	286091	27.52	ng/ul	0.00
29) 4-Chloroaniline-d4	10.41	131	19264	1.83	ng/ul	0.00
44) Dimethylphthalate-d6	13.57	166	982697	33.47	ng/ul	0.00
47) Acenaphthylene-d8	13.83	160	1125631	32.36	ng/ul	0.00
52) 4-Nitrophenol-d4	14.36	143	28248	5.43	ng/ul	0.00
58) Fluorene-d10	15.15	176	826245	32.35	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.27	200	137940	30.24	ng/ul	0.00
71) Anthracene-d10	16.99	188	1283298	33.29	ng/ul	0.00
79) Pyrene-d10	19.29	212	1463868	34.42	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.09	264	1582902	34.31	ng/ul	0.00

Target Compounds						Qvalue
28) Naphthalene	10.32	128	79967	2.04	ng/ul	99
45) Dimethylphthalate	13.62	163	153553	4.98	ng/ul	98
56) 2,3,4,6-Tetrachlorophenol	14.78	232	20645	2.91	ng/ul#	95
69) Pentachlorophenol	16.56	266	387574	68.70	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM051920\
Data File : BM026020.D
Acq On : 19 May 2020 14:48
Operator : MA/SJ
Sample : L2681-04
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C5CA6

Quant Time: May 19 15:28:25 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM051320MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 19 10:05:34 2020
Response via : Initial Calibration

