

Data Path : Z:\HPCHEM1\BNA_M\Data\BM052015\
 Data File : BM001376.D
 Acq On : 22 May 2015 01:18
 Operator : TP/IZ
 Sample : G2255-14
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 AOC50MW002-150512

Quant Time: May 22 05:33:18 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM052015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 10:49:04 2015
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Dichlorobenzene-d4	7.432	152	12209	5.00 ng	0.00
6) Naphthalene-d8	10.199	136	52588	5.00 ng	0.00
12) Acenaphthene-d10	14.107	164	31339	5.00 ng	0.00
18) Phenanthrene-d10	16.851	188	62349	5.00 ng	# 0.00
25) Chrysene-d12	21.051	240	50455	5.00 ng	#-0.02
32) Perylene-d12	23.167	264	36823	5.00 ng	0.00
System Monitoring Compounds					
4) 2-Fluorophenol	5.020	112	10424	4.40 ng	0.00
5) Phenol-d6	6.628	99	7648	2.59 ng	0.02
7) Nitrobenzene-d5	8.586	82	3931	1.39 ng	0.00
13) 2,4,6-Tribromophenol	15.595	330	2682	2.14 ng	0.00
14) 2-Fluorobiphenyl	12.719	172	58015	6.21 ng	0.00
27) Terphenyl-d14	19.503	244	71087	10.71 ng	0.00
Target Compounds					Qvalue
30) Bis(2-ethylhexyl)phtha...	20.974	149	6618	0.90 ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\Data\BM052015\
Data File : BM001376.D
Acq On : 22 May 2015 01:18
Operator : TP/IZ
Sample : G2255-14
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
AOC50MW002-150512

Quant Time: May 22 05:33:18 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM052015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 25 10:49:04 2015
Response via : Initial Calibration

