

Data Path : Z:\HPCHEM1\BNA M\DATA\BM052015\
 Data File : BM001354.D
 Acq On : 21 May 2015 10:59
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM052015

Quant Time: May 21 12:30:37 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM052015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 21 11:08:14 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	99	0.00
2	1,4-Dioxane	1.000	0.765	23.5	74	0.00
3	n-Nitrosodimethylamine	1.000	0.880	12.0	81	0.00
4 S	2-Fluorophenol	1.000	0.927	7.3	93	0.00
5 S	Phenol-d6	1.000	0.928	7.2	92	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	98	0.00
7 S	Nitrobenzene-d5	1.000	0.924	7.6	94	0.00
8	Nitrobenzene	1.000	0.876	12.4	94	0.00
9	Naphthalene	1.000	0.970	3.0	95	0.00
10	Hexachlorobutadiene	1.000	0.954	4.6	94	0.00
11	2-Methylnaphthalene	1.000	0.957	4.3	95	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	100	0.00
13 S	2,4,6-Tribromophenol	1.000	0.926	7.4	96	0.00
14 S	2-Fluorobiphenyl	1.000	0.947	5.3	95	0.00
15	Acenaphthylene	1.000	0.927	7.3	96	0.00
16	Acenaphthene	1.000	0.940	6.0	96	0.00
17	Fluorene	1.000	0.859	14.1	88	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	96	0.00
19	4-Bromophenyl-phenylether	1.000	1.082	-8.2	107	0.00
20	Hexachlorobenzene	1.000	0.933	6.7	89	0.00
21	Pentachlorophenol	1.000	0.846	15.4	90	0.00
22	Phenanthrene	1.000	0.912	8.8	86	0.00
23	Anthracene	1.000	1.020	-2.0	91	0.00
24	Fluoranthene	1.000	1.027	-2.7	97	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	102	0.00
26	Pyrene	1.000	0.942	5.8	98	0.00
27 S	Terphenyl-d14	1.000	0.956	4.4	100	0.00
28	Benzo(a)anthracene	1.000	0.936	6.4	101	0.00
29	Chrysene	1.000	0.956	4.4	101	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.821	17.9	94	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.963	3.7	106	0.00
32 I	Perylene-d12	5.000	5.000	0.0	108	0.00
33	Benzo(b)fluoranthene	1.000	0.938	6.2	105	0.00
34	Benzo(k)fluoranthene	1.000	0.942	5.8	103	0.00
35 C	Benzo(a)pyrene	1.000	0.922	7.8	104	0.00
36	Dibenzo(a,h)anthracene	1.000	0.934	6.6	106	0.00
37	Benzo(g,h,i)perylene	1.000	0.930	7.0	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0