

Data Path : Z:\HPCHEM1\BNA_M\Data\BM052015\
 Data File : BM001365.D
 Acq On : 21 May 2015 18:09
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 22 03:54:20 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM052015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 10:49:04 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.512	0.597	-16.6	101	0.00
3	n-Nitrosodimethylamine	0.564	0.607	-7.6	91	0.00
4 S	2-Fluorophenol	0.971	0.998	-2.8	94	0.00
5 S	Phenol-d6	1.208	1.257	-4.1	95	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
7 S	Nitrobenzene-d5	0.270	0.278	-3.0	96	0.00
8	Nitrobenzene	0.282	0.289	-2.5	96	0.00
9	Naphthalene	1.067	1.108	-3.8	93	0.00
10	Hexachlorobutadiene	0.206	0.216	-4.9	95	0.00
11	2-Methylnaphthalene	0.758	0.788	-4.0	94	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
13 S	2,4,6-Tribromophenol	0.175	0.198	-13.1	99	0.00
14 S	2-Fluorobiphenyl	1.491	1.559	-4.6	95	0.00
15	Acenaphthylene	2.104	2.188	-4.0	97	0.00
16	Acenaphthene	1.321	1.351	-2.3	95	0.00
17	Fluorene	1.057	1.037	1.9	91	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
19	4-Bromophenyl-phenylether	0.251	0.322	-28.3#	106	0.00
20	Hexachlorobenzene	0.369	0.376	-1.9	88	0.00
21	Pentachlorophenol	0.133	0.133	0.0	109	0.00
22	Phenanthrene	1.758	1.830	-4.1	89	0.00
23	Anthracene	0.680	0.799	-17.5	95	0.00
24	Fluoranthene	1.506	1.740	-15.5	99	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
26	Pyrene	1.528	1.546	-1.2	98	0.00
27 S	Terphenyl-d14	0.658	0.680	-3.3	101	0.00
28	Benzo(a)anthracene	1.430	1.478	-3.4	105	0.00
29	Chrysene	1.369	1.430	-4.5	103	0.00
30	Bis(2-ethylhexyl)phthalate	0.705	0.669	5.1	102	0.00
31	Indeno(1,2,3-cd)pyrene	1.475	1.584	-7.4	111	0.00
32 I	Perylene-d12	1.000	1.000	0.0	103	0.00
33	Benzo(b)fluoranthene	1.541	1.597	-3.6	110	0.00
34	Benzo(k)fluoranthene	1.357	1.357	0.0	104	0.00
35 C	Benzo(a)pyrene	1.311	1.339	-2.1	110	0.00
36	Dibenzo(a,h)anthracene	1.204	1.236	-2.7	111	0.00
37	Benzo(g,h,i)perylene	1.293	1.323	-2.3	111	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0