

Data Path : Z:\HPCHEM1\BNA_M\Data\BM052015\
 Data File : BM001365.D
 Acq On : 21 May 2015 18:09
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 22 03:54:20 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM052015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 10:49:04 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	90	0.00
2	1,4-Dioxane	1.000	1.323	-32.3#	101	0.00
3	n-Nitrosodimethylamine	1.000	1.076	-7.6	91	0.00
4 S	2-Fluorophenol	1.000	1.028	-2.8	94	0.00
5 S	Phenol-d6	1.000	1.040	-4.0	95	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	90	0.00
7 S	Nitrobenzene-d5	1.000	1.031	-3.1	96	0.00
8	Nitrobenzene	1.000	0.959	4.1	96	0.00
9	Naphthalene	1.000	1.039	-3.9	93	0.00
10	Hexachlorobutadiene	1.000	1.046	-4.6	95	0.00
11	2-Methylnaphthalene	1.000	1.039	-3.9	94	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	91	0.00
13 S	2,4,6-Tribromophenol	1.000	1.041	-4.1	99	0.00
14 S	2-Fluorobiphenyl	1.000	1.045	-4.5	95	0.00
15	Acenaphthylene	1.000	1.040	-4.0	97	0.00
16	Acenaphthene	1.000	1.023	-2.3	95	0.00
17	Fluorene	1.000	0.981	1.9	91	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	86	0.00
19	4-Bromophenyl-phenylether	1.000	1.181	-18.1	106	0.00
20	Hexachlorobenzene	1.000	1.019	-1.9	88	0.00
21	Pentachlorophenol	1.000	1.002	-0.2	109	0.00
22	Phenanthrene	1.000	1.041	-4.1	89	0.00
23	Anthracene	1.000	1.174	-17.4	95	0.00
24	Fluoranthene	1.000	1.155	-15.5	99	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	96	0.00
26	Pyrene	1.000	1.012	-1.2	98	0.00
27 S	Terphenyl-d14	1.000	1.034	-3.4	101	0.00
28	Benzo(a)anthracene	1.000	1.033	-3.3	105	0.00
29	Chrysene	1.000	1.044	-4.4	103	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.918	8.2	102	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.074	-7.4	111	0.00
32 I	Perylene-d12	5.000	5.000	0.0	103	0.00
33	Benzo(b)fluoranthene	1.000	1.036	-3.6	110	0.00
34	Benzo(k)fluoranthene	1.000	1.000	0.0	104	0.00
35 C	Benzo(a)pyrene	1.000	1.022	-2.2	110	0.00
36	Dibenzo(a,h)anthracene	1.000	1.026	-2.6	111	0.00
37	Benzo(g,h,i)perylene	1.000	1.023	-2.3	111	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0