

Data Path : Z:\HPCHEM1\BNA M\DATA\BM052015\
 Data File : BM001349.D
 Acq On : 21 May 2015 07:24
 Operator : TP/IZ
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.8

Manual Integrations
 APPROVED

apatel
 5/21/2015 8:53:05 PM

Quant Time: May 21 09:04:19 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM052015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 21 08:43:35 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.43	152	14942	5.00	ng	0.00
6) Naphthalene-d8	10.20	136	60277	5.00	ng	0.00
12) Acenaphthene-d10	14.11	164	37961	5.00	ng	0.00
18) Phenanthrene-d10	16.85	188	85109	5.00	ng	0.00
25) Chrysene-d12	21.07	240	85438	5.00	ng	0.00
32) Perylene-d12	23.17	264	80061	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.02	112	2279	0.79	ng	0.00
5) Phenol-d6	6.61	99	2775	0.75	ng	0.00
7) Nitrobenzene-d5	8.59	82	2469	0.71	ng	0.00
13) 2,4,6-Tribromophenol	15.59	330	931	0.79	ng	0.00
14) 2-Fluorobiphenyl	12.72	172	8963	0.81	ng	0.00
27) Terphenyl-d14	19.50	244	8824	0.81	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.90	88	1040	0.77	ng	# 85
3) n-Nitrosodimethylamine	3.23	42	1311	0.73	ng	# 90
8) Nitrobenzene	8.63	77	2603	0.71	ng	100
9) Naphthalene	10.25	128	10132	0.81	ng	99
10) Hexachlorobutadiene	10.54	225	1967	0.82	ng	99
11) 2-Methylnaphthalene	11.89	142	7228	0.81	ng	99
15) Acenaphthylene	13.82	152	12588	0.77	ng	100
16) Acenaphthene	14.17	154	7921	0.81	ng	100
17) Fluorene	15.15	166	6601	0.69	ng	92
19) 4-Bromophenyl-phenylether	16.04	248	3018	0.98	ng	92
20) Hexachlorobenzene	16.17	284	4777	0.85	ng	# 96
21) Pentachlorophenol	16.51	266	1237	0.62	ng	99
22) Phenanthrene	16.89	178	22983m	0.78	ng	
23) Anthracene	16.99	178	8829m	0.84	ng	
24) Fluoranthene	18.92	202	19037	0.87	ng	100
26) Pyrene	19.29	202	20371	0.81	ng	100
28) Benzo(a)anthracene	21.04	228	18843	0.81	ng	98
29) Chrysene	21.10	228	18216	0.81	ng	99
30) Bis(2-ethylhexyl)phthalate	20.99	149	8223	0.63	ng	100
31) Indeno(1,2,3-cd)pyrene	25.24	276	19517	0.78	ng	100
33) Benzo(b)fluoranthene	22.54	252	19737	0.85	ng	99
34) Benzo(k)fluoranthene	22.58	252	16725	0.75	ng	100
35) Benzo(a)pyrene	23.07	252	16299	0.78	ng	100
36) Dibenzo(a,h)anthracene	25.25	278	15041	0.77	ng	100
37) Benzo(g,h,i)perylene	25.88	276	16189	0.79	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA M\DATA\BM052015\
Data File : BM001349.D
Acq On : 21 May 2015 07:24
Operator : TP/IZ
Sample : SSTDIC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8

Manual Integrations
APPROVED

apatel
5/21/2015 8:53:05 PM

Quant Time: May 21 09:04:19 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM052015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 21 08:43:35 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.43 min Scan# 308

Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Instrument :

BNA_M





ClientSampleId :

SSTDIC0.8

Tot Ion	Ratio	Lower	Upper
152	100		
150	143.9	114.1	171.1
115	49.6	39.2	58.8

Manual Integrations
APPROVED

apatel
5/21/2015 8:53:05 PM



#2

1,4-Dioxane

Concen: 0.77 ng




RT: 2.90 min Scan# 15


Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Tgt Ion	Ratio	Lower	Upper
88	100		
58	78.7	53.1	79.7
43	43.3	27.8	41.8





#3





n-Nitrosodimethylamine
Concen: 0.73 ng
RT: 3.23 min Scan# 36
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8

Tot Ion	Ratio	Lower	Upper
42	100		
74	96.2	70.0	105.0
44	8.5	3.6	5.4

Manual Integrations
APPROVED




apatel
5/21/2015 8:53:05 PM




#4

2-Fluorophenol
Concen: 0.79 ng
RT: 5.02 min Scan# 152
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Tgt Ion	Ratio	Lower	Upper
112	100		
64	64.3	52.9	79.3
63	36.1	29.6	44.4





#5

Phenol-d6

Concen: 0.75 ng

RT: 6.61 min Scan# 255

Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Instrument :

BNA_M





ClientSampleId :

SSTDICC0.8

Tot Ion:	99	Resp:	2775
Ion	Ratio	Lower	Upper
99	100		
42	26.5	21.4	32.2
71	36.6	29.0	43.4

Manual Integrations
APPROVED

apatel
5/21/2015 8:53:05 PM



#6

Naphthalene-d8

Concen: 5.00 ng




RT: 10.20 min Scan# 536


Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Tgt Ion:	136	Resp:	60277
Ion	Ratio	Lower	Upper
136	100		
137	10.2	8.2	12.4
54	12.5	9.4	14.0
68	6.8	5.2	7.8





#7

Nitrobenzene-d5

Concen: 0.71 ng

RT: 8.59 min Scan# 396

Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Instrument :

BNA_M





ClientSampleId :

SSTDIC0.8

Tot Ion:	82	Resp:	2469
Ion	Ratio	Lower	Upper
82	100		
128	41.3	33.5	50.3
54	59.3	45.1	67.7

Manual Integrations
APPROVED

apatel
5/21/2015 8:53:05 PM



#8

Nitrobenzene

Concen: 0.71 ng




RT: 8.63 min Scan# 400


Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Tgt Ion:	77	Resp:	2603
Ion	Ratio	Lower	Upper
77	100		
123	42.1	33.6	50.4
65	13.8	11.3	16.9





#9

Naphthalene

Concen: 0.81 ng

RT: 10.25 min Scan# 539

Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Instrument :

BNA_M





ClientSampleId :

SSTDIC0.8

Tot Ion:128	Resp:	10132
Ion Ratio	Lower	Upper
128 100		
129 11.4	9.1	13.7
127 13.4	10.4	15.6

Manual Integrations
APPROVED

apatel
5/21/2015 8:53:05 PM



#10

Hexachlorobutadiene

Concen: 0.82 ng




RT: 10.54 min Scan# 556


Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Tgt Ion:225	Resp:	1967
Ion Ratio	Lower	Upper
225 100		
223 62.6	50.3	75.5
227 63.9	50.6	75.8









#11
2-Methylnaphthalene
Concen: 0.81 ng
RT: 11.89 min Scan# 656
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8

Tot Ion	Ratio	Lower	Upper
142	100		
141	85.3	69.0	103.4
115	34.6	27.1	40.7




Manual Integrations
APPROVED




apatel
5/21/2015 8:53:05 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.11 min Scan# 860
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Tgt Ion	Ratio	Lower	Upper
164	100		
162	104.3	81.8	122.8
160	46.0	35.8	53.8









#13
2,4,6-Tribromophenol
Concen: 0.79 ng
RT: 15.59 min Scan# 936
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Tot Ion	Ratio	Lower	Upper
330	100		
332	82.9	65.4	98.2
141	30.0	20.2	30.4

Instrument :
BNA_M
ClientSampleId :
SSTDICC0.8


Manual Integrations
APPROVED


apatel
5/21/2015 8:53:05 PM



#14
2-Fluorobiphenyl
Concen: 0.81 ng
RT: 12.72 min Scan# 736
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Tgt Ion	Ratio	Lower	Upper
172	100		
171	32.1	25.4	38.2
170	20.0	15.7	23.5





#15

Acenaphthylene

Concen: 0.77 ng

RT: 13.82 min Scan# 836

Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Instrument :

BNA_M





ClientSampleId :

SSTDIC0.8

Tot Ion:	152	Resp:	12588
Ion Ratio	Lower	Upper	
152	100		
151	18.8	15.3	22.9
153	12.9	10.3	15.5

Manual Integrations
APPROVED

apatel
5/21/2015 8:53:05 PM



#16

Acenaphthene

Concen: 0.81 ng




RT: 14.17 min Scan# 865


Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Tgt Ion:	154	Resp:	7921
Ion Ratio	Lower	Upper	
154	100		
153	113.1	90.6	135.8
152	54.2	43.6	65.4





#17

Fluorene

Concen: 0.69 ng

RT: 15.15 min Scan# 923

Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Instrument :

BNA_M





ClientSampleId :

SSTDIC0.8

Tot Ion:	166	Resp:	6601
Ion Ratio	Lower	Upper	
166	100		
165	82.2	72.5	108.7
167	17.7	16.0	24.0

Manual Integrations
APPROVED

apatel
5/21/2015 8:53:05 PM



#18

Phenanthrene-d10

Concen: 5.00 ng




RT: 16.85 min Scan# 973


Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Tgt Ion:	188	Resp:	85109
Ion Ratio	Lower	Upper	
188	100		
94	0.8	0.6	1.0
80	0.7	0.6	0.8









#19
4-Bromophenyl-phenylether
Concen: 0.98 ng
RT: 16.04 min Scan# 949
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8

Tot Ion	Ratio	Lower	Upper
248	100		
250	63.8	51.5	77.3
141	85.9	59.0	88.6




Manual Integrations
APPROVED




apatel
5/21/2015 8:53:05 PM



#20
Hexachlorobenzene
Concen: 0.85 ng
RT: 16.17 min Scan# 953
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Tgt Ion	Ratio	Lower	Upper
284	100		
142	0.0	0.0	0.0
249	15.4	11.1	16.7









#21
 Pentachlorophenol
 Concen: 0.62 ng
 RT: 16.51 min Scan# 963
 Delta R.T. -0.00 min
 Lab File: BM001349.D
 Acq: 21 May 2015 07:24

Tot Ion:	266	Resp:	1237
Ion Ratio	Lower	Upper	
266	100		
268	63.8	51.2	76.8
264	67.8	53.7	80.5

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.8


Manual Integrations
 APPROVED


apatel
 5/21/2015 8:53:05 PM



#22
 Phenanthrene
 Concen: 0.78 ng m
 RT: 16.89 min Scan# 974
 Delta R.T. -0.00 min
 Lab File: BM001349.D
 Acq: 21 May 2015 07:24

Tgt Ion:	178	Resp:	22983
Ion Ratio	Lower	Upper	
178	100		
176	23.3	16.6	25.0
179	17.9	12.8	19.2





#23

Anthracene

Concen: 0.84 ng m

RT: 16.99 min Scan# 977

Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Instrument :

BNA_M




ClientSampleId :

SSTDIC0.8

Tot Ion:	178	Resp:	8829
Ion Ratio	Lower	Upper	
178	100		
176	60.5	46.1	69.1
179	46.7	35.4	53.0

Manual Integrations
APPROVED

apatel
5/21/2015 8:53:05 PM



#24

Fluoranthene

Concen: 0.87 ng




RT: 18.92 min Scan# 1070


Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Tgt Ion:	202	Resp:	19037
Ion Ratio	Lower	Upper	
202	100		
101	10.9	8.7	13.1
203	17.4	13.8	20.6









#25
Chrysene-d12
Concen: 5.00 ng
RT: 21.07 min Scan# 1238
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8

Tot Ion	Ratio	Lower	Upper
240	100		
120	1.7	1.6	2.4
236	23.2	18.5	27.7




Manual Integrations
APPROVED


apatel
5/21/2015 8:53:05 PM



#26
Pyrene
Concen: 0.81 ng
RT: 19.29 min Scan# 1100
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.4	16.2	24.4
203	17.1	13.9	20.9





#27

Terphenyl-d14

Concen: 0.81 ng

RT: 19.50 min Scan# 1118

Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Instrument :

BNA_M





ClientSampleId :

SSTDIC0.8

Tot Ion:	244	Resp:	8824
Ion Ratio	Lower	Upper	
244	100		
212	10.6	8.5	12.7
122	16.8	13.4	20.2

Manual Integrations
APPROVED

apatel
5/21/2015 8:53:05 PM



#28

Benzo(a)anthracene

Concen: 0.81 ng




RT: 21.04 min Scan# 1236


Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Tgt Ion:	228	Resp:	18843
Ion Ratio	Lower	Upper	
228	100		
226	29.3	24.5	36.7
229	17.4	13.3	19.9





#29

Chrysene

Concen: 0.81 ng

RT: 21.10 min Scan# 1240

Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24

Instrument :

BNA_M


ClientSampleId :

SSTDICC0.8

Tot Ion:	228	Resp:	18216
Ion Ratio	Lower	Upper	
228	100		
226	24.4	20.0	30.0
229	22.5	17.6	26.4

Manual Integrations
APPROVED




apatel
5/21/2015 8:53:05 PM



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#30

Bis(2-ethylhexyl)phthalate

Concen: 0.63 ng


RT: 20.99 min Scan# 1233

Delta R.T. -0.00 min

Lab File: BM001349.D

Acq: 21 May 2015 07:24



Tgt Ion:	149	Resp:	8223
Ion Ratio	Lower	Upper	
149	100		
167	26.9	21.6	32.4
279	2.4	1.9	2.9




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E









#31
Indeno(1,2,3-cd)pyrene
Concen: 0.78 ng
RT: 25.24 min Scan# 1609
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8

Tot Ion	Ratio	Lower	Upper
276	100		
138	30.2	24.2	36.4
277	24.5	19.7	29.5




Manual Integrations
APPROVED


apatel
5/21/2015 8:53:05 PM



#32
Perylene-d12
Concen: 5.00 ng
RT: 23.17 min Scan# 1410
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.4	20.4	30.6
265	21.3	16.6	24.8









#33
Benzo(b)fluoranthene
Concen: 0.85 ng
RT: 22.54 min Scan# 1350
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.1	17.9	26.9
125	11.7	9.8	14.8




Manual Integrations
APPROVED


apatel
5/21/2015 8:53:05 PM



#34
Benzo(k)fluoranthene
Concen: 0.75 ng
RT: 22.58 min Scan# 1354
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.5	18.6	28.0
125	11.4	9.3	13.9









#35
Benzo(a)pyrene
Concen: 0.78 ng
RT: 23.07 min Scan# 1401
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.4	18.9	28.3
125	13.3	10.6	15.8




Manual Integrations
APPROVED


apatel
5/21/2015 8:53:05 PM



#36
Dibenzo(a,h)anthracene
Concen: 0.77 ng
RT: 25.25 min Scan# 1610
Delta R.T. -0.00 min
Lab File: BM001349.D
Acq: 21 May 2015 07:24

Tgt Ion	Ratio	Lower	Upper
278	100		
139	21.7	17.5	26.3
279	23.6	18.7	28.1





#37
 Benzo(a,h,i)perylene
 Concen: 0.79 ng
 RT: 25.88 min Scan# 1670
 Delta R.T. -0.00 min
 Lab File: BM001349.D
 Acq: 21 May 2015 07:24

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.8

Tot Ion	Ratio	Lower	Upper
276	100		
277	24.6	19.2	28.8
138	25.6	20.2	30.4

Manual Integrations
 APPROVED

apatel
 5/21/2015 8:53:05 PM

