

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM052020\
 Data File : BM026046.D
 Acq On : 20 May 2020 19:25
 Operator : MA/SJ
 Sample : L2681-02DL2 50X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C5CA4DL2

Quant Time: May 20 19:57:40 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM051320MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 20 11:35:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	143668	20.00	ng/ul	0.00
18) Naphthalene-d8	10.26	136	604397	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.14	164	370534	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.89	188	786854	20.00	ng/ul	0.00
78) Chrysene-d12	21.09	240	806286	20.00	ng/ul	0.00
86) Perylene-d12	23.21	264	824098	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0d	0.00	ng/uL	
5) Phenol-d5	6.69	99	1381	0.11	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.86	67	5615	0.62	ng/ul	0.00
9) 2-Chlorophenol-d4	7.04	132	3918	0.38	ng/ul	0.00
13) 4-Methylphenol-d8	8.22	113	2565	0.26	ng/ul	0.01
19) Nitrobenzene-d5	8.65	128	2789	0.57	ng/ul	0.00
22) 2-Nitrophenol-d4	9.36	143	1852	0.38	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.90	165	3848	0.39	ng/ul	0.01
29) 4-Chloroaniline-d4	0.00	131	0d	0.00	ng/ul	
44) Dimethylphthalate-d6	13.57	166	19912	0.67	ng/ul	0.00
47) Acenaphthylene-d8	13.83	160	19288	0.55	ng/ul	0.00
52) 4-Nitrophenol-d4	0.00	143	0d	0.00	ng/ul	
58) Fluorene-d10	15.14	176	18154	0.70	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	0.00	200	0d	0.00	ng/ul	
71) Anthracene-d10	16.99	188	27003	0.70	ng/ul	0.00
79) Pyrene-d10	19.29	212	29047	0.70	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.08	264	27497	0.63	ng/ul	0.00

Target Compounds

					Ovalue	
28) Naphthalene	10.31	128	708214	19.165	ng/ul	99
34) 2-Methylnaphthalene	11.93	142	95674	3.849	ng/ul	99
56) 2,3,4,6-Tetrachlorophenol	14.78	232	28031	3.909	ng/ul#	95
69) Pentachlorophenol	16.55	266	520830	91.790	ng/ul	97
70) Phenanthrene	16.93	178	79882	1.612	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM052020\
Data File : BM026046.D
Acq On : 20 May 2020 19:25
Operator : MA/SJ
Sample : L2681-02DL2 50X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C5CA4DL2

Quant Time: May 20 19:57:40 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM051320MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed May 20 11:35:33 2020
Response via : Initial Calibration

