

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM052021\
 Data File : BM030125.D
 Acq On : 21 May 2021 22:42
 Operator : CG/JU
 Sample : M2474-09
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 GB408

Manual Integrations
 APPROVED

mohammad
 5/24/2021 1:53:01 PM

Quant Time: May 22 00:37:09 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 21 23:57:54 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.51	152	126331	20.00	ng/ul	0.00
20) Naphthalene-d8	10.28	136	560217	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.16	164	374852	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.91	188	763514	20.00	ng/ul	0.00
79) Chrysene-d12	21.10	240	672296	20.00	ng/ul	0.00
88) Perylene-d12	23.24	264	615087	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.03	96	9182	2.90	ng/uL	0.00
4) Pyridine-d5	0.00	84	0d	0.00	ng/ul	0.00
7) Phenol-d5	6.71	99	66643	5.67	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	6.86	67	222760	30.58	ng/ul	0.00
11) 2-Chlorophenol-d4	7.05	132	227704	25.11	ng/ul	0.00
15) 4-Methylphenol-d8	8.23	113	142985	14.72	ng/ul	0.00
21) Nitrobenzene-d5	8.67	128	138873	35.97	ng/ul	0.00
24) 2-Nitrophenol-d4	9.38	143	153761	35.29	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.92	165	241184	27.60	ng/ul	0.00
31) 4-Chloroaniline-d4	10.49	131	11217m	0.84	ng/ul	0.05
46) Dimethylphthalate-d6	13.59	166	1049520	36.12	ng/ul	0.00
49) Acenaphthylene-d8	13.85	160	1244971	33.81	ng/ul	0.00
54) 4-Nitrophenol-d4	14.49	143	14107m	2.98	ng/ul	0.09
60) Fluorene-d10	15.16	176	927208	37.59	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.30	200	181201	36.54	ng/ul	0.00
73) Anthracene-d10	17.01	188	1590709	41.37	ng/ul	0.00
81) Pyrene-d10	19.31	212	1763243	51.13	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.11	264	1453277	41.68	ng/ul	0.00
Target Compounds						
47) Dimethylphthalate	13.63	163	72553	2.494	ng/ul	Ovalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM052021\
Data File : BM030125.D
Acq On : 21 May 2021 22:42
Operator : CG/JU
Sample : M2474-09
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
GB408

Manual Integrations
APPROVED

mohammad
5/24/2021 1:53:01 PM

Quant Time: May 22 00:37:09 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050721.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 21 23:57:54 2021
Response via : Initial Calibration

