

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052116\
 Data File : BM005554.D
 Acq On : 21 May 2016 17:20
 Operator : UM/SJ
 Sample : H3058-10
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BD3E7

Quant Time: May 22 04:57:18 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM052016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun May 22 04:56:37 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	67216	20.00	ng/ul	0.00
18) Naphthalene-d8	10.61	136	302240	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.45	164	186591	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.19	188	448388	20.00	ng/ul	0.00
75) Chrysene-d12	21.36	240	520922	20.00	ng/ul	0.00
83) Perylene-d12	23.64	264	479975	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.31	96	2572	1.68	ng/uL	0.00
5) Phenol-d5	6.98	99	45947	8.34	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.15	67	95061	29.77	ng/ul	0.00
9) 2-Chlorophenol-d4	7.35	132	111836	24.33	ng/ul	0.00
13) 4-Methylphenol-d8	8.52	113	78090	17.30	ng/ul	0.00
19) Nitrobenzene-d5	8.97	128	62828	27.19	ng/ul	0.00
22) 2-Nitrophenol-d4	9.69	143	69483	27.32	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.23	165	120822	25.22	ng/ul	0.00
29) 4-Chloroaniline-d4	10.75	131	9240	1.93	ng/ul	0.00
43) Dimethylphthalate-d6	13.86	166	480564	31.55	ng/ul	0.00
46) Acenaphthylene-d8	14.14	160	551240	28.95	ng/ul	0.00
51) 4-Nitrophenol-d4	14.64	143	22705	7.88	ng/ul	0.00
57) Fluorene-d10	15.44	176	436637	32.95	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.56	200	73827	28.63	ng/ul	0.00
70) Anthracene-d10	17.29	188	771235	36.11	ng/ul	0.00
76) Pyrene-d10	19.58	212	894442	37.92	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.50	264	817023	36.45	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052116\
Data File : BM005554.D
Acq On : 21 May 2016 17:20
Operator : UM/SJ
Sample : H3058-10
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BD3E7

Quant Time: May 22 04:57:18 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM052016.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun May 22 04:56:37 2016
Response via : Initial Calibration

