

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052116\
 Data File : BM005567.D
 Acq On : 22 May 2016 01:12
 Operator : UM/SJ
 Sample : H2922-02MS 2X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E5NG0MS

Quant Time: May 22 05:34:22 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM052016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun May 22 05:33:20 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	76198	20.00	ng/ul	0.00
18) Naphthalene-d8	10.61	136	312783	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.46	164	154765	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.21	188	263903	20.00	ng/ul	0.02
75) Chrysene-d12	21.41	240	233320	20.00	ng/ul	0.04
83) Perylene-d12	23.71	264	309193	20.00	ng/ul	0.08

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.31	96	892	0.51	ng/uL	0.00
5) Phenol-d5	6.99	99	16490	2.64	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.16	67	16584	4.58	ng/ul	0.00
9) 2-Chlorophenol-d4	7.36	132	12835	2.46	ng/ul	0.00
13) 4-Methylphenol-d8	8.52	113	5539	1.08	ng/ul	0.00
19) Nitrobenzene-d5	8.97	128	12101	5.06	ng/ul	0.00
22) 2-Nitrophenol-d4	9.69	143	5056	1.92	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.23	165	12022	2.42	ng/ul	0.00
29) 4-Chloroaniline-d4	10.75	131	29989	6.04	ng/ul	0.00
43) Dimethylphthalate-d6	13.86	166	96091	7.61	ng/ul	0.00
46) Acenaphthylene-d8	14.14	160	128547	8.14	ng/ul	0.00
51) 4-Nitrophenol-d4	14.64	143	366	0.15	ng/ul	0.00
57) Fluorene-d10	15.45	176	86451	7.86	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.56	200	10190	6.71	ng/ul	0.00
70) Anthracene-d10	17.30	188	101717	8.09	ng/ul	0.01
76) Pyrene-d10	19.62	212	111088	10.51	ng/ul	0.04
87) Benzo(a)pyrene-d12	23.57	264	90772	6.29	ng/ul	0.07

Target Compounds

						Qvalue
6) Phenol	7.01	94	15756	2.45	ng/ul	96
10) 2-Chlorophenol	7.39	128	11530	2.21	ng/ul	97
15) N-Nitroso-di-n-propylamine	8.62	70	20178	4.96	ng/ul	95
28) Naphthalene	10.66	128	26295	1.67	ng/ul#	96
33) 4-Chloro-3-methylphenol	11.89	107	14743	2.75	ng/ul	95
34) 2-Methylnaphthalene	12.27	142	57642	5.01	ng/ul	97
40) 1,1'-Biphenyl	13.29	154	65361	5.03	ng/ul	96
44) Dimethylphthalate	13.91	163	21547	1.73	ng/ul#	81
47) Acenaphthylene	14.17	152	19806	1.23	ng/ul#	14
48) 3-Nitroaniline	14.36	138	3957	1.69	ng/ul#	48
49) Acenaphthene	14.52	153	119487	11.29	ng/ul	97
52) 4-Nitrophenol	14.66	109	9103	3.79	ng/ul#	31
53) Dibenzofuran	14.86	168	50198	3.32	ng/ul#	48
54) 2,4-Dinitrotoluene	14.82	165	25443	6.91	ng/ul#	90
58) Fluorene	15.50	166	47160	3.90	ng/ul#	61
63) 4,6-Dinitro-2-methylphenol	15.55	198	2183	1.37	ng/ul#	1
64) N-Nitrosodiphenylamine	15.72	169	120174	15.01	ng/ul#	51
69) Phenanthrene	17.25	178	620279	42.06	ng/ul#	92
71) Anthracene	17.34	178	93477	6.22	ng/ul#	61
73) Di-n-butylphthalate	18.16	149	19991	1.26	ng/ul#	7
74) Fluoranthene	19.27	202	342858	20.65	ng/ul#	93
77) Pyrene	19.65	202	453707	33.83	ng/ul	96
79) 3,3'-Dichlorobenzidine	21.34	252	6516	1.54	ng/ul#	1

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052116\
Data File : BM005567.D
Acq On : 22 May 2016 01:12
Operator : UM/SJ
Sample : H2922-02MS 2X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E5NG0MS

Quant Time: May 22 05:34:22 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM052016.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun May 22 05:33:20 2016
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
80) Benzo(a)anthracene	21.40	228	64718	4.72	ng/ul#	14
81) Bis(2-ethylhexyl)phthalate	21.32	149	19817	2.48	ng/ul#	8
82) Chrysene	21.44	228	143434	10.99	ng/ul#	49
85) Benzo(b)fluoranthene	23.02	252	80655	4.42	ng/ul#	1
86) Benzo(k)fluoranthene	23.02	252	80655	4.67	ng/ul#	1
88) Benzo(a)pyrene	23.61	252	20506	1.16	ng/ul#	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052116\
Data File : BM005567.D
Acq On : 22 May 2016 01:12
Operator : UM/SJ
Sample : H2922-02MS 2X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleID :
E5NG0MS

Quant Time: May 22 05:34:22 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM052016.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun May 22 05:33:20 2016
Response via : Initial Calibration





#6

Phenol

Concen: 2.45 ng/ul

RT: 7.01 min Scan# 752

Delta R.T. -0.00 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12

Instrument :

BNA_M

ClientSampled :

E5NG0MS


Tgt Ion: 94 Resp: 15756

Ion Ratio Lower Upper

94 100

65 29.1 24.2 36.2




66 38.8 33.9 50.9



Abundance Ion 94.00 (93.70 to 94.70): BM005567.D (-718) (-)

Ion 65.00 (64.70 to 65.70): BM005567.D (-718) (-)

Ion 66.00 (65.70 to 66.70): BM005567.D (-718) (-)



#10

2-Chlorophenol

Concen: 2.21 ng/ul

RT: 7.39 min Scan# 816

Delta R.T. -0.00 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12


Tgt Ion: 128 Resp: 11530

Ion Ratio Lower Upper

128 100

64 47.6 37.3 55.9



130 30.8 26.6 40.0




Abundance Ion 128.00 (127.70 to 128.70): BM005567.D (-782) (-)

Ion 64.00 (63.70 to 64.70): BM005567.D (-782) (-)

Ion 130.00 (129.70 to 130.70): BM005567.D (-782) (-)





#15

N-Nitroso-di-n-propylamine

Concen: 4.96 ng/ul

RT: 8.62 min Scan# 1025

Delta R.T. -0.01 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12

Instrument :

BNA_M

ClientSampleId :

E5NG0MS

Tgt Ion: 70 Resp: 20178


Ion Ratio Lower Upper

70 100

42 52.6 46.1 69.1

101 9.7 8.0 12.0

130 22.0 16.8 25.2






Abundance Ion 70.00 (69.70 to 70.70): BM005567.D (-993) (-)

Ion 42.00 (41.70 to 42.70): BM005567.D (-993) (-)

Ion 101.00 (100.70 to 101.70): BM005567.D (-993) (-)

Ion 130.00 (129.70 to 130.70): BM005567.D (-993) (-)



#28

Naphthalene

Concen: 1.67 ng/ul

RT: 10.66 min Scan# 1373

Delta R.T. -0.00 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12


Tgt Ion: 128 Resp: 26295

Ion Ratio Lower Upper

128 100

129 13.7 8.6 13.0#



127 13.7 10.6 15.8




Abundance Ion 128.00 (127.70 to 128.70): BM005567.D (-1339) (-)

Ion 129.00 (128.70 to 129.70): BM005567.D (-1339) (-)

Ion 127.00 (126.70 to 127.70): BM005567.D (-1339) (-)









#33
 4-Chloro-3-methylphenol
 Concen: 2.75 ng/ul
 RT: 11.89 min Scan# 1581
 Delta R.T. -0.01 min
 Lab File: BM005567.D
 Acq: 22 May 2016 01:12

Instrument :
 BNA_M
 ClientSampleId :
 E5NG0MS

Tgt Ion:107 Resp: 14743
 Ion Ratio Lower Upper
 107 100
 144 25.3 20.6 31.0
 142 74.2 63.5 95.3




Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E






#34
 2-Methylnaphthalene
 Concen: 5.01 ng/ul
 RT: 12.27 min Scan# 1647
 Delta R.T. 0.01 min
 Lab File: BM005567.D
 Acq: 22 May 2016 01:12

Tgt Ion:142 Resp: 57642
 Ion Ratio Lower Upper
 142 100
 141 85.9 70.6 106.0



Abundance Ion 142.00 (141.70 to 142.70): E
 Ion 141.00 (140.70 to 141.70): E









#40
 1,1'-Biphenyl
 Concen: 5.03 ng/ul
 RT: 13.29 min Scan# 1819
 Delta R.T. -0.00 min
 Lab File: BM005567.D
 Acq: 22 May 2016 01:12




Instrument :
 BNA_M
 ClientSampleId :
 E5NG0MS


Tgt Ion:	154	Resp:	65361
Ion	Ratio	Lower	Upper
154	100		
153	37.0	31.7	47.5
76	11.9	10.1	15.1



#44
 Dimethylphthalate
 Concen: 1.73 ng/ul
 RT: 13.91 min Scan# 1925
 Delta R.T. -0.00 min
 Lab File: BM005567.D
 Acq: 22 May 2016 01:12

Tgt Ion:	163	Resp:	21547
Ion	Ratio	Lower	Upper
163	100		
194	14.5	3.4	5.0#
164	15.1	8.0	12.0#





#47

Acenaphthylene

Concen: 1.23 ng/ul

RT: 14.17 min Scan# 1970

Delta R.T. -0.00 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12

Instrument :

BNA_M

ClientSampleId :

E5NG0MS


Tgt Ion:152 Resp: 19806

Ion Ratio Lower Upper

152 100

151 50.2 15.7 23.5#




153 59.0 10.5 15.7#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#48

3-Nitroaniline

Concen: 1.69 ng/ul

RT: 14.36 min Scan# 2001

Delta R.T. -0.00 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12


Tgt Ion:138 Resp: 3957

Ion Ratio Lower Upper

138 100

108 71.7 9.1 13.7#



92 70.7 92.8 139.2#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BM





#49

Acenaphthene

Concen: 11.29 ng/ul

RT: 14.52 min Scan# 2029

Delta R.T. 0.01 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12

Instrument :

BNA_M

ClientSampleId :

E5NG0MS


Tgt Ion:153 Resp: 119487

Ion Ratio Lower Upper

153 100

152 48.7 37.8 56.6




154 86.3 71.1 106.7



Abundance Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E



#52

4-Nitrophenol

Concen: 3.79 ng/ul

RT: 14.66 min Scan# 2053

Delta R.T. -0.00 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12


Tgt Ion:109 Resp: 9103

Ion Ratio Lower Upper

109 100

139 48.3 87.5 131.3#



65 28.3 90.2 135.2#




Abundance Ion 109.00 (108.70 to 109.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 65.00 (64.70 to 65.70): BM





#53

Dibenzenofuran

Concen: 3.32 ng/ul

RT: 14.86 min Scan# 2086

Delta R.T. 0.01 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12

Instrument :

BNA_M

ClientSampleId :





E5NG0MS

Tgt Ion:168 Resp: 50198

Ion Ratio Lower Upper

168 100

139 69.5 30.6 46.0#



#54

2,4-Dinitrotoluene

Concen: 6.91 ng/ul

RT: 14.82 min Scan# 2079

Delta R.T. 0.01 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12




Tgt Ion:165 Resp: 25443


Ion Ratio Lower Upper

165 100

63 40.6 33.9 50.9

182 39.1 2.3 3.5#









#58
 Fluorene
 Concen: 3.90 ng/ul
 RT: 15.50 min Scan# 2196
 Delta R.T. 0.01 min
 Lab File: BM005567.D
 Acq: 22 May 2016 01:12




Instrument :
 BNA_M
 ClientSampleId :
 E5NG0MS


Tgt Ion:166 Resp: 47160
 Ion Ratio Lower Upper
 166 100
 165 128.6 78.0 117.0#
 167 50.4 10.5 15.7#



#63
 4,6-Dinitro-2-methylphenol
 Concen: 1.37 ng/ul
 RT: 15.55 min Scan# 2204
 Delta R.T. -0.02 min
 Lab File: BM005567.D
 Acq: 22 May 2016 01:12

Tgt Ion:198 Resp: 2183
 Ion Ratio Lower Upper
 198 100
 182 908.6 2.5 3.7#
 77 830.7 21.8 32.8#





#64

N-Nitrosodiphenylamine

Concen: 15.01 ng/ul

RT: 15.72 min Scan# 2233

Delta R.T. 0.01 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12

Instrument :

BNA_M

ClientSampleId :

E5NG0MS





Tgt Ion:169 Resp: 120174

Ion Ratio Lower Upper

169 100

168 23.3 53.2 79.8#

167 13.3 28.9 43.3#



#69

Phenanthrene

Concen: 42.06 ng/ul

RT: 17.25 min Scan# 2493

Delta R.T. 0.02 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12




Tgt Ion:178 Resp: 620279


Ion Ratio Lower Upper

178 100

179 21.7 12.2 18.2#

176 20.4 15.8 23.6





#71

Anthracene

Concen: 6.22 ng/ul

RT: 17.34 min Scan# 2508

Delta R.T. 0.01 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12

Instrument :

BNA_M

ClientSampleId :

E5NG0MS


Tgt Ion:178 Resp: 93477

Ion Ratio Lower Upper

178 100

179 44.2 12.6 18.8#




176 25.5 15.0 22.4#



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#73

Di-n-butylphthalate

Concen: 1.26 ng/ul

RT: 18.16 min Scan# 2648

Delta R.T. 0.01 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12


Tgt Ion:149 Resp: 19991

Ion Ratio Lower Upper

149 100

150 47.9 7.4 11.0#



104 29.4 4.4 6.6#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E





#74

Fluoranthene

Concen: 20.65 ng/ul

RT: 19.27 min Scan# 2837

Delta R.T. 0.03 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12

Instrument :

BNA_M

ClientSampleId :

E5NG0MS


Tgt Ion:202 Resp: 342858

Ion Ratio Lower Upper

202 100

101 11.2 7.9 11.9




100 10.8 5.6 8.4#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B



#77

Pyrene

Concen: 33.83 ng/ul

RT: 19.65 min Scan# 2901

Delta R.T. 0.04 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12


Tgt Ion:202 Resp: 453707

Ion Ratio Lower Upper

202 100

101 13.6 9.6 14.4



100 11.5 7.8 11.6




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B





#79

3,3'-Dichlorobenzidine

Concen: 1.54 ng/ul

RT: 21.34 min Scan# 3189

Delta R.T. 0.06 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12

Instrument :

BNA_M

ClientSampleId :

E5NG0MS





Tgt Ion:252 Resp: 6516

Ion Ratio Lower Upper

252 100

254 47.1 51.7 77.5#

126 272.4 7.9 11.9#



#80

Benzo(a)anthracene

Concen: 4.72 ng/ul

RT: 21.40 min Scan# 3198

Delta R.T. 0.05 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12




Tgt Ion:228 Resp: 64718


Ion Ratio Lower Upper

228 100

229 80.8 15.4 23.2#

226 51.8 21.0 31.4#









#81
 Bis(2-ethylhexyl)phthalate
 Concen: 2.48 ng/ul
 RT: 21.32 min Scan# 3184
 Delta R.T. 0.04 min
 Lab File: BM005567.D
 Acq: 22 May 2016 01:12




Instrument :
 BNA_M
 ClientSampleId :
 E5NG0MS


Tgt Ion:149 Resp: 19817
 Ion Ratio Lower Upper
 149 100
 167 79.4 22.0 33.0#
 279 21.8 3.8 5.8#



#82
 Chrysene
 Concen: 10.99 ng/ul
 RT: 21.44 min Scan# 3205
 Delta R.T. 0.04 min
 Lab File: BM005567.D
 Acq: 22 May 2016 01:12

Tgt Ion:228 Resp: 143434
 Ion Ratio Lower Upper
 228 100
 226 43.9 23.7 35.5#
 229 60.0 15.5 23.3#





#85

Benzo(b)fluoranthene

Concen: 4.42 ng/ul

RT: 23.02 min Scan# 3474

Delta R.T. 0.06 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12





Instrument :

BNA_M

ClientSampleId :

E5NG0MS

Tgt Ion:	252	Resp:	80655
Ion Ratio	Lower	Upper	
252	100		
253	109.3	17.3	25.9#
125	338.6	7.0	10.4#



#86

Benzo(k)fluoranthene

Concen: 4.67 ng/ul




RT: 23.02 min Scan# 3474


Delta R.T. 0.02 min

Lab File: BM005567.D

Acq: 22 May 2016 01:12

Tgt Ion:	252	Resp:	80655
Ion Ratio	Lower	Upper	
252	100		
253	109.3	17.5	26.3#
125	338.6	7.0	10.6#





#88

Benzo(a)pyrene

Concen: 1.16 ng/ul

RT: 23.61 min Scan# 3575

Delta R.T. 0.08 min

Lab File: BM005567.D


Acq: 22 May 2016 01:12

Instrument :

BNA_M

ClientSampleId :

E5NG0MS



Tgt Ion:	252	Resp:	20506
Ion	Ratio	Lower	Upper
252	100		
253	100.9	17.6	26.4#
125	591.1	7.6	11.4#

