

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052116\
 Data File : BM005557.D
 Acq On : 21 May 2016 19:09
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02059

Quant Time: May 22 05:07:39 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM052016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun May 22 04:56:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.02
2	1,4-Dioxane	0.806	0.762	5.5	86	0.04
3 S	1,4-Dioxane-d8	0.457	0.437	4.4	87	0.04
4	Benzaldehyde	1.064	0.749	29.6#	58	0.02
5 S	Phenol-d5	1.639	1.688	-3.0	89	0.00
6	Phenol	1.689	1.762	-4.3	89	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.950	0.996	-4.8	90	0.02
8	Bis(2-Chloroethyl)ether	1.276	1.333	-4.5	90	0.02
9 S	2-Chlorophenol-d4	1.368	1.383	-1.1	89	0.01
10	2-Chlorophenol	1.372	1.371	0.1	88	0.02
11	2-Methylphenol	1.280	1.327	-3.7	89	0.00
12	2,2'-oxybis(1-Chloropropane	1.704	1.795	-5.3	90	0.02
13 S	4-Methylphenol-d8	1.343	1.405	-4.6	89	0.00
14	Acetophenone	2.023	2.182	-7.9	90	0.02
15 P	N-Nitroso-di-n-propylamine	1.068	1.115	-4.4	89	0.02
16	4-Methylphenol	1.390	1.460	-5.0	89	0.00
17	Hexachloroethane	0.549	0.530	3.5	84	0.02
18 I	Naphthalene-d8	1.000	1.000	0.0	95	0.02
19 S	Nitrobenzene-d5	0.153	0.146	4.6	89	0.01
20	Nitrobenzene	0.370	0.353	4.6	88	0.01
21	Isophorone	0.693	0.674	2.7	88	0.02
22 S	2-Nitrophenol-d4	0.168	0.157	6.5	86	0.01
23 C	2-Nitrophenol	0.181	0.170	6.1	86	0.02
24	2,4-Dimethylphenol	0.374	0.366	2.1	90	0.01
25	Bis(2-Chloroethoxy)methane	0.408	0.408	0.0	92	0.02
26 S	2,4-Dichlorophenol-d3	0.317	0.310	2.2	90	0.00
27 C	2,4-Dichlorophenol	0.312	0.305	2.2	88	0.00
28	Naphthalene	1.005	0.982	2.3	90	0.02
29 S	4-Chloroaniline-d4	0.318	0.329	-3.5	95	0.01
30	4-Chloroaniline	0.325	0.337	-3.7	94	0.01
31 C	Hexachlorobutadiene	0.209	0.192	8.1	86	0.02
32	Caprolactam	0.104	0.107	-2.9	94	0.00
33 C	4-Chloro-3-methylphenol	0.342	0.346	-1.2	91	0.00
34	2-Methylnaphthalene	0.735	0.744	-1.2	93	0.01
35 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.01
36	1,2,4,5-Tetrachlorobenzene	0.683	0.637	6.7	89	0.01
37	Hexachlorocyclopentadiene	0.429	0.350	18.4	81	0.02
38 C	2,4,6-Trichlorophenol	0.445	0.418	6.1	90	0.00
39	2,4,5-Trichlorophenol	0.490	0.468	4.5	92	0.00
40	1,1'-Biphenyl	1.679	1.604	4.5	93	0.01
41	2-Chloronaphthalene	1.294	1.228	5.1	92	0.01
42	2-Nitroaniline	0.389	0.364	6.4	89	0.00
43 S	Dimethylphthalate-d6	1.633	1.613	1.2	96	0.02
44	Dimethylphthalate	1.613	1.602	0.7	98	0.01

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052116\
 Data File : BM005557.D
 Acq On : 21 May 2016 19:09
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02059

Quant Time: May 22 05:07:39 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM052016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun May 22 04:56:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.326	0.321	1.5	96	0.00
46 S	Acenaphthylene-d8	2.041	1.974	3.3	93	0.01
47	Acenaphthylene	2.074	2.007	3.2	93	0.01
48	3-Nitroaniline	0.302	0.296	2.0	92	0.01
49 C	Acenaphthene	1.367	1.339	2.0	94	0.01
50	2,4-Dinitrophenol	0.183	0.144	21.3#	82	0.00
51 S	4-Nitrophenol-d4	0.309	0.300	2.9	93	-0.01
52	4-Nitrophenol	0.310	0.272	12.3	87	-0.01
53	Dibenzofuran	1.953	1.905	2.5	94	0.01
54	2,4-Dinitrotoluene	0.476	0.486	-2.1	98	0.00
55	2,3,4,6-Tetrachlorophenol	0.429	0.408	4.9	92	0.00
56	Diethylphthalate	1.645	1.663	-1.1	98	0.02
57 S	Fluorene-d10	1.421	1.417	0.3	98	0.01
58	Fluorene	1.563	1.565	-0.1	97	0.00
59	4-Chlorophenyl-phenylether	0.781	0.765	2.0	94	0.01
60	4-Nitroaniline	0.369	0.327	11.4	84	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.115	0.101	12.2	92	0.00
63	4,6-Dinitro-2-methylphenol	0.121	0.110	9.1	93	0.00
64	N-Nitrosodiphenylamine	0.607	0.596	1.8	99	0.01
65	4-Bromophenyl-phenylether	0.223	0.215	3.6	97	0.01
66	Hexachlorobenzene	0.254	0.242	4.7	96	0.00
67	Atrazine	0.227	0.227	0.0	99	0.01
68 C	Pentachlorophenol	0.152	0.141	7.2	94	0.00
69	Phenanthrene	1.118	1.101	1.5	100	0.00
70 S	Anthracene-d10	0.953	0.941	1.3	100	0.01
71	Anthracene	1.139	1.118	1.8	99	0.01
72	Carbazole	0.989	1.031	-4.2	102	0.00
73	Di-n-butylphthalate	1.200	1.220	-1.7	102	0.01
74 C	Fluoranthene	1.258	1.307	-3.9	101	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
76 S	Pyrene-d10	0.906	0.936	-3.3	102	0.00
77	Pyrene	1.150	1.191	-3.6	102	0.00
78	Butylbenzylphthalate	0.485	0.495	-2.1	100	0.01
79	3,3'-Dichlorobenzidine	0.363	0.300	17.4	74	0.00
80	Benzo(a)anthracene	1.176	1.138	3.2	96	0.00
81	Bis(2-ethylhexyl)phthalate	0.684	0.689	-0.7	98	0.02
82	Chrysene	1.118	1.085	3.0	95	0.00
83 I	Perylene-d12	1.000	1.000	0.0	87	0.00
84	Di-n-octyl phthalate	1.147	1.293	-12.7	92	0.02
85	Benzo(b)fluoranthene	1.180	1.230	-4.2	89	0.00
86	Benzo(k)fluoranthene	1.117	1.132	-1.3	85	0.00
87 S	Benzo(a)pyrene-d12	0.934	0.916	1.9	84	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052116\
Data File : BM005557.D
Acq On : 21 May 2016 19:09
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02059

Quant Time: May 22 05:07:39 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM052016.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun May 22 04:56:37 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.144	1.133	1.0	84	0.00
89	Indeno(1,2,3-cd)pyrene	1.359	1.209	11.0	75	0.00
90	Dibenzo(a,h)anthracene	1.132	1.013	10.5	75	0.00
91	Benzo(g,h,i)perylene	1.143	1.004	12.2	74	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0