


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052116\
Data File : BM005549.D
Acq On : 21 May 2016 11:17
Operator : UM/SJ
Sample : DFTPP43
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP43

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM052016.M
Title : SVOA CALIBRATION
Last Update : Tue May 24 06:55:29 2016



AutoFind: Scans 2557, 2558, 2559; Background Corrected with Scan 2551

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.0	75970	PASS
68	69	0.00	2	1.7	1322	PASS
69	198	0.00	100	35.1	78394	PASS
70	69	0.00	2	0.5	402	PASS
127	198	10	80	45.1	100618	PASS
197	198	0.00	2	0.4	934	PASS
198	198	100	100	100.0	223232	PASS
199	198	5	9	6.8	15209	PASS
275	198	10	60	27.2	60661	PASS
365	198	1	100	4.0	8947	PASS
441	443	0.01	100	81.5	31157	PASS
442	198	50	100	89.5	199829	PASS
443	442	15	24	19.1	38208	PASS

Scan 2551 (17.592 min): BM005549.D
DFTPP43

m/z	Abundance
40.00	391.0
44.00	742.0
207.10	1743.0
208.00	409.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

m/z	Abundance
37.00	362.0
38.00	1113.0
39.10	5039.0
40.00	622.0
44.00	582.0
49.10	585.0
50.00	19576.0
51.00	77088.0
52.00	4149.0
55.00	533.0
56.00	2065.0
57.00	4951.0
58.00	304.0
61.00	857.0
62.00	1090.0
63.00	3044.0
64.00	385.0
65.10	1473.0
68.00	1451.0
69.00	78080.0
70.00	411.0
73.00	898.0
74.00	8177.0
75.00	13177.0
76.00	4396.0
77.00	91784.0
78.00	6025.0
79.00	6069.0
80.00	4534.0
81.00	6432.0
82.00	1614.0
83.00	1362.0
85.00	1361.0
86.00	2097.0
87.00	1071.0
87.90	364.0
90.90	1579.0
91.90	1920.0
93.00	11452.0
94.00	788.0
96.00	621.0
98.00	8184.0
99.00	7540.0
99.90	743.0
101.00	3908.0
103.00	1079.0
104.00	2200.0
105.00	2473.0
106.00	731.0
107.00	28976.0
108.00	4249.0
109.00	767.0
110.00	51440.0
111.00	7802.0
111.90	1123.0
112.90	317.0
116.00	1540.0
117.00	25440.0
118.00	1974.0
119.00	313.0
120.00	436.0
122.00	1867.0
123.00	3136.0
124.00	1497.0
125.00	1086.0
126.20	462.0
127.00	95112.0
128.00	7230.0
129.00	38152.0
130.00	3344.0
130.90	555.0
132.00	375.0
133.00	326.0
134.00	1105.0
135.00	3062.0
136.00	1367.0
136.90	1995.0
137.80	341.0
140.00	479.0
141.00	4544.0
142.00	1538.0
142.90	1016.0
143.90	320.0
144.90	383.0
146.00	940.0
147.00	2872.0
148.00	6237.0
149.00	1310.0
149.90	324.0
151.00	490.0
151.80	325.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

153.00	1376.0
153.90	1124.0
155.00	2505.0
156.00	4248.0
157.10	782.0
157.90	891.0
158.90	756.0
160.00	1735.0
161.00	2451.0
162.00	377.0
165.00	1983.0
166.00	1496.0
167.00	10810.0
168.00	6753.0
169.00	1037.0
170.90	474.0
171.90	979.0
173.00	1221.0
174.00	2125.0
175.00	4017.0
176.00	1111.0
176.90	1620.0
177.90	564.0
179.00	7319.0
180.00	5636.0
181.00	2434.0
182.00	410.0
184.00	709.0
185.00	3760.0
186.00	29080.0
187.00	8320.0
188.00	1100.0
189.00	1609.0
190.00	386.0
191.00	1119.0
192.00	2475.0
193.00	2660.0
194.00	672.0
194.90	465.0
196.00	5386.0
196.80	1167.0
198.00	196416.0
199.00	14221.0
200.00	1014.0
201.50	835.0
203.00	1448.0
204.00	7611.0
205.00	12769.0
206.10	51624.0
207.00	8724.0
208.00	2445.0
208.90	748.0
210.10	885.0
211.00	2095.0
211.70	560.0
215.00	532.0
216.00	1174.0
217.00	14269.0
218.00	1868.0
221.00	7355.0
221.80	1941.0
223.00	3083.0
224.00	28536.0
225.00	7948.0
226.00	972.0
227.00	12633.0
228.00	1663.0
229.00	2888.0
230.00	340.0
231.00	1027.0
234.00	808.0
235.00	932.0
236.00	732.0
237.00	1420.0
239.00	589.0
239.80	517.0
240.90	814.0
242.00	1545.0
243.00	1725.0
244.10	20808.0
245.00	3072.0
246.00	4644.0
247.00	717.0
249.00	1034.0
252.90	506.0
254.10	871.0
255.00	115104.0
256.00	17528.0
257.00	1538.0
258.00	7045.0
259.00	963.0
265.00	2315.0
265.90	475.0
271.90	414.0
273.00	3284.0

Instrument :

BNA_M

ClientSampleId :

DFTPP43

274.00	8984.0
275.00	49048.0
276.00	6611.0
277.00	4814.0
278.00	866.0
282.90	641.0
284.00	375.0
285.00	839.0
293.00	1265.0
295.00	368.0
296.00	15010.0
297.00	2527.0
303.00	1575.0
304.00	410.0
314.00	783.0
314.90	1936.0
316.00	1009.0
320.90	343.0
323.10	4375.0
324.10	825.0
326.90	985.0
328.00	472.0
331.90	306.0
333.10	379.0
334.00	2959.0
335.00	666.0
341.00	437.0
346.00	939.0
352.00	1312.0
353.10	903.0
354.00	1381.0
365.00	6223.0
366.00	850.0
371.00	448.0
372.00	2684.0
373.10	558.0
382.90	650.0
402.10	928.0
403.00	1333.0
404.00	659.0
421.00	1172.0
422.00	1356.0
423.10	7652.0
424.00	2442.0
425.10	343.0
441.00	21680.0
442.10	137088.0
443.10	27288.0
444.00	2865.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

m/z	Abundance
37.00	500.0
38.10	1087.0
39.10	6174.0
40.00	640.0
41.10	324.0
44.00	806.0
49.00	777.0
50.00	22568.0
51.00	88352.0
52.00	4638.0
54.90	565.0
56.00	2437.0
57.00	5988.0
61.00	1249.0
62.00	1479.0
63.10	3856.0
64.00	483.0
65.00	1897.0
68.10	1474.0
69.00	91896.0
70.00	487.0
73.00	1138.0
74.00	10004.0
75.00	15603.0
76.10	5532.0
77.10	107416.0
78.00	7423.0
79.00	7203.0
80.00	5256.0
81.00	8050.0
82.00	2310.0
83.00	1554.0
85.00	1602.0
86.00	2187.0
86.90	1144.0
88.00	427.0
91.00	1967.0
92.00	1962.0
93.00	13947.0
94.10	838.0
95.00	327.0
96.00	621.0
98.00	10404.0
99.00	8326.0
100.00	732.0
101.00	4445.0
101.90	318.0
103.00	1323.0
104.00	2672.0
105.00	2701.0
106.00	1036.0
107.00	36672.0
108.00	5047.0
108.90	678.0
110.00	59464.0
111.00	9380.0
112.00	1150.0
113.10	335.0
116.00	1758.0
117.00	31584.0
118.00	2132.0
120.00	452.0
122.00	2353.0
123.00	3947.0
124.00	1718.0
125.00	1658.0
126.10	605.0
127.00	117392.0
128.00	8634.0
129.00	47568.0
130.00	3834.0
131.00	699.0
132.10	355.0
132.90	319.0
133.90	1669.0
135.00	3824.0
136.00	1654.0
137.10	2058.0
138.00	497.0
140.00	641.0
141.00	6603.0
142.00	1932.0
142.90	1373.0
143.90	517.0
144.90	354.0
145.90	1090.0
147.00	3354.0
148.00	7595.0
149.00	1665.0
149.90	332.0
151.00	700.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

151.90	535.0
153.00	1764.0
154.00	1525.0
155.00	3722.0
156.00	5367.0
157.00	1092.0
157.90	1306.0
159.00	747.0
160.00	1769.0
161.00	3103.0
162.00	918.0
163.90	332.0
165.00	2812.0
165.90	1634.0
167.00	13162.0
168.00	7965.0
169.00	1189.0
170.00	420.0
171.00	543.0
172.00	1420.0
173.00	1487.0
174.00	2596.0
175.00	5292.0
176.00	1202.0
177.00	2350.0
178.00	802.0
179.00	9572.0
180.00	6838.0
181.00	3154.0
181.90	465.0
184.00	794.0
185.00	4593.0
186.00	38784.0
187.00	11154.0
188.00	1125.0
189.00	2592.0
190.00	363.0
191.00	1091.0
192.00	3193.0
193.00	3681.0
194.00	819.0
196.00	6929.0
197.10	1502.0
198.00	263616.0
199.00	17912.0
200.00	1450.0
201.20	787.0
201.60	725.0
203.00	1887.0
204.00	10211.0
205.00	16728.0
206.00	69800.0
207.00	11137.0
208.00	2502.0
209.00	1074.0
209.90	1012.0
211.00	2824.0
212.00	320.0
214.90	719.0
216.00	1748.0
217.00	19648.0
218.00	2592.0
221.00	9678.0
222.00	2148.0
223.00	4347.0
224.00	39544.0
225.10	9997.0
226.00	1142.0
227.00	18160.0
228.00	2629.0
229.00	3769.0
230.00	439.0
231.00	1286.0
234.00	1143.0
235.00	1264.0
236.00	1050.0
237.00	1725.0
239.00	623.0
240.00	632.0
241.00	933.0
242.00	1977.0
243.00	2060.0
244.10	30688.0
245.10	3821.0
246.00	6560.0
247.00	1410.0
247.90	326.0
249.00	1420.0
252.00	436.0
252.90	690.0
254.10	1315.0
255.00	162176.0
256.00	24080.0
257.10	2000.0
258.00	10219.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

259.00	1601.0
260.00	339.0
264.00	324.0
265.00	4384.0
265.90	483.0
271.00	387.0
272.00	481.0
273.00	5029.0
274.00	13831.0
275.00	71256.0
276.00	9787.0
277.00	6499.0
278.00	1255.0
279.00	313.0
283.10	770.0
284.00	501.0
285.00	1172.0
289.00	346.0
293.00	1432.0
294.00	537.0
295.10	565.0
296.00	22624.0
297.00	3019.0
301.00	309.0
302.00	399.0
303.00	2326.0
304.00	770.0
307.90	380.0
310.00	324.0
314.00	970.0
315.00	2329.0
316.10	1422.0
321.00	628.0
322.00	348.0
323.10	6234.0
324.00	1152.0
326.90	1318.0
328.00	747.0
332.00	572.0
333.00	557.0
334.00	4695.0
335.00	1077.0
341.00	778.0
342.10	342.0
346.00	1378.0
346.90	348.0
352.00	2199.0
353.00	1639.0
354.10	2238.0
355.10	319.0
365.00	10072.0
366.00	1358.0
371.10	492.0
372.00	3718.0
373.00	864.0
383.10	940.0
390.10	562.0
402.00	1500.0
403.00	2076.0
404.00	923.0
421.00	1877.0
422.00	2190.0
423.10	11904.0
424.00	3491.0
425.00	473.0
440.00	320.0
441.00	35872.0
442.10	231744.0
443.00	44064.0
444.00	4021.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

m/z	Abundance
38.00	747.0
39.10	4039.0
40.00	576.0
44.00	722.0
49.10	462.0
50.00	16317.0
51.10	62472.0
52.00	3343.0
54.90	348.0
56.00	1884.0
57.00	4011.0
61.00	820.0
62.00	908.0
63.10	2523.0
64.00	319.0
65.00	1435.0
68.00	1043.0
69.00	65208.0
70.10	308.0
73.00	775.0
74.00	6388.0
75.00	11163.0
76.00	3970.0
77.10	76288.0
78.00	5411.0
79.00	5182.0
80.00	3955.0
81.00	5661.0
82.00	1301.0
83.00	1173.0
85.00	1230.0
86.00	1638.0
87.00	694.0
91.00	1303.0
92.00	1419.0
93.00	9253.0
94.00	644.0
96.10	415.0
98.00	7663.0
99.00	6168.0
99.90	506.0
101.00	3471.0
103.00	1123.0
103.90	2182.0
105.00	2084.0
106.00	714.0
107.00	26272.0
108.00	3994.0
109.00	635.0
110.00	45952.0
111.00	7061.0
111.90	946.0
116.00	1279.0
117.00	23776.0
118.00	1541.0
121.90	1736.0
123.00	2898.0
124.00	1118.0
125.00	1068.0
126.10	379.0
127.00	89352.0
128.00	6687.0
129.00	34624.0
130.00	2867.0
131.00	609.0
131.90	335.0
133.90	949.0
135.00	2973.0
136.00	1284.0
137.00	1647.0
140.00	517.0
141.00	5080.0
142.00	1643.0
143.00	1143.0
143.90	341.0
146.00	771.0
147.00	2503.0
148.00	5799.0
149.00	1491.0
149.90	421.0
150.90	437.0
151.80	324.0
153.10	1440.0
154.00	1208.0
155.00	2717.0
156.00	4021.0
157.10	962.0
157.90	616.0
159.00	672.0
159.90	1643.0
161.00	2482.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

162.00	804.0
165.00	1859.0
166.00	1470.0
167.00	10248.0
168.00	5894.0
169.00	944.0
169.90	300.0
171.00	483.0
171.90	870.0
173.00	1324.0
174.00	2355.0
175.00	4280.0
175.90	893.0
177.00	2475.0
178.00	690.0
179.00	7956.0
180.00	5253.0
181.00	2391.0
182.00	444.0
183.90	670.0
185.00	3979.0
186.00	29536.0
187.00	8229.0
188.00	971.0
189.00	1847.0
191.00	951.0
192.00	2533.0
193.00	3234.0
194.00	591.0
196.00	5175.0
197.00	1301.0
198.00	209664.0
199.00	13495.0
200.00	1053.0
201.30	776.0
202.90	1585.0
204.00	6819.0
205.00	12932.0
206.00	55832.0
207.00	8392.0
208.00	2415.0
208.90	771.0
210.20	736.0
211.00	2065.0
215.10	783.0
216.00	1149.0
217.00	15785.0
218.00	2217.0
221.00	8341.0
222.00	1743.0
223.00	3963.0
224.00	32400.0
225.00	8522.0
226.00	1095.0
227.00	15270.0
228.00	2290.0
229.00	3479.0
231.00	1214.0
233.90	857.0
235.00	1229.0
235.90	735.0
237.00	1240.0
239.00	652.0
240.00	487.0
241.00	793.0
242.00	1545.0
243.10	1775.0
244.10	24984.0
245.00	3513.0
246.00	5275.0
246.90	1003.0
249.00	1012.0
253.00	642.0
254.10	1399.0
255.00	137600.0
256.00	20032.0
257.00	1497.0
258.00	8666.0
259.00	1604.0
265.00	3757.0
265.90	510.0
272.00	487.0
273.00	4371.0
274.00	10964.0
275.00	61680.0
276.00	8486.0
277.00	5359.0
278.00	1150.0
281.00	388.0
283.00	793.0
283.90	506.0
285.10	977.0
293.00	1560.0
294.00	512.0
295.00	332.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43

296.00	19960.0
297.00	2504.0
302.00	401.0
303.00	2391.0
304.00	632.0
314.00	935.0
315.00	2197.0
316.00	1231.0
321.00	671.0
322.10	315.0
323.00	6084.0
324.10	1088.0
327.00	1230.0
328.00	627.0
331.90	428.0
333.00	559.0
334.00	4179.0
335.00	1142.0
341.00	699.0
346.00	1404.0
352.00	1888.0
353.00	1466.0
354.10	2125.0
355.00	451.0
365.00	10548.0
366.00	1534.0
371.00	483.0
372.00	3258.0
373.00	941.0
383.00	783.0
390.00	388.0
391.00	389.0
402.00	1255.0
403.00	2286.0
404.00	879.0
421.00	1800.0
422.00	1664.0
423.10	11571.0
424.10	3173.0
425.00	446.0
439.20	314.0
440.10	330.0
441.10	35920.0
442.10	230656.0
443.10	43272.0
444.10	4328.0
445.00	302.0

Instrument :
BNA_M
ClientSampleId :
DFTPP43