

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052215\
 Data File : BM001409.D
 Acq On : 23 May 2015 08:11
 Operator : TP/IZ
 Sample : G2352-03
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 052115-P20-E-U-B

Quant Time: May 25 08:57:35 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM052215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 23 04:45:39 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	9486	5.00	ng	0.00
6) Naphthalene-d8	10.18	136	34122	5.00	ng	0.00
12) Acenaphthene-d10	14.08	164	20364	5.00	ng	0.00
18) Phenanthrene-d10	16.82	188	54465	5.00	ng	0.00
25) Chrysene-d12	21.03	240	39703	5.00	ng	0.00
32) Perylene-d12	23.13	264	35366	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	4.99	112	7546	4.03	ng	0.00
5) Phenol-d6	6.60	99	5345	2.46	ng	0.00
7) Nitrobenzene-d5	8.55	82	14274	7.25	ng	0.00
13) 2,4,6-Tribromophenol	15.57	330	7298	11.54	ng	-0.03
14) 2-Fluorobiphenyl	12.68	172	49847	7.95	ng	0.00
27) Terphenyl-d14	19.49	244	52124	9.70	ng	0.00
Target Compounds						
24) Fluoranthene	18.89	202	290	0.03	ng	# 95
26) Pyrene	19.26	202	315m	0.03	ng	
28) Benzo(a)anthracene	21.02	228	331m	0.03	ng	98
29) Chrysene	21.06	228	200m	0.02	ng	
30) Bis(2-ethylhexyl)phthalate	20.97	149	2641	0.42	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052215\
Data File : BM001409.D
Acq On : 23 May 2015 08:11
Operator : TP/IZ
Sample : G2352-03
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
052115-P20-E-U-B

Quant Time: May 25 08:57:35 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM052215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 23 04:45:39 2015
Response via : Initial Calibration

