

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052215\
 Data File : BM001393.D
 Acq On : 22 May 2015 21:29
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM052215

Quant Time: May 23 05:00:51 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM052215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 23 04:45:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.449	0.398	11.4	88	0.00
3	n-Nitrosodimethylamine	0.639	0.583	8.8	86	0.00
4 S	2-Fluorophenol	0.988	0.924	6.5	86	0.00
5 S	Phenol-d6	1.146	1.047	8.6	85	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
7 S	Nitrobenzene-d5	0.289	0.262	9.3	85	0.00
8	Nitrobenzene	0.305	0.272	10.8	84	0.00
9	Naphthalene	1.064	1.024	3.8	87	0.00
10	Hexachlorobutadiene	0.217	0.207	4.6	86	0.00
11	2-Methylnaphthalene	0.718	0.683	4.9	86	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
13 S	2,4,6-Tribromophenol	0.155	0.123	20.6	74	0.00
14 S	2-Fluorobiphenyl	1.539	1.460	5.1	87	0.00
15	Acenaphthylene	2.115	1.988	6.0	88	0.00
16	Acenaphthene	1.317	1.266	3.9	89	0.00
17	Fluorene	2.246	2.261	-0.7	85	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
19	4-Bromophenyl-phenylether	0.086	0.067	22.1	84	0.00
20	Hexachlorobenzene	0.219	0.195	11.0	95	0.00
21	Pentachlorophenol	0.043	0.030	30.2#	88	0.00
22	Phenanthrene	1.066	0.967	9.3	96	0.00
23	Anthracene	0.893	0.830	7.1	81	0.00
24	Fluoranthene	0.922	0.825	10.5	88	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
26	Pyrene	1.509	1.389	8.0	88	0.00
27 S	Terphenyl-d14	0.677	0.616	9.0	87	0.00
28	Benzo(a)anthracene	1.476	1.377	6.7	92	0.00
29	Chrysene	1.323	1.197	9.5	87	0.00
30	Bis(2-ethylhexyl)phthalate	0.797	0.622	22.0	84	-0.02
31	Indeno(1,2,3-cd)pyrene	1.460	1.346	7.8	87	-0.01
32 I	Perylene-d12	1.000	1.000	0.0	95	0.00
33	Benzo(b)fluoranthene	1.515	1.414	6.7	91	0.00
34	Benzo(k)fluoranthene	1.412	1.321	6.4	85	0.00
35 C	Benzo(a)pyrene	1.337	1.234	7.7	88	0.00
36	Dibenzo(a,h)anthracene	1.211	1.133	6.4	88	-0.01
37	Benzo(g,h,i)perylene	1.287	1.207	6.2	88	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0