

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052215\
 Data File : BM001393.D
 Acq On : 22 May 2015 21:29
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM052215

Quant Time: May 23 05:00:51 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM052215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 23 04:45:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	95	0.00
2	1,4-Dioxane	1.000	0.845	15.5	88	0.00
3	n-Nitrosodimethylamine	1.000	0.913	8.7	86	0.00
4 S	2-Fluorophenol	1.000	0.935	6.5	86	0.00
5 S	Phenol-d6	1.000	0.914	8.6	85	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	94	0.00
7 S	Nitrobenzene-d5	1.000	0.909	9.1	85	0.00
8	Nitrobenzene	1.000	0.893	10.7	84	0.00
9	Naphthalene	1.000	0.962	3.8	87	0.00
10	Hexachlorobutadiene	1.000	0.956	4.4	86	0.00
11	2-Methylnaphthalene	1.000	0.951	4.9	86	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	95	0.00
13 S	2,4,6-Tribromophenol	1.000	0.790	21.0	74	0.00
14 S	2-Fluorobiphenyl	1.000	0.949	5.1	87	0.00
15	Acenaphthylene	1.000	0.940	6.0	88	0.00
16	Acenaphthene	1.000	0.961	3.9	89	0.00
17	Fluorene	1.000	0.872	12.8	85	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	97	0.00
19	4-Bromophenyl-phenylether	1.000	1.060	-6.0	84	0.00
20	Hexachlorobenzene	1.000	1.023	-2.3	95	0.00
21	Pentachlorophenol	1.000	1.122	-12.2	88	0.00
22	Phenanthrene	1.000	1.043	-4.3	96	0.00
23	Anthracene	1.000	0.929	7.1	81	0.00
24	Fluoranthene	1.000	0.895	10.5	88	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	98	0.00
26	Pyrene	1.000	0.921	7.9	88	0.00
27 S	Terphenyl-d14	1.000	0.910	9.0	87	0.00
28	Benzo(a)anthracene	1.000	0.932	6.8	92	0.00
29	Chrysene	1.000	0.905	9.5	87	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.781	21.9	84	-0.02
31	Indeno(1,2,3-cd)pyrene	1.000	0.922	7.8	87	-0.01
32 I	Perylene-d12	5.000	5.000	0.0	95	0.00
33	Benzo(b)fluoranthene	1.000	0.933	6.7	91	0.00
34	Benzo(k)fluoranthene	1.000	0.936	6.4	85	0.00
35 C	Benzo(a)pyrene	1.000	0.923	7.7	88	0.00
36	Dibenzo(a,h)anthracene	1.000	0.935	6.5	88	-0.01
37	Benzo(g,h,i)perylene	1.000	0.937	6.3	88	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0