

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052215\
 Data File : BM001395.D
 Acq On : 22 May 2015 23:16
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 23 05:14:18 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM052215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 23 04:45:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.449	0.489	-8.9	110	0.00
3	n-Nitrosodimethylamine	0.639	0.710	-11.1	106	-0.02
4 S	2-Fluorophenol	0.988	1.015	-2.7	96	0.00
5 S	Phenol-d6	1.146	1.174	-2.4	97	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
7 S	Nitrobenzene-d5	0.289	0.284	1.7	96	0.00
8	Nitrobenzene	0.305	0.298	2.3	97	0.00
9	Naphthalene	1.064	1.094	-2.8	98	0.00
10	Hexachlorobutadiene	0.217	0.226	-4.1	98	0.00
11	2-Methylnaphthalene	0.718	0.746	-3.9	99	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
13 S	2,4,6-Tribromophenol	0.155	0.123	20.6	78	0.00
14 S	2-Fluorobiphenyl	1.539	1.583	-2.9	99	0.00
15	Acenaphthylene	2.115	2.179	-3.0	100	0.00
16	Acenaphthene	1.317	1.349	-2.4	99	0.00
17	Fluorene	2.246	2.389	-6.4	94	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
19	4-Bromophenyl-phenylether	0.086	0.066	23.3	86	0.00
20	Hexachlorobenzene	0.219	0.238	-8.7	122	0.00
21	Pentachlorophenol	0.043	0.032	25.6#	100	0.00
22	Phenanthrene	1.066	1.133	-6.3	118	0.00
23	Anthracene	0.893	0.843	5.6	87	0.00
24	Fluoranthene	0.922	0.910	1.3	103	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
26	Pyrene	1.509	1.516	-0.5	103	0.00
27 S	Terphenyl-d14	0.677	0.672	0.7	101	-0.01
28	Benzo(a)anthracene	1.476	1.549	-4.9	110	0.00
29	Chrysene	1.323	1.250	5.5	96	0.00
30	Bis(2-ethylhexyl)phthalate	0.797	0.679	14.8	98	-0.02
31	Indeno(1,2,3-cd)pyrene	1.460	1.446	1.0	100	-0.01
32 I	Perylene-d12	1.000	1.000	0.0	100	0.00
33	Benzo(b)fluoranthene	1.515	1.610	-6.3	109	0.00
34	Benzo(k)fluoranthene	1.412	1.369	3.0	93	0.00
35 C	Benzo(a)pyrene	1.337	1.352	-1.1	101	-0.01
36	Dibenzo(a,h)anthracene	1.211	1.229	-1.5	100	-0.01
37	Benzo(g,h,i)perylene	1.287	1.301	-1.1	100	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0