

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052215\
 Data File : BM001411.D
 Acq On : 23 May 2015 09:23
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 25 08:28:05 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM052215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 23 04:45:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	0.00
2	1,4-Dioxane	0.449	0.532	-18.5	154#	-0.02
3	n-Nitrosodimethylamine	0.639	0.657	-2.8	126	-0.02
4 S	2-Fluorophenol	0.988	1.016	-2.8	124	0.00
5 S	Phenol-d6	1.146	1.205	-5.1	128	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	128	-0.02
7 S	Nitrobenzene-d5	0.289	0.289	0.0	128	0.00
8	Nitrobenzene	0.305	0.323	-5.9	136	0.00
9	Naphthalene	1.064	1.113	-4.6	130	0.00
10	Hexachlorobutadiene	0.217	0.226	-4.1	128	0.00
11	2-Methylnaphthalene	0.718	0.764	-6.4	132	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	133	0.00
13 S	2,4,6-Tribromophenol	0.155	0.112	27.7#	94	0.00
14 S	2-Fluorobiphenyl	1.539	1.587	-3.1	132	0.00
15	Acenaphthylene	2.115	2.250	-6.4	138	0.00
16	Acenaphthene	1.317	1.358	-3.1	134	0.00
17	Fluorene	2.246	2.156	4.0	113	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	135	0.00
19	4-Bromophenyl-phenylether	0.086	0.065	24.4	114	0.00
20	Hexachlorobenzene	0.219	0.244	-11.4	166#	0.00
21	Pentachlorophenol	0.043	0.031	27.9#	129	-0.03
22	Phenanthrene	1.066	1.181	-10.8	163#	0.00
23	Anthracene	0.893	0.764	14.4	104	0.00
24	Fluoranthene	0.922	0.846	8.2	127	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
26	Pyrene	1.509	1.681	-11.4	125	0.00
27 S	Terphenyl-d14	0.677	0.723	-6.8	119	-0.01
28	Benzo(a)anthracene	1.476	1.511	-2.4	118	0.00
29	Chrysene	1.323	1.304	1.4	110	0.00
30	Bis(2-ethylhexyl)phthalate	0.797	0.706	11.4	112	-0.02
31	Indeno(1,2,3-cd)pyrene	1.460	1.250	14.4	95	-0.01
32 I	Perylene-d12	1.000	1.000	0.0	102	0.00
33	Benzo(b)fluoranthene	1.515	1.594	-5.2	109	0.00
34	Benzo(k)fluoranthene	1.412	1.460	-3.4	101	0.00
35 C	Benzo(a)pyrene	1.337	1.351	-1.0	103	0.00
36	Dibenzo(a,h)anthracene	1.211	1.124	7.2	93	-0.01
37	Benzo(g,h,i)perylene	1.287	1.249	3.0	97	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0