

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052215\
 Data File : BM001411.D
 Acq On : 23 May 2015 09:23
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 25 08:28:05 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM052215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 23 04:45:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	125	0.00
2	1,4-Dioxane	1.000	1.217	-21.7	154	-0.02
3	n-Nitrosodimethylamine	1.000	1.028	-2.8	126	-0.02
4 S	2-Fluorophenol	1.000	1.028	-2.8	124	0.00
5 S	Phenol-d6	1.000	1.051	-5.1	128	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	128	-0.02
7 S	Nitrobenzene-d5	1.000	1.002	-0.2	128	0.00
8	Nitrobenzene	1.000	1.059	-5.9	136	0.00
9	Naphthalene	1.000	1.046	-4.6	130	0.00
10	Hexachlorobutadiene	1.000	1.041	-4.1	128	0.00
11	2-Methylnaphthalene	1.000	1.063	-6.3	132	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	133	0.00
13 S	2,4,6-Tribromophenol	1.000	0.719	28.1#	94	0.00
14 S	2-Fluorobiphenyl	1.000	1.031	-3.1	132	0.00
15	Acenaphthylene	1.000	1.063	-6.3	138	0.00
16	Acenaphthene	1.000	1.031	-3.1	134	0.00
17	Fluorene	1.000	0.830	17.0	113	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	135	0.00
19	4-Bromophenyl-phenylether	1.000	1.032	-3.2	114	0.00
20	Hexachlorobenzene	1.000	1.249	-24.9	166	0.00
21	Pentachlorophenol	1.000	1.159	-15.9	129	-0.03
22	Phenanthrene	1.000	1.244	-24.4	163	0.00
23	Anthracene	1.000	0.856	14.4	104	0.00
24	Fluoranthene	1.000	0.918	8.2	127	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	114	0.00
26	Pyrene	1.000	1.114	-11.4	125	0.00
27 S	Terphenyl-d14	1.000	1.068	-6.8	119	-0.01
28	Benzo(a)anthracene	1.000	1.023	-2.3	118	0.00
29	Chrysene	1.000	0.986	1.4	110	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.886	11.4	112	-0.02
31	Indeno(1,2,3-cd)pyrene	1.000	0.856	14.4	95	-0.01
32 I	Perylene-d12	5.000	5.000	0.0	102	0.00
33	Benzo(b)fluoranthene	1.000	1.052	-5.2	109	0.00
34	Benzo(k)fluoranthene	1.000	1.034	-3.4	101	0.00
35 C	Benzo(a)pyrene	1.000	1.010	-1.0	103	0.00
36	Dibenzo(a,h)anthracene	1.000	0.928	7.2	93	-0.01
37	Benzo(g,h,i)perylene	1.000	0.970	3.0	97	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0