

Data Path : Z:\HPCHEM1\BNA_M\Data\BM052217\
 Data File : BM010131.D
 Acq On : 23 May 2017 08:22
 Operator : SJ/MA
 Sample : I3245-05MSD
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 GMW-7MSD

Quant Time: May 23 13:06:04 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM051917.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 20 04:25:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.96	152	145844	20.00	ng	-0.02
21) Naphthalene-d8	10.77	136	501375	20.00	ng	-0.02
38) Acenaphthene-d10	14.59	164	317354	20.00	ng	-0.02
63) Phenanthrene-d10	17.34	188	773050	20.00	ng	-0.01
75) Chrysene-d12	21.50	240	1221904	20.00	ng	-0.01
86) Perylene-d12	23.86	264	1334349	20.00	ng	-0.02

System Monitoring Compounds						
5) 2-Fluorophenol	5.54	112	593507	73.49	ng	-0.02
7) Phenol-d6	7.15	99	464138	44.66	ng	-0.02
23) Nitrobenzene-d5	9.15	82	1027836	110.60	ng	-0.02
41) 2,4,6-Tribromophenol	16.09	330	749927	155.59	ng	-0.01
44) 2-Fluorobiphenyl	13.22	172	2565058	103.91	ng	-0.02
78) Terphenyl-d14	19.94	244	4497083	83.68	ng	-0.01

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.43	88	79305	21.37	ng	# 100
3) Pyridine	3.86	79	166729	16.78	ng	89
4) n-Nitrosodimethylamine	3.77	42	90260	25.12	ng	# 85
6) Aniline	7.31	93	251782	19.55	ng	92
8) 2-Chlorophenol	7.53	128	332237	37.81	ng	97
9) Benzaldehyde	7.12	77	151770	23.46	ng	87
10) Phenol	7.17	94	166684	15.22	ng	88
11) bis(2-Chloroethyl)ether	7.39	93	361330	44.26	ng	97
12) 1,3-Dichlorobenzene	7.85	146	439409	39.35	ng	# 93
13) 1,4-Dichlorobenzene	7.99	146	458660	39.99	ng	96
14) 1,2-Dichlorobenzene	8.31	146	446365	40.54	ng	96
15) Benzyl Alcohol	8.22	79	242390	30.83	ng	# 86
16) 2,2'-oxybis(1-Chloropropan	8.49	45	303985	37.87	ng	77
17) 2-Methylphenol	8.42	107	225321	28.91	ng	# 86
18) Hexachloroethane	9.03	117	145437	39.11	ng	# 71
19) n-Nitroso-di-n-propylamine	8.77	70	334468	45.60	ng	90
20) 3+4-Methylphenols	8.75	107	282684	26.86	ng	87
22) Acetophenone	8.80	105	607138	47.24	ng	# 96
24) Nitrobenzene	9.19	77	553407	57.80	ng	97
25) Isophorone	9.70	82	811911	50.56	ng	# 94
26) 2-Nitrophenol	9.89	139	189950	47.07	ng	# 69
27) 2,4-Dimethylphenol	9.95	122	288057	39.53	ng	# 81
28) bis(2-Chloroethoxy)methane	10.18	93	461713	45.31	ng	99
29) 2,4-Dichlorophenol	10.42	162	383479	45.71	ng	97
30) 1,2,4-Trichlorobenzene	10.62	180	482266	44.43	ng	98
31) Naphthalene	10.82	128	1152914	42.99	ng	100
32) Benzoic acid	10.04	122	61349	12.35	ng	82
33) 4-Chloroaniline	10.96	127	137543	13.31	ng	# 86
34) Hexachlorobutadiene	11.06	225	328525	45.37	ng	96
35) Caprolactam	11.77	113	16206	6.73	ng	# 78
36) 4-Chloro-3-methylphenol	12.07	107	352556	38.80	ng	85
37) 2-Methylnaphthalene	12.42	142	889936	46.08	ng	# 94
39) 1,2,4,5-Tetrachlorobenzene	12.77	216	577245	47.30	ng	# 100
40) Hexachlorocyclopentadiene	12.73	237	724352	102.80	ng	98


Data Path : Z:\HPCHEM1\BNA_M\Data\BM052217\
 Data File : BM010131.D
 Acq On : 23 May 2017 08:22
 Operator : SJ/MA
 Sample : I3245-05MSD
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 GMW-7MSD

Quant Time: May 23 13:06:04 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM051917.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 20 04:25:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.03	196	355408	49.44	ng	99
43) 2,4,5-Trichlorophenol	13.10	196	374138	47.11	ng #	94
45) 1,1'-Biphenyl	13.43	154	1229238	47.10	ng	100
46) 2-Chloronaphthalene	13.47	162	969243	45.87	ng	97
47) 2-Nitroaniline	13.71	65	175839	32.36	ng #	78
48) Acenaphthylene	14.32	152	1479746	47.36	ng	99
49) Dimethylphthalate	14.06	163	1274133	45.01	ng #	99
50) 2,6-Dinitrotoluene	14.19	165	262546	48.84	ng	89
51) Acenaphthene	14.66	154	910447	46.87	ng	99
52) 3-Nitroaniline	14.54	138	146212	29.27	ng #	73
53) 2,4-Dinitrophenol	14.73	184	322307	89.43	ng #	84
54) Dibenzofuran	14.99	168	1551045	50.78	ng	98
55) 4-Nitrophenol	14.84	139	141417	35.17	ng #	41
56) 2,4-Dinitrotoluene	14.97	165	374530	49.25	ng #	83
57) Fluorene	15.64	166	1279839	48.19	ng	99
58) 2,3,4,6-Tetrachlorophenol	15.22	232	368999	54.54	ng #	100
59) Diethylphthalate	15.40	149	1212238	45.29	ng	98
60) 4-Chlorophenyl-phenylether	15.63	204	705657	47.32	ng	99
61) 4-Nitroaniline	15.70	138	99614	19.64	ng #	31
62) Azobenzene	15.92	77	1169860	59.44	ng	90
64) 4,6-Dinitro-2-methylphenol	15.72	198	221300	46.86	ng	76
65) n-Nitrosodiphenylamine	15.85	169	1066190	46.02	ng	99
66) 4-Bromophenyl-phenylether	16.52	248	472261	47.72	ng	93
67) Hexachlorobenzene	16.62	284	532187	44.26	ng	96
68) Atrazine	16.80	200	435999	49.63	ng	98
69) Pentachlorophenol	16.98	266	646281	97.86	ng	97
70) Phenanthrene	17.38	178	2077181	48.03	ng	99
71) Anthracene	17.47	178	2091843	48.73	ng	99
72) Carbazole	17.76	167	1885820	49.62	ng	99
73) Di-n-butylphthalate	18.27	149	2098606	46.58	ng #	98
74) Fluoranthene	19.38	202	2742335	48.83	ng	97
76) Benzidine	19.59	184	193772	8.60	ng #	96
77) Pyrene	19.75	202	2843312	43.40	ng	99
79) Butylbenzylphthalate	20.62	149	1059952	43.71	ng #	86
80) Benzo(a)anthracene	21.48	228	3183739	47.28	ng	99
81) 3,3'-Dichlorobenzidine	21.42	252	205948	7.60	ng #	96
82) Chrysene	21.53	228	2915371	45.65	ng	100
83) Bis(2-ethylhexyl)phthalate	21.36	149	1571774	43.34	ng #	97
84) Di-n-octyl phthalate	22.27	149	2885912	44.45	ng	97
85) Indeno(1,2,3-cd)pyrene	26.29	276	4827366	51.54	ng #	100
87) Benzo(b)fluoranthene	23.14	252	3855624	49.37	ng #	93
88) Benzo(k)fluoranthene	23.19	252	3607522	47.15	ng #	95
89) Benzo(a)pyrene	23.76	252	3610221	48.51	ng #	96
90) Dibenzo(a,h)anthracene	26.30	278	4112120	49.42	ng #	93
91) Benzo(g,h,i)perylene	27.04	276	3977816	48.79	ng #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 7.96 min Scan# 828

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion:152 Resp: 145844

Ion Ratio Lower Upper

152 100

150 153.4 119.5 179.3




115 60.9 43.0 64.4



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 21.37 ng

RT: 3.43 min Scan# 59

Delta R.T. -0.01 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion: 88 Resp: 79305

Ion Ratio Lower Upper

88 100

58 62.1 0.0 0.0#



43 29.9 0.0 0.0#




Abundance Ion 88.00 (87.70 to 88.70): BM0

Ion 58.00 (57.70 to 58.70): BM0

Ion 43.00 (42.70 to 43.70): BM0





#3

Pyridine

Concen: 16.78 ng

RT: 3.86 min Scan# 131

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion: 79 Resp: 166729

Ion Ratio Lower Upper

79 100

52 52.4 51.1 76.7




51 34.9 26.1 39.1



Abundance Ion 79.00 (78.70 to 79.70): BM010069.D (-130) (-)

Ion 52.00 (51.70 to 52.70): BM010069.D (-130) (-)

Ion 51.00 (50.70 to 51.70): BM010069.D (-130) (-)



#4

n-Nitrosodimethylamine

Concen: 25.12 ng

RT: 3.77 min Scan# 116

Delta R.T. -0.01 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion: 42 Resp: 90260

Ion Ratio Lower Upper

42 100

74 129.0 119.3 178.9



44 4.6 5.4 8.2#




Abundance Ion 42.00 (41.70 to 42.70): BM010069.D (-114) (-)

Ion 74.00 (73.70 to 74.70): BM010069.D (-114) (-)

Ion 44.00 (43.70 to 44.70): BM010069.D (-114) (-)





#5

2-Fluorophenol

Concen: 73.49 ng

RT: 5.54 min Scan# 417

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion: 112 Resp: 593507

Ion Ratio Lower Upper

112 100

64 63.1 38.6 57.8#




63 38.6 19.3 28.9#



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): B

Ion 63.00 (62.70 to 63.70): B



#6

Aniline

Concen: 19.55 ng

RT: 7.31 min Scan# 718

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion: 93 Resp: 251782

Ion Ratio Lower Upper

93 100

66 45.5 30.8 46.2



65 20.9 16.0 24.0




Abundance Ion 93.00 (92.70 to 93.70): B

Ion 66.00 (65.70 to 66.70): B

Ion 65.00 (64.70 to 65.70): B





#7

Phenol-d6

Concen: 44.66 ng

RT: 7.15 min Scan# 690

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion: 99 Resp: 464138

Ion Ratio Lower Upper

99 100

42 20.5 11.0 16.4#




71 47.2 23.4 35.2#



Abundance Ion 99.00 (98.70 to 99.70): BM010131.D (-677) (-)

Ion 42.00 (41.70 to 42.70): BM010131.D (-677) (-)

Ion 71.00 (70.70 to 71.70): BM010131.D (-677) (-)



#8

2-Chlorophenol

Concen: 37.81 ng

RT: 7.53 min Scan# 756

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion: 128 Resp: 332237

Ion Ratio Lower Upper

128 100

130 32.5 12.0 52.0



64 47.9 24.9 64.9




Abundance Ion 128.00 (127.70 to 128.70): BM010131.D (-742) (-)

Ion 130.00 (129.70 to 130.70): BM010131.D (-742) (-)

Ion 64.00 (63.70 to 64.70): BM010131.D (-742) (-)





#9

Benzaldehyde

Concen: 23.46 ng

RT: 7.12 min Scan# 686

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion: 77 Resp: 151770

Ion Ratio Lower Upper

77 100

105 88.1 78.8 118.8

106 79.2 73.7 113.7






Abundance Ion 77.00 (76.70 to 77.70): BM010131.D (-672) (-)

Ion 105.00 (104.70 to 105.70): BM010131.D (-672) (-)

Ion 106.00 (105.70 to 106.70): BM010131.D (-672) (-)

7.12



#10

Phenol

Concen: 15.22 ng

RT: 7.17 min Scan# 695

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion: 94 Resp: 166684

Ion Ratio Lower Upper

94 100

65 33.6 18.8 58.8

66 49.5 39.9 79.9






Abundance Ion 94.00 (93.70 to 94.70): BM010131.D (-682) (-)

Ion 65.00 (64.70 to 65.70): BM010131.D (-682) (-)

Ion 66.00 (65.70 to 66.70): BM010131.D (-682) (-)

7.17









#11
bis(2-Chloroethyl)ether
Concen: 44.26 ng
RT: 7.39 min Scan# 732
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22




Instrument :
BNA_M
ClientSampleId :
GMW-7MSD


Tgt Ion:	93	Resp:	361330
Ion	Ratio	Lower	Upper
93	100		
63	73.3	49.5	89.5
95	33.3	13.5	53.5



#12
1,3-Dichlorobenzene
Concen: 39.35 ng
RT: 7.85 min Scan# 809
Delta R.T. -0.01 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22

Tgt Ion:	146	Resp:	439409
Ion	Ratio	Lower	Upper
146	100		
148	64.7	50.9	76.3
75	37.1	21.4	32.2









#13
 1,4-Dichlorobenzene
 Concen: 39.99 ng
 RT: 7.99 min Scan# 834
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22




Instrument :
 BNA_M
 ClientSampleId :
 GMW-7MSD


Tgt Ion:146 Resp: 458660
 Ion Ratio Lower Upper
 146 100
 148 62.9 51.9 77.9
 111 40.3 29.2 43.8



#14
 1,2-Dichlorobenzene
 Concen: 40.54 ng
 RT: 8.31 min Scan# 888
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22

Tgt Ion:146 Resp: 446365
 Ion Ratio Lower Upper
 146 100
 148 65.1 52.3 78.5
 111 43.5 30.6 45.8









#15
Benzyl Alcohol
Concen: 30.83 ng
RT: 8.22 min Scan# 873
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22




Instrument :
BNA_M
ClientSampleId :
GMW-7MSD


Tgt Ion: 79 Resp: 242390
Ion Ratio Lower Upper
79 100
108 62.0 65.0 97.4#
77 63.3 53.0 79.6



#16
2,2'-oxybis(1-chloropropane)
Concen: 37.87 ng
RT: 8.49 min Scan# 918
Delta R.T. -0.01 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22

Tgt Ion: 45 Resp: 303985
Ion Ratio Lower Upper
45 100
77 26.0 0.0 35.2
79 19.2 0.0 32.0









#17
2-Methylphenol
Concen: 28.91 ng
RT: 8.42 min Scan# 907
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22




Instrument :
BNA_M
ClientSampleId :
GMW-7MSD


Tgt Ion:	107	Resp:	225321
Ion	Ratio	Lower	Upper
107	100		
108	108.4	89.8	134.8
77	57.0	32.6	48.8#
79	59.4	32.8	49.2#



#18
Hexachloroethane
Concen: 39.11 ng
RT: 9.03 min Scan# 1011
Delta R.T. -0.01 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22

Tgt Ion:	117	Resp:	145437
Ion	Ratio	Lower	Upper
117	100		
119	97.0	79.2	118.8
201	142.3	69.8	104.6#





#19

n-Nitroso-di-n-propylamine

Concen: 45.60 ng

RT: 8.77 min Scan# 966

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD

Tgt Ion: 70 Resp: 334468


Ion Ratio Lower Upper

70 100

42 45.4 44.5 66.7

101 8.2 7.4 11.2

130 18.9 15.3 22.9






Abundance Ion 70.00 (69.70 to 70.70): BM0

Ion 42.00 (41.70 to 42.70): BM0

Ion 101.00 (100.70 to 101.70): E

Ion 130.00 (129.70 to 130.70): E



#20

3+4-Methylphenols

Concen: 26.86 ng

RT: 8.75 min Scan# 963

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Tgt Ion: 107 Resp: 282684


Ion Ratio Lower Upper

107 100

108 84.3 73.2 113.2

77 43.9 10.7 50.7

79 37.4 9.6 49.6






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BM0

Ion 79.00 (78.70 to 79.70): BM0









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.77 min Scan# 1306
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22

Instrument :
BNA_M
ClientSampleId :
GMW-7MSD

Tgt Ion:	136	Resp:	501375
Ion	Ratio	Lower	Upper
136	100		
137	11.7	8.7	13.1
54	6.8	4.6	6.8
68	4.7	3.7	5.5




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BM
Ion 68.00 (67.70 to 68.70): BM






#22
Acetophenone
Concen: 47.24 ng
RT: 8.80 min Scan# 971
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22

Tgt Ion:	105	Resp:	607138
Ion	Ratio	Lower	Upper
105	100		
71	0.2	0.6	1.0#
51	29.3	21.6	32.4
120	21.4	16.1	24.1



Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BM
Ion 51.00 (50.70 to 51.70): BM
Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 110.60 ng
 RT: 9.15 min Scan# 1031
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22




Instrument :
 BNA_M
 ClientSampleId :
 GMW-7MSD


Tgt Ion: 82 Resp: 1027836
 Ion Ratio Lower Upper
 82 100
 128 36.6 32.8 49.2
 54 47.0 40.6 60.8



#24
 Nitrobenzene
 Concen: 57.80 ng
 RT: 9.19 min Scan# 1038
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22

Tgt Ion: 77 Resp: 553407
 Ion Ratio Lower Upper
 77 100
 123 38.4 32.6 49.0
 65 14.1 10.9 16.3





#25

Isophorone

Concen: 50.56 ng

RT: 9.70 min Scan# 1124

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion: 82 Resp: 811911

Ion Ratio Lower Upper

82 100

95 8.8 5.8 8.6#




138 17.1 16.3 24.5



Abundance Ion 82.00 (81.70 to 82.70): BM010131.D (-1124) (-)

Ion 95.00 (94.70 to 95.70): BM010131.D (-1124) (-)

Ion 138.00 (137.70 to 138.70): BM010131.D (-1124) (-)



#26

2-Nitrophenol

Concen: 47.07 ng

RT: 9.89 min Scan# 1157

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion: 139 Resp: 189950

Ion Ratio Lower Upper

139 100

109 39.8 20.1 30.1#



65 59.5 31.5 47.3#




Abundance Ion 139.00 (138.70 to 139.70): BM010131.D (-1157) (-)

Ion 109.00 (108.70 to 109.70): BM010131.D (-1157) (-)

Ion 65.00 (64.70 to 65.70): BM010131.D (-1157) (-)









#27
 2,4-Dimethylphenol
 Concen: 39.53 ng
 RT: 9.95 min Scan# 1166
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22




Instrument :
 BNA_M
 ClientSampleId :
 GMW-7MSD


Tgt Ion	Ratio	Lower	Upper
122	100		
107	135.9	84.7	127.1#
121	58.8	46.6	69.8



#28
 bis(2-Chloroethoxy)methane
 Concen: 45.31 ng
 RT: 10.18 min Scan# 1206
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22

Tgt Ion	Ratio	Lower	Upper
93	100		
95	31.8	25.5	38.3
123	11.6	8.6	13.0









#29
 2,4-Dichlorophenol
 Concen: 45.71 ng
 RT: 10.42 min Scan# 1247
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22




Instrument :
 BNA_M
 ClientSampleId :
 GMW-7MSD


Tgt Ion	Ratio	Lower	Upper
162	100		
164	65.4	42.8	82.8
98	33.8	14.5	54.5



#30
 1,2,4-Trichlorobenzene
 Concen: 44.43 ng
 RT: 10.62 min Scan# 1280
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22

Tgt Ion	Ratio	Lower	Upper
180	100		
182	98.9	77.6	116.4
145	29.0	23.4	35.2





#31

Naphthalene

Concen: 42.99 ng

RT: 10.82 min Scan# 1315

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion:128 Resp: 1152914

Ion Ratio Lower Upper

128 100

129 10.7 8.9 13.3




127 13.0 10.4 15.6



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 12.35 ng

RT: 10.04 min Scan# 1182

Delta R.T. -0.07 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion:122 Resp: 61349

Ion Ratio Lower Upper

122 100

105 138.0 99.9 139.9



77 112.3 73.7 113.7




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BM









#33
4-Chloroaniline
Concen: 13.31 ng
RT: 10.96 min Scan# 1338
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22




Instrument :
BNA_M
ClientSampleId :
GMW-7MSD


Tgt Ion:	127	Resp:	137543
Ion	Ratio	Lower	Upper
127	100		
129	32.4	25.3	37.9
65	37.1	17.6	26.4#
92	22.2	13.1	19.7#



#34
Hexachlorobutadiene
Concen: 45.37 ng
RT: 11.06 min Scan# 1356
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22

Tgt Ion:	225	Resp:	328525
Ion	Ratio	Lower	Upper
225	100		
223	63.8	47.9	71.9
227	65.0	50.1	75.1





#35

Caprolactam

Concen: 6.73 ng

RT: 11.77 min Scan# 1477

Delta R.T. -0.04 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion:113 Resp: 16206

Ion Ratio Lower Upper

113 100

55 137.1 145.9 185.9#




56 95.6 101.0 141.0#



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BM0

Ion 56.00 (55.70 to 56.70): BM0



#36

4-Chloro-3-methylphenol

Concen: 38.80 ng

RT: 12.07 min Scan# 1527

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion:107 Resp: 352556

Ion Ratio Lower Upper

107 100

144 24.2 23.3 34.9



142 74.6 72.6 109.0




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#37
2-Methylnaphthalene
Concen: 46.08 ng
RT: 12.42 min Scan# 1586
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22




Instrument :
BNA_M
ClientSampleId :
GMW-7MSD


Tgt Ion	Ratio	Lower	Upper
142	100		
141	88.7	71.9	107.9
115	43.0	25.4	38.0



#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.59 min Scan# 1956
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22

Tgt Ion	Ratio	Lower	Upper
164	100		
162	103.3	82.4	123.6
160	44.0	35.8	53.6





#39

1,2,4,5-Tetrachlorobenzene

Concen: 47.30 ng

RT: 12.77 min Scan# 1646

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD

Tgt Ion:216 Resp: 577245


Ion Ratio Lower Upper

216 100

214 77.3 0.0 0.0#

179 21.0 0.0 0.0#

108 15.1 0.0 0.0#






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 102.80 ng

RT: 12.73 min Scan# 1640

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion:237 Resp: 724352

Ion Ratio Lower Upper

237 100

235 63.7 42.0 82.0



272 12.6 0.0 33.4




Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 155.59 ng
 RT: 16.09 min Scan# 2210
 Delta R.T. -0.01 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22




Instrument :
 BNA_M
 ClientSampleId :
 GMW-7MSD


Tgt Ion	Ratio	Lower	Upper
330	100		
332	95.8	0.0	0.0#
141	33.1	0.0	0.0#



#42
 2,4,6-Trichlorophenol
 Concen: 49.44 ng
 RT: 13.03 min Scan# 1690
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22

Tgt Ion	Ratio	Lower	Upper
196	100		
198	94.9	76.3	114.5
200	30.7	25.6	38.4









#43
 2,4,5-Trichlorophenol
 Concen: 47.11 ng
 RT: 13.10 min Scan# 1703
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22




Instrument :
 BNA_M
 ClientSampleId :
 GMW-7MSD


Tgt Ion	Ratio	Lower	Upper
196	100		
198	100.2	79.4	119.0
97	45.3	26.8	40.2
132	21.8	18.6	28.0



#44
 2-Fluorobiphenyl
 Concen: 103.91 ng
 RT: 13.22 min Scan# 1722
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22

Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.2	28.5	42.7
170	23.0	18.8	28.2









#45
 1,1'-Biphenyl
 Concen: 47.10 ng
 RT: 13.43 min Scan# 1758
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22




Instrument :
 BNA_M
 ClientSampleId :
 GMW-7MSD


Tgt Ion:154 Resp: 1229238
 Ion Ratio Lower Upper
 154 100
 153 40.4 20.7 60.7
 76 11.1 0.0 30.9



#46
 2-Chloronaphthalene
 Concen: 45.87 ng
 RT: 13.47 min Scan# 1766
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22

Tgt Ion:162 Resp: 969243
 Ion Ratio Lower Upper
 162 100
 127 36.5 26.9 40.3
 164 33.3 26.8 40.2









#47
2-Nitroaniline
Concen: 32.36 ng
RT: 13.71 min Scan# 1806
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22




Instrument :
BNA_M
ClientSampleId :
GMW-7MSD


Tgt Ion: 65 Resp: 175839
Ion Ratio Lower Upper
65 100
92 61.2 59.1 88.7
138 88.9 93.9 140.9#



#48
Acenaphthylene
Concen: 47.36 ng
RT: 14.32 min Scan# 1910
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22

Tgt Ion: 152 Resp: 1479746
Ion Ratio Lower Upper
152 100
151 19.5 15.9 23.9
153 12.8 10.6 16.0





#49

Dimethylphthalate

Concen: 45.01 ng

RT: 14.06 min Scan# 1865

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion:163 Resp: 1274133

Ion Ratio Lower Upper

163 100

194 4.6 2.7 4.1#




164 10.3 8.2 12.2



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 48.84 ng

RT: 14.19 min Scan# 1888

Delta R.T. -0.01 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion:165 Resp: 262546

Ion Ratio Lower Upper

165 100

63 63.7 44.1 66.1



89 62.8 44.3 66.5




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BM0

Ion 89.00 (88.70 to 89.70): BM0





#51

Acenaphthene

Concen: 46.87 ng

RT: 14.66 min Scan# 1967

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion:154 Resp: 910447

Ion Ratio Lower Upper

154 100

153 113.9 90.0 135.0




152 52.0 42.6 63.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 29.27 ng

RT: 14.54 min Scan# 1947

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion:138 Resp: 146212

Ion Ratio Lower Upper

138 100

108 12.3 7.3 10.9#



92 123.5 76.6 114.8#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BM









#53
2,4-Dinitrophenol
Concen: 89.43 ng
RT: 14.73 min Scan# 1979
Delta R.T. -0.01 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22




Instrument :
BNA_M
ClientSampleId :
GMW-7MSD


Tgt Ion:184 Resp: 322307
Ion Ratio Lower Upper
184 100
63 62.7 69.2 103.8#
154 64.2 53.3 79.9



#54
Dibenzofuran
Concen: 50.78 ng
RT: 14.99 min Scan# 2024
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22

Tgt Ion:168 Resp: 1551045
Ion Ratio Lower Upper
168 100
139 35.7 27.3 40.9
169 13.1 10.6 16.0





#55

4-Nitrophenol

Concen: 35.17 ng

RT: 14.84 min Scan# 1999

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD





Tgt Ion:139 Resp: 141417

Ion Ratio Lower Upper

139 100

109 100.8 37.7 77.7#

65 118.4 49.9 89.9#



#56

2,4-Dinitrotoluene

Concen: 49.25 ng

RT: 14.97 min Scan# 2020

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Tgt Ion:165 Resp: 374530




Ion Ratio Lower Upper


165 100

63 50.2 52.4 78.6#

89 74.8 72.4 108.6

182 3.0 2.0 3.0#









#57
Fluorene
 Concen: 48.19 ng
 RT: 15.64 min Scan# 2134
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22




Instrument :
 BNA_M
 ClientSampleId :
 GMW-7MSD


Tgt Ion:166 Resp: 1279839
 Ion Ratio Lower Upper
 166 100
 165 97.9 79.0 118.4
 167 13.2 10.3 15.5



#58
2,3,4,6-Tetrachlorophenol
 Concen: 54.54 ng
 RT: 15.22 min Scan# 2062
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22

Tgt Ion:232 Resp: 368999
 Ion Ratio Lower Upper
 232 100
 131 35.7 0.0 0.0#
 130 2.5 0.0 0.0#
 166 27.4 0.0 0.0#





#59

Diethylphthalate

Concen: 45.29 ng

RT: 15.40 min Scan# 2094

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion:149 Resp: 1212238

Ion Ratio Lower Upper

149 100

177 24.4 18.3 27.5




150 12.4 9.8 14.8



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#60

4-Chlorophenyl-phenylether

Concen: 47.32 ng

RT: 15.63 min Scan# 2132

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion:204 Resp: 705657

Ion Ratio Lower Upper

204 100

206 33.0 26.6 39.8



141 56.0 45.2 67.8




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E









#61
4-Nitroaniline
Concen: 19.64 ng
RT: 15.70 min Scan# 2145
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22




Instrument :
BNA_M
ClientSampleId :
GMW-7MSD


Tgt Ion:138 Resp: 99614
Ion Ratio Lower Upper
138 100
92 63.7 16.7 56.7#
108 120.1 37.6 77.6#



#62
Azobenzene
Concen: 59.44 ng
RT: 15.92 min Scan# 2182
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22

Tgt Ion: 77 Resp: 1169860
Ion Ratio Lower Upper
77 100
182 28.1 6.5 46.5
105 17.4 3.8 43.8
51 32.7 5.5 45.5









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.34 min Scan# 2423
Delta R.T. -0.01 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22




Instrument :
BNA_M
ClientSampleId :
GMW-7MSD


Tgt Ion:188 Resp: 773050
Ion Ratio Lower Upper
188 100
94 5.8 8.7 13.1#
80 7.1 9.3 13.9#



#64
4,6-Dinitro-2-methylphenol
Concen: 46.86 ng
RT: 15.72 min Scan# 2148
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22

Tgt Ion:198 Resp: 221300
Ion Ratio Lower Upper
198 100
51 32.8 28.8 68.8
105 32.9 30.5 70.5





#65

n-Nitrosodiphenylamine

Concen: 46.02 ng

RT: 15.85 min Scan# 2170

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion:169 Resp: 1066190

Ion Ratio Lower Upper

169 100

168 68.2 53.9 80.9




167 35.4 28.9 43.3



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 47.72 ng

RT: 16.52 min Scan# 2284

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion:248 Resp: 472261

Ion Ratio Lower Upper

248 100

250 98.7 77.4 116.0



141 67.4 45.2 67.8




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E





#67

Hexachlorobenzene

Concen: 44.26 ng

RT: 16.62 min Scan# 2300

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion:284 Resp: 532187

Ion Ratio Lower Upper

284 100

142 26.9 20.4 30.6




249 31.9 23.8 35.6



Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#68

Atrazine

Concen: 49.63 ng

RT: 16.80 min Scan# 2331

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion:200 Resp: 435999

Ion Ratio Lower Upper

200 100

173 23.4 4.8 44.8



215 47.5 26.9 66.9




Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E





#69

Pentachlorophenol

Concen: 97.86 ng

RT: 16.98 min Scan# 2362

Delta R.T. -0.01 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion:266 Resp: 646281

Ion Ratio Lower Upper

266 100

268 62.9 52.0 78.0




264 63.2 49.0 73.4



Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E



#70

Phenanthrene

Concen: 48.03 ng

RT: 17.38 min Scan# 2430

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion:178 Resp: 2077181

Ion Ratio Lower Upper

178 100

176 19.2 15.2 22.8



179 15.3 12.1 18.1




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#71

Anthracene

Concen: 48.73 ng

RT: 17.47 min Scan# 2445

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD





Tgt Ion:178 Resp: 2091843

Ion Ratio Lower Upper

178 100

176 18.6 14.6 22.0

179 14.8 12.6 19.0



#72

Carbazole

Concen: 49.62 ng

RT: 17.76 min Scan# 2494

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22




Tgt Ion:167 Resp: 1885820


Ion Ratio Lower Upper

167 100

166 21.5 17.2 25.8

139 12.1 10.8 16.2





#73

Di-n-butylphthalate

Concen: 46.58 ng

RT: 18.27 min Scan# 2582

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion:149 Resp: 2098606

Ion Ratio Lower Upper




149 100

150 9.1 7.5 11.3

104 5.9 3.7 5.5#



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 48.83 ng

RT: 19.38 min Scan# 2770

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion:202 Resp: 2742335

Ion Ratio Lower Upper



202 100


101 6.6 0.0 28.7

203 17.6 0.0 37.2



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 21.50 min Scan# 3130

Delta R.T. -0.01 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion:240 Resp: 1221904

Ion Ratio Lower Upper

240 100

120 5.4 8.2 12.2#




236 26.0 20.0 30.0



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 8.60 ng

RT: 19.59 min Scan# 2806

Delta R.T. -0.01 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion:184 Resp: 193772

Ion Ratio Lower Upper

184 100

185 14.1 11.3 16.9



183 14.4 8.9 13.3#




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#77

Pyrene

Concen: 43.40 ng

RT: 19.75 min Scan# 2833

Delta R.T. -0.01 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :


BNA_M

ClientSampleId :

GMW-7MSD

Tgt Ion:202 Resp: 2843312




Ion	Ratio	Lower	Upper
202	100		
200	20.6	16.9	25.3
203	17.5	14.1	21.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 83.68 ng

RT: 19.94 min Scan# 2865


Delta R.T. -0.01 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Tgt Ion:244 Resp: 4497083



Ion	Ratio	Lower	Upper
244	100		
212	7.2	4.9	7.3
122	6.1	6.2	9.4




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#79

Butylbenzylphthalate

Concen: 43.71 ng

RT: 20.62 min Scan# 2981

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD


Tgt Ion:149 Resp: 1059952

Ion Ratio Lower Upper

149 100

91 73.2 50.1 75.1




206 26.3 16.2 24.2#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BM0

Ion 206.00 (205.70 to 206.70): E



#80

Benzo(a)anthracene

Concen: 47.28 ng

RT: 21.48 min Scan# 3127

Delta R.T. -0.02 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22


Tgt Ion:228 Resp: 3183739

Ion Ratio Lower Upper

228 100

226 26.7 21.7 32.5



229 19.3 16.0 24.0




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#81
 3,3'-Dichlorobenzidine
 Concen: 7.60 ng
 RT: 21.42 min Scan# 3117
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22




Instrument :
 BNA_M
 ClientSampleId :
 GMW-7MSD


Tgt Ion	Ratio	Lower	Upper
252	100		
254	63.0	51.9	77.9
126	6.5	9.8	14.8



#82
 Chrysene
 Concen: 45.65 ng
 RT: 21.53 min Scan# 3136
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.8	24.1	36.1
229	19.6	15.5	23.3









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 43.34 ng
 RT: 21.36 min Scan# 3107
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22




Instrument :
 BNA_M
 ClientSampleId :
 GMW-7MSD


Tgt Ion:149 Resp: 1571774
 Ion Ratio Lower Upper
 149 100
 167 29.8 22.4 33.6
 279 5.6 3.4 5.0#



#84
 Di-n-octyl phthalate
 Concen: 44.45 ng
 RT: 22.27 min Scan# 3262
 Delta R.T. -0.02 min
 Lab File: BM010131.D
 Acq: 23 May 2017 08:22

Tgt Ion:149 Resp: 2885912
 Ion Ratio Lower Upper
 149 100
 167 1.8 1.2 1.8
 43 8.0 7.3 10.9









#85
Indeno(1,2,3-cd)pyrene
Concen: 51.54 ng
RT: 26.29 min Scan# 3944
Delta R.T. -0.04 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22




Instrument :
BNA_M
ClientSampleId :
GMW-7MSD


Tgt Ion:276 Resp: 4827366
Ion Ratio Lower Upper
276 100
138 13.8 0.0 0.0#
277 25.6 0.0 0.0#



#86
Perylene-d12
Concen: 20.00 ng
RT: 23.86 min Scan# 3532
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22

Tgt Ion:264 Resp: 1334349
Ion Ratio Lower Upper
264 100
260 23.2 18.6 27.8
265 22.3 17.3 25.9









#87
Benzo(b)fluoranthene
Concen: 49.37 ng
RT: 23.14 min Scan# 3409
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22




Instrument :
BNA_M
ClientSampleId :
GMW-7MSD


Tgt Ion:252 Resp: 3855624
Ion Ratio Lower Upper
252 100
253 22.0 18.0 27.0
125 5.3 9.9 14.9#



#88
Benzo(k)fluoranthene
Concen: 47.15 ng
RT: 23.19 min Scan# 3417
Delta R.T. -0.02 min
Lab File: BM010131.D
Acq: 23 May 2017 08:22

Tgt Ion:252 Resp: 3607522
Ion Ratio Lower Upper
252 100
253 21.9 17.2 25.8
125 5.1 8.1 12.1#





#89

Benzo(a)pyrene

Concen: 48.51 ng

RT: 23.76 min Scan# 3514

Delta R.T. -0.03 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD





Tgt Ion:252 Resp: 3610221

Ion Ratio Lower Upper

252 100

253 21.2 17.6 26.4

125 5.6 7.4 11.2#



#90

Dibenzo(a,h)anthracene

Concen: 49.42 ng

RT: 26.30 min Scan# 3946

Delta R.T. -0.04 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22




Tgt Ion:278 Resp: 4112120


Ion Ratio Lower Upper

278 100

139 9.4 13.4 20.0#

279 23.8 19.0 28.4





#91

Benzo(g,h,i)perylene

Concen: 48.79 ng

RT: 27.04 min Scan# 4073

Delta R.T. -0.04 min

Lab File: BM010131.D

Acq: 23 May 2017 08:22

Instrument :

BNA_M

ClientSampleId :

GMW-7MSD

Tgt Ion:276 Resp: 3977816

Ion Ratio Lower Upper

276 100

277 23.2 18.5 27.7

138 11.6 19.0 28.6

