

Data Path : Z:\HPCHEM1\BNA\_M\DATA\BM052218\  
 Data File : BM015238.D  
 Acq On : 22 May 2018 13:51  
 Operator : SJ/JU  
 Sample : SSTDICC050  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC050

Quant Time: May 22 14:22:35 2018  
 Quant Method : Z:\HPCHEM1\BNA\_M\METHODS\8270-BM052218.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue May 22 13:50:55 2018  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|---------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 8.41  | 152  | 201329   | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8        | 11.25 | 136  | 810219   | 20.00 | ng    | 0.00     |
| 38) Acenaphthene-d10      | 15.02 | 164  | 411491   | 20.00 | ng    | 0.00     |
| 63) Phenanthrene-d10      | 17.74 | 188  | 843914   | 20.00 | ng    | 0.00     |
| 75) Chrysene-d12          | 21.83 | 240  | 880185   | 20.00 | ng    | 0.00     |
| 86) Perylene-d12          | 24.27 | 264  | 901546   | 20.00 | ng    | 0.00     |

## System Monitoring Compounds

|                          |       |     |         |        |    |      |
|--------------------------|-------|-----|---------|--------|----|------|
| 5) 2-Fluorophenol        | 5.89  | 112 | 1279660 | 110.78 | ng | 0.00 |
| 7) Phenol-d6             | 7.55  | 99  | 1653219 | 101.06 | ng | 0.00 |
| 23) Nitrobenzene-d5      | 9.60  | 82  | 1523754 | 84.00  | ng | 0.00 |
| 41) 2,4,6-Tribromophenol | 16.49 | 330 | 550941  | 94.36  | ng | 0.00 |
| 44) 2-Fluorobiphenyl     | 13.65 | 172 | 3317416 | 100.64 | ng | 0.00 |
| 78) Terphenyl-d14        | 20.29 | 244 | 3960810 | 87.93  | ng | 0.00 |

## Target Compounds

|                                |       |     |         |        |    | Qvalue |
|--------------------------------|-------|-----|---------|--------|----|--------|
| 2) 1,4-Dioxane                 | 3.62  | 88  | 258758  | 54.515 | ng | # 100  |
| 3) Pyridine                    | 4.07  | 79  | 695769  | 52.935 | ng | # 88   |
| 4) n-Nitrosodimethylamine      | 3.98  | 42  | 369423  | 44.121 | ng | # 69   |
| 6) Aniline                     | 7.72  | 93  | 1017420 | 48.784 | ng | # 97   |
| 8) 2-Chlorophenol              | 7.96  | 128 | 713079  | 54.315 | ng | # 97   |
| 9) Benzaldehyde                | 7.53  | 77  | 464535  | 39.456 | ng | # 95   |
| 10) Phenol                     | 7.58  | 94  | 836011  | 51.765 | ng | # 78   |
| 11) bis(2-Chloroethyl)ether    | 7.82  | 93  | 651497  | 51.033 | ng | # 91   |
| 12) 1,3-Dichlorobenzene        | 8.29  | 146 | 800401  | 48.777 | ng | # 96   |
| 13) 1,4-Dichlorobenzene        | 8.45  | 146 | 808762  | 49.058 | ng | # 96   |
| 14) 1,2-Dichlorobenzene        | 8.77  | 146 | 767774  | 46.279 | ng | # 96   |
| 15) Benzyl Alcohol             | 8.65  | 79  | 603061  | 39.784 | ng | # 95   |
| 16) 2,2'-oxybis(1-Chloropropan | 8.94  | 45  | 1009725 | 77.760 | ng | # 96   |
| 17) 2-Methylphenol             | 8.85  | 107 | 555312  | 44.927 | ng | # 94   |
| 18) Hexachloroethane           | 9.50  | 117 | 294554  | 41.336 | ng | # 92   |
| 19) n-Nitroso-di-n-propylamine | 9.22  | 70  | 525389  | 37.281 | ng | # 89   |
| 20) 3+4-Methylphenols          | 9.18  | 107 | 738723  | 41.267 | ng | # 94   |
| 22) Acetophenone               | 9.25  | 105 | 985279  | 42.292 | ng | # 94   |
| 24) Nitrobenzene               | 9.64  | 77  | 786711  | 42.124 | ng | # 96   |
| 25) Isophorone                 | 10.16 | 82  | 1331123 | 44.298 | ng | # 96   |
| 26) 2-Nitrophenol              | 10.35 | 139 | 358890  | 50.941 | ng | # 81   |
| 27) 2,4-Dimethylphenol         | 10.40 | 122 | 651880  | 60.708 | ng | # 92   |
| 28) bis(2-Chloroethoxy)methane | 10.63 | 93  | 842889  | 54.696 | ng | # 99   |
| 29) 2,4-Dichlorophenol         | 10.89 | 162 | 619691  | 51.647 | ng | # 98   |
| 30) 1,2,4-Trichlorobenzene     | 11.10 | 180 | 708315  | 44.781 | ng | # 99   |
| 31) Naphthalene                | 11.30 | 128 | 2119068 | 47.766 | ng | # 100  |
| 32) Benzoic acid               | 10.55 | 122 | 375148  | 44.282 | ng | # 93   |
| 33) 4-Chloroaniline            | 11.40 | 127 | 865241  | 46.920 | ng | # 91   |
| 34) Hexachlorobutadiene        | 11.56 | 225 | 418035  | 38.528 | ng | # 96   |
| 35) Caprolactam                | 12.19 | 113 | 174414  | 40.648 | ng | # 91   |
| 36) 4-Chloro-3-methylphenol    | 12.50 | 107 | 631988  | 42.514 | ng | # 91   |
| 37) 2-Methylnaphthalene        | 12.87 | 142 | 1472176 | 44.063 | ng | # 98   |
| 39) 1,2,4,5-Tetrachlorobenzene | 13.23 | 216 | 734773  | 50.961 | ng | # 100  |
| 40) Hexachlorocyclopentadiene  | 13.20 | 237 | 443375  | 61.060 | ng | # 99   |

Data Path : Z:\HPCHEM1\BNA\_M\DATA\BM052218\  
 Data File : BM015238.D  
 Acq On : 22 May 2018 13:51  
 Operator : SJ/JU  
 Sample : SSTDICC050  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 1


Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC050

Quant Time: May 22 14:22:35 2018  
 Quant Method : Z:\HPCHEM1\BNA\_M\METHODS\8270-BM052218.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue May 22 13:50:55 2018  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|--------------------------------|-------|------|----------|--------|-------|----------|
| 42) 2,4,6-Trichlorophenol      | 13.47 | 196  | 462083   | 54.554 | ng    | 98       |
| 43) 2,4,5-Trichlorophenol      | 13.55 | 196  | 492576   | 52.672 | ng    | # 91     |
| 45) 1,1'-Biphenyl              | 13.86 | 154  | 1811869  | 50.831 | ng    | 98       |
| 46) 2-Chloronaphthalene        | 13.92 | 162  | 1396657  | 51.617 | ng    | 94       |
| 47) 2-Nitroaniline             | 14.12 | 65   | 411370   | 41.315 | ng    | 90       |
| 48) Acenaphthylene             | 14.75 | 152  | 2127274  | 47.946 | ng    | 99       |
| 49) Dimethylphthalate          | 14.47 | 163  | 1665104  | 45.201 | ng    | 99       |
| 50) 2,6-Dinitrotoluene         | 14.60 | 165  | 364563   | 48.229 | ng    | 99       |
| 51) Acenaphthene               | 15.09 | 154  | 1298455  | 47.477 | ng    | 99       |
| 52) 3-Nitroaniline             | 14.93 | 138  | 375037   | 47.965 | ng    | 83       |
| 53) 2,4-Dinitrophenol          | 15.13 | 184  | 189181   | 53.024 | ng    | # 87     |
| 54) Dibenzofuran               | 15.42 | 168  | 2013421  | 47.115 | ng    | 94       |
| 55) 4-Nitrophenol              | 15.22 | 139  | 309013   | 57.505 | ng    | # 51     |
| 56) 2,4-Dinitrotoluene         | 15.37 | 165  | 483362   | 41.435 | ng    | # 78     |
| 57) Fluorene                   | 16.06 | 166  | 1612689  | 42.895 | ng    | 99       |
| 58) 2,3,4,6-Tetrachlorophenol  | 15.63 | 232  | 420347   | 45.624 | ng    | # 100    |
| 59) Diethylphthalate           | 15.81 | 149  | 1660716  | 40.613 | ng    | 98       |
| 60) 4-Chlorophenyl-phenylether | 16.04 | 204  | 815431   | 41.725 | ng    | 97       |
| 61) 4-Nitroaniline             | 16.08 | 138  | 394153   | 50.379 | ng    | # 70     |
| 62) Azobenzene                 | 16.33 | 77   | 1611863  | 37.955 | ng    | 91       |
| 64) 4,6-Dinitro-2-methylphenol | 16.13 | 198  | 249976   | 46.977 | ng    | 94       |
| 65) n-Nitrosodiphenylamine     | 16.25 | 169  | 1391556  | 52.232 | ng    | 100      |
| 66) 4-Bromophenyl-phenylether  | 16.93 | 248  | 484177   | 48.950 | ng    | # 92     |
| 67) Hexachlorobenzene          | 17.05 | 284  | 546920   | 50.363 | ng    | # 92     |
| 68) Atrazine                   | 17.18 | 200  | 450381   | 45.071 | ng    | 98       |
| 69) Pentachlorophenol          | 17.39 | 266  | 316268   | 48.645 | ng    | 99       |
| 70) Phenanthrene               | 17.78 | 178  | 2428686  | 47.707 | ng    | 100      |
| 71) Anthracene                 | 17.87 | 178  | 2482265  | 50.042 | ng    | 99       |
| 72) Carbazole                  | 18.13 | 167  | 2218467  | 47.579 | ng    | 99       |
| 73) Di-n-butylphthalate        | 18.66 | 149  | 2682220  | 44.950 | ng    | 99       |
| 74) Fluoranthene               | 19.75 | 202  | 2730136  | 45.553 | ng    | 97       |
| 76) Benzidine                  | 19.92 | 184  | 1434294  | 54.717 | ng    | 99       |
| 77) Pyrene                     | 20.11 | 202  | 2773421  | 45.356 | ng    | 100      |
| 79) Butylbenzylphthalate       | 20.96 | 149  | 1220804  | 44.332 | ng    | 93       |
| 80) Benzo(a)anthracene         | 21.82 | 228  | 2798956  | 48.939 | ng    | 100      |
| 81) 3,3'-Dichlorobenzidine     | 21.74 | 252  | 1068069  | 49.966 | ng    | 99       |
| 82) Chrysene                   | 21.87 | 228  | 2635484  | 47.503 | ng    | 100      |
| 83) Bis(2-ethylhexyl)phthalate | 21.70 | 149  | 1778296  | 44.271 | ng    | 99       |
| 84) Di-n-octyl phthalate       | 22.63 | 149  | 3014166  | 54.204 | ng    | # 95     |
| 85) Indeno(1,2,3-cd)pyrene     | 26.80 | 276  | 3288894  | 75.402 | ng    | # 90     |
| 87) Benzo(b)fluoranthene       | 23.53 | 252  | 2909241  | 46.506 | ng    | 97       |
| 88) Benzo(k)fluoranthene       | 23.58 | 252  | 2735285  | 45.995 | ng    | 99       |
| 89) Benzo(a)pyrene             | 24.17 | 252  | 2747102  | 49.831 | ng    | # 98     |
| 90) Dibenzo(a,h)anthracene     | 26.81 | 278  | 2792437  | 61.983 | ng    | 99       |
| 91) Benzo(g,h,i)perylene       | 27.58 | 276  | 2701931  | 64.147 | ng    | 97       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1




1,4-Dichlorobenzene-d4  
Concen: 20.000 ng  
RT: 8.41 min Scan# 871  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDIC050

Tgt Ion: 152 Resp: 201329  
Ion Ratio Lower Upper  
152 100  
150 157.8 119.5 179.3  
115 59.5 43.0 64.4




Abundance Ion 152.00 (151.70 to 152.70): E  
Ion 150.00 (149.70 to 150.70): E  
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane  
Concen: 54.515 ng  
RT: 3.62 min Scan# 57  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51

Tgt Ion: 88 Resp: 258758  
Ion Ratio Lower Upper  
88 100  
58 76.0 0.0 0.0#  
43 34.0 0.0 0.0#



Abundance Ion 88.00 (87.70 to 88.70): BM0  
Ion 58.00 (57.70 to 58.70): BM0  
Ion 43.00 (42.70 to 43.70): BM0





#3

Pyridine

Concen: 52.935 ng

RT: 4.07 min Scan# 133

Delta R.T. -0.01 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDICC050


Tgt Ion: 79 Resp: 695769

Ion Ratio Lower Upper

79 100

52 73.7 51.1 76.7




51 38.8 26.1 39.1



Abundance Ion 79.00 (78.70 to 79.70): BM015237.D (-130) (-)

Ion 52.00 (51.70 to 52.70): BM015237.D (-130) (-)

Ion 51.00 (50.70 to 51.70): BM015237.D (-130) (-)



#4

n-Nitrosodimethylamine

Concen: 44.121 ng

RT: 3.98 min Scan# 117

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion: 42 Resp: 369423

Ion Ratio Lower Upper

42 100

74 108.6 119.3 178.9#



44 6.2 5.4 8.2




Abundance Ion 42.00 (41.70 to 42.70): BM015237.D (-112) (-)

Ion 74.00 (73.70 to 74.70): BM015237.D (-112) (-)

Ion 44.00 (43.70 to 44.70): BM015237.D (-112) (-)





#5

2-Fluorophenol

Concen: 110.785 ng

RT: 5.89 min Scan# 443

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDICC050


Tgt Ion:112 Resp: 1279660

Ion Ratio Lower Upper

112 100

64 61.6 38.6 57.8#




63 33.0 19.3 28.9#



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BM0

Ion 63.00 (62.70 to 63.70): BM0



#6

Aniline

Concen: 48.784 ng

RT: 7.72 min Scan# 754

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion: 93 Resp: 1017420

Ion Ratio Lower Upper

93 100

66 41.2 30.8 46.2



65 20.6 16.0 24.0






Abundance Ion 93.00 (92.70 to 93.70): BM0

Ion 66.00 (65.70 to 66.70): BM0

Ion 65.00 (64.70 to 65.70): BM0





#7

Phenol-d6

Concen: 101.057 ng

RT: 7.55 min Scan# 725

Delta R.T. 0.01 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050





Tgt Ion: 99 Resp: 1653219

Ion Ratio Lower Upper

99 100

42 22.0 11.0 16.4#

71 36.2 23.4 35.2#



#8

2-Chlorophenol

Concen: 54.315 ng

RT: 7.96 min Scan# 795

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion: 128 Resp: 713079


Ion Ratio Lower Upper

128 100

130 32.5 12.0 52.0

64 47.6 24.9 64.9





#9

Benzaldehyde

Concen: 39.456 ng

RT: 7.53 min Scan# 722

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050





Tgt Ion: 77 Resp: 464535

Ion Ratio Lower Upper

77 100

105 93.4 78.8 118.8

106 89.6 73.7 113.7



#10

Phenol

Concen: 51.765 ng

RT: 7.58 min Scan# 730

Delta R.T. 0.01 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51




Tgt Ion: 94 Resp: 836011


Ion Ratio Lower Upper

94 100

65 30.0 18.8 58.8

66 39.6 39.9 79.9#









#11  
 bis(2-Chloroethyl)ether  
 Concen: 51.033 ng  
 RT: 7.82 min Scan# 770  
 Delta R.T. 0.01 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC050


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 93    | Resp: | 651497 |
| Ion      | Ratio | Lower | Upper  |
| 93       | 100   |       |        |
| 63       | 80.1  | 49.5  | 89.5   |
| 95       | 33.1  | 13.5  | 53.5   |



#12  
 1,3-Dichlorobenzene  
 Concen: 48.777 ng  
 RT: 8.29 min Scan# 851  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 146   | Resp: | 800401 |
| Ion      | Ratio | Lower | Upper  |
| 146      | 100   |       |        |
| 148      | 63.7  | 50.9  | 76.3   |
| 75       | 32.9  | 21.4  | 32.2#  |





#13

1,4-Dichlorobenzene

Concen: 49.058 ng

RT: 8.45 min Scan# 877

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050





Tgt Ion:146 Resp: 808762

Ion Ratio Lower Upper

146 100

148 63.8 51.9 77.9

111 41.1 29.2 43.8



#14

1,2-Dichlorobenzene

Concen: 46.279 ng

RT: 8.77 min Scan# 932

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51




Tgt Ion:146 Resp: 767774




Ion Ratio Lower Upper

146 100

148 64.8 52.3 78.5

111 43.2 30.6 45.8





#15

Benzyl Alcohol

Concen: 39.784 ng

RT: 8.65 min Scan# 912

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050





Tgt Ion: 79 Resp: 603061

Ion Ratio Lower Upper

79 100

108 74.0 65.0 97.4

77 65.0 53.0 79.6



#16

2,2'-oxybis(1-Chloropropane)

Concen: 77.760 ng

RT: 8.94 min Scan# 961

Delta R.T. 0.01 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion: 45 Resp: 1009725


Ion Ratio Lower Upper

45 100

77 13.6 0.0 35.2

79 10.3 0.0 32.0





#17

2-Methylphenol

Concen: 44.927 ng

RT: 8.85 min Scan# 946

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51





Instrument :

BNA\_M

ClientSampleId :

SSTDIC050

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 107   | Resp: | 555312 |
| Ion      | Ratio | Lower | Upper  |
| 107      | 100   |       |        |
| 108      | 111.7 | 89.8  | 134.8  |
| 77       | 48.9  | 32.6  | 48.8#  |
| 79       | 48.3  | 32.8  | 49.2   |



#18

Hexachloroethane

Concen: 41.336 ng




RT: 9.50 min Scan# 1057




Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 117   | Resp: | 294554 |
| Ion      | Ratio | Lower | Upper  |
| 117      | 100   |       |        |
| 119      | 97.0  | 79.2  | 118.8  |
| 201      | 101.1 | 69.8  | 104.6  |





#19

n-Nitroso-di-n-propylamine

Concen: 37.281 ng

RT: 9.22 min Scan# 1009

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050

Tgt Ion: 70 Resp: 525389





Ion Ratio Lower Upper

70 100

42 67.0 44.5 66.7#

101 9.5 7.4 11.2

130 20.6 15.3 22.9



#20

3+4-Methylphenols

Concen: 41.267 ng

RT: 9.18 min Scan# 1002

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Tgt Ion: 107 Resp: 738723


Ion Ratio Lower Upper


107 100

108 86.9 73.2 113.2

77 34.9 10.7 50.7

79 29.2 9.6 49.6









#21  
Naphthalene-d8  
Concen: 20.000 ng  
RT: 11.25 min Scan# 1353  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDIC050

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 136   | Resp: | 810219 |
| Ion      | Ratio | Lower | Upper  |
| 136      | 100   |       |        |
| 137      | 10.3  | 8.7   | 13.1   |
| 54       | 8.5   | 4.6   | 6.8#   |
| 68       | 6.2   | 3.7   | 5.5#   |




Abundance Ion 136.00 (135.70 to 136.70): E  
Ion 137.00 (136.70 to 137.70): E  
Ion 54.00 (53.70 to 54.70): BM0  
Ion 68.00 (67.70 to 68.70): BM0






#22  
Acetophenone  
Concen: 42.292 ng  
RT: 9.25 min Scan# 1013  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 105   | Resp: | 985279 |
| Ion      | Ratio | Lower | Upper  |
| 105      | 100   |       |        |
| 71       | 1.0   | 0.6   | 1.0#   |
| 51       | 30.3  | 21.6  | 32.4   |
| 120      | 22.4  | 16.1  | 24.1   |



Abundance Ion 105.00 (104.70 to 105.70): E  
Ion 71.00 (70.70 to 71.70): BM0  
Ion 51.00 (50.70 to 51.70): BM0  
Ion 120.00 (119.70 to 120.70): E









#23  
 Nitrobenzene-d5  
 Concen: 84.005 ng  
 RT: 9.60 min Scan# 1073  
 Delta R.T. 0.01 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC050

Tgt Ion: 82 Resp: 1523754  
 Ion Ratio Lower Upper  
 82 100  
 128 43.3 32.8 49.2  
 54 51.8 40.6 60.8




Abundance Ion 82.00 (81.70 to 82.70): BM015238.D (-1055) (-)  
 Ion 128.00 (127.70 to 128.70): BM015238.D (-1055) (-)  
 Ion 54.00 (53.70 to 54.70): BM015238.D (-1055) (-)






#24  
 Nitrobenzene  
 Concen: 42.124 ng  
 RT: 9.64 min Scan# 1080  
 Delta R.T. 0.01 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

Tgt Ion: 77 Resp: 786711  
 Ion Ratio Lower Upper  
 77 100  
 123 43.6 32.6 49.0  
 65 14.1 10.9 16.3



Abundance Ion 77.00 (76.70 to 77.70): BM015238.D (-1062) (-)  
 Ion 123.00 (122.70 to 123.70): BM015238.D (-1062) (-)  
 Ion 65.00 (64.70 to 65.70): BM015238.D (-1062) (-)









#25  
Isophorone  
Concen: 44.298 ng  
RT: 10.16 min Scan# 1169  
Delta R.T. 0.01 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDICC050


Tgt Ion: 82 Resp: 1331123  
Ion Ratio Lower Upper  
82 100  
95 7.4 5.8 8.6  
138 18.2 16.3 24.5



#26  
2-Nitrophenol  
Concen: 50.941 ng  
RT: 10.35 min Scan# 1201  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51

Tgt Ion: 139 Resp: 358890  
Ion Ratio Lower Upper  
139 100  
109 31.5 20.1 30.1#  
65 53.4 31.5 47.3#





#27

2,4-Dimethylphenol

Concen: 60.708 ng

RT: 10.40 min Scan# 1209

Delta R.T. 0.01 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050


Tgt Ion:122 Resp: 651880

Ion Ratio Lower Upper

122 100

107 118.0 84.7 127.1




121 59.4 46.6 69.8



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 54.696 ng

RT: 10.63 min Scan# 1249

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion: 93 Resp: 842889

Ion Ratio Lower Upper

93 100

95 31.5 25.5 38.3



123 11.7 8.6 13.0




Abundance Ion 93.00 (92.70 to 93.70): BM0

Ion 95.00 (94.70 to 95.70): BM0

Ion 123.00 (122.70 to 123.70): E





#29

2,4-Dichlorophenol

Concen: 51.647 ng

RT: 10.89 min Scan# 1292

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050


Tgt Ion:162 Resp: 619691

Ion Ratio Lower Upper

162 100

164 64.3 42.8 82.8




98 36.1 14.5 54.5



Abundance Ion 162.00 (161.70 to 162.70): E

Ion 164.00 (163.70 to 164.70): E

Ion 98.00 (97.70 to 98.70): BM



#30

1,2,4-Trichlorobenzene

Concen: 44.781 ng

RT: 11.10 min Scan# 1329

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion:180 Resp: 708315

Ion Ratio Lower Upper

180 100

182 96.7 77.6 116.4



145 30.3 23.4 35.2




Abundance Ion 180.00 (179.70 to 180.70): E

Ion 182.00 (181.70 to 182.70): E

Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 47.766 ng

RT: 11.30 min Scan# 1362

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050





Tgt Ion:128 Resp: 2119068

Ion Ratio Lower Upper

128 100

129 10.9 8.9 13.3

127 13.0 10.4 15.6



#32

Benzoic acid

Concen: 44.282 ng

RT: 10.55 min Scan# 1234

Delta R.T. 0.03 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51




Tgt Ion:122 Resp: 375148


Ion Ratio Lower Upper

122 100

105 125.7 99.9 139.9

77 101.7 73.7 113.7









#33  
4-Chloroaniline  
Concen: 46.920 ng  
RT: 11.40 min Scan# 1380  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDIC050

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 127     | 100   |       |       |
| 129     | 33.0  | 25.3  | 37.9  |
| 65      | 31.4  | 17.6  | 26.4  |
| 92      | 19.7  | 13.1  | 19.7  |




Abundance Ion 127.00 (126.70 to 127.70): E  
Ion 129.00 (128.70 to 129.70): E  
Ion 65.00 (64.70 to 65.70): BM0  
Ion 92.00 (91.70 to 92.70): BM0






#34  
Hexachlorobutadiene  
Concen: 38.528 ng  
RT: 11.56 min Scan# 1407  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 225     | 100   |       |       |
| 223     | 64.4  | 47.9  | 71.9  |
| 227     | 64.2  | 50.1  | 75.1  |



Abundance Ion 225.00 (224.70 to 225.70): E  
Ion 223.00 (222.70 to 223.70): E  
Ion 227.00 (226.70 to 227.70): E









#35  
 Caprolactam  
 Concen: 40.648 ng  
 RT: 12.19 min Scan# 1514  
 Delta R.T. 0.02 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC050


Tgt Ion:113 Resp: 174414  
 Ion Ratio Lower Upper  
 113 100  
 55 178.0 145.9 185.9  
 56 131.1 101.0 141.0



#36  
 4-Chloro-3-methylphenol  
 Concen: 42.514 ng  
 RT: 12.50 min Scan# 1566  
 Delta R.T. 0.01 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

Tgt Ion:107 Resp: 631988  
 Ion Ratio Lower Upper  
 107 100  
 144 26.5 23.3 34.9  
 142 81.3 72.6 109.0









#37  
 2-Methylnaphthalene  
 Concen: 44.063 ng  
 RT: 12.87 min Scan# 1630  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC050


Tgt Ion:142 Resp: 1472176  
 Ion Ratio Lower Upper  
 142 100  
 141 90.1 71.9 107.9  
 115 35.6 25.4 38.0



#38  
 Acenaphthene-d10  
 Concen: 20.000 ng  
 RT: 15.02 min Scan# 1995  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

Tgt Ion:164 Resp: 411491  
 Ion Ratio Lower Upper  
 164 100  
 162 103.5 82.4 123.6  
 160 45.9 35.8 53.6









#39  
 1,2,4,5-Tetrachlorobenzene  
 Concen: 50.961 ng  
 RT: 13.23 min Scan# 1691  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC050

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 216   | Resp: | 734773 |
| Ion      | Ratio | Lower | Upper  |
| 216      | 100   |       |        |
| 214      | 78.5  | 0.0   | 0.0#   |
| 179      | 21.9  | 0.0   | 0.0#   |
| 108      | 19.7  | 0.0   | 0.0#   |




Abundance Ion 216.00 (215.70 to 216.70): E  
 Ion 214.00 (213.70 to 214.70): E  
 Ion 179.00 (178.70 to 179.70): E  
 Ion 108.00 (107.70 to 108.70): E






#40  
 Hexachlorocyclopentadiene  
 Concen: 61.060 ng  
 RT: 13.20 min Scan# 1686  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 237   | Resp: | 443375 |
| Ion      | Ratio | Lower | Upper  |
| 237      | 100   |       |        |
| 235      | 62.9  | 42.0  | 82.0   |
| 272      | 13.4  | 0.0   | 33.4   |



Abundance Ion 236.80 (236.50 to 237.50): E  
 Ion 234.90 (234.60 to 235.60): E  
 Ion 271.75 (271.45 to 272.45): E









#41  
 2,4,6-Tribromophenol  
 Concen: 94.356 ng  
 RT: 16.49 min Scan# 2245  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC050

Tgt Ion: 330 Resp: 550941  
 Ion Ratio Lower Upper  
 330 100  
 332 96.3 0.0 0.0#  
 141 37.1 0.0 0.0#




Abundance Ion 329.80 (329.50 to 330.50): E  
 Ion 331.80 (331.50 to 332.50): E  
 Ion 141.00 (140.70 to 141.70): E






#42  
 2,4,6-Trichlorophenol  
 Concen: 54.554 ng  
 RT: 13.47 min Scan# 1731  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

Tgt Ion: 196 Resp: 462083  
 Ion Ratio Lower Upper  
 196 100  
 198 97.2 76.3 114.5  
 200 30.9 25.6 38.4



Abundance Ion 196.00 (195.70 to 196.70): E  
 Ion 198.00 (197.70 to 198.70): E  
 Ion 200.00 (199.70 to 200.70): E









#43  
 2,4,5-Trichlorophenol  
 Concen: 52.672 ng  
 RT: 13.55 min Scan# 1744  
 Delta R.T. 0.01 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC050


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 196     | 100   |       |       |
| 198     | 96.6  | 79.4  | 119.0 |
| 97      | 49.6  | 26.8  | 40.2  |
| 132     | 26.9  | 18.6  | 28.0  |



#44  
 2-Fluorobiphenyl  
 Concen: 100.638 ng  
 RT: 13.65 min Scan# 1762  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 172     | 100   |       |       |
| 171     | 35.5  | 28.5  | 42.7  |
| 170     | 23.5  | 18.8  | 28.2  |









#45  
 1,1'-Biphenyl  
 Concen: 50.831 ng  
 RT: 13.86 min Scan# 1798  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC050


Tgt Ion:154 Resp: 1811869  
 Ion Ratio Lower Upper  
 154 100  
 153 41.0 20.7 60.7  
 76 13.5 0.0 30.9



#46  
 2-Chloronaphthalene  
 Concen: 51.617 ng  
 RT: 13.92 min Scan# 1807  
 Delta R.T. 0.01 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

Tgt Ion:162 Resp: 1396657  
 Ion Ratio Lower Upper  
 162 100  
 127 39.2 26.9 40.3  
 164 32.7 26.8 40.2





#47

2-Nitroaniline

Concen: 41.315 ng

RT: 14.12 min Scan# 1841

Delta R.T. 0.01 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050


Tgt Ion: 65 Resp: 411370

Ion Ratio Lower Upper

65 100

92 66.0 59.1 88.7




138 105.5 93.9 140.9



Abundance Ion 65.00 (64.70 to 65.70): BM015238.D (-1830) (-)

Ion 92.00 (91.70 to 92.70): BM015238.D (-1830) (-)

Ion 138.00 (137.70 to 138.70): BM015238.D (-1830) (-)



#48

Acenaphthylene

Concen: 47.946 ng

RT: 14.75 min Scan# 1949

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion: 152 Resp: 2127274

Ion Ratio Lower Upper

152 100

151 20.3 15.9 23.9



153 13.1 10.6 16.0




Abundance Ion 152.00 (151.70 to 152.70): BM015238.D (-1932) (-)

Ion 151.00 (150.70 to 151.70): BM015238.D (-1932) (-)

Ion 153.00 (152.70 to 153.70): BM015238.D (-1932) (-)





#49

Dimethylphthalate

Concen: 45.201 ng

RT: 14.47 min Scan# 1901

Delta R.T. 0.01 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050


Tgt Ion:163 Resp: 1665104

Ion Ratio Lower Upper

163 100

194 4.0 2.7 4.1




164 10.0 8.2 12.2



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 48.229 ng

RT: 14.60 min Scan# 1923

Delta R.T. 0.01 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion:165 Resp: 364563

Ion Ratio Lower Upper

165 100

63 56.9 44.1 66.1



89 55.2 44.3 66.5




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BM0

Ion 89.00 (88.70 to 89.70): BM0





#51

Acenaphthene

Concen: 47.477 ng

RT: 15.09 min Scan# 2006

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050


Tgt Ion:154 Resp: 1298455

Ion Ratio Lower Upper

154 100

153 114.0 90.0 135.0




152 53.2 42.6 63.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 47.965 ng

RT: 14.93 min Scan# 1979

Delta R.T. 0.01 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion:138 Resp: 375037

Ion Ratio Lower Upper

138 100

108 10.6 7.3 10.9



92 113.8 76.6 114.8




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BM









#53  
2,4-Dinitrophenol  
Concen: 53.024 ng  
RT: 15.13 min Scan# 2014  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDIC050


Tgt Ion:184 Resp: 189181  
Ion Ratio Lower Upper  
184 100  
63 65.8 69.2 103.8#  
154 65.5 53.3 79.9



#54  
Dibenzofuran  
Concen: 47.115 ng  
RT: 15.42 min Scan# 2062  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51

Tgt Ion:168 Resp: 2013421  
Ion Ratio Lower Upper  
168 100  
139 38.7 27.3 40.9  
169 13.2 10.6 16.0









#55  
4-Nitrophenol  
Concen: 57.505 ng  
RT: 15.22 min Scan# 2028  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDICC050


Tgt Ion:139 Resp: 309013  
Ion Ratio Lower Upper  
139 100  
109 75.3 130.0 170.0#  
65 101.1 130.0 170.0#



#56  
2,4-Dinitrotoluene  
Concen: 41.435 ng  
RT: 15.37 min Scan# 2055  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51

Tgt Ion:165 Resp: 483362  
Ion Ratio Lower Upper  
165 100  
63 45.1 52.4 78.6#  
89 71.2 72.4 108.6#  
182 2.9 2.0 3.0









#57  
 Fluorene  
 Concen: 42.895 ng  
 RT: 16.06 min Scan# 2171  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC050


Tgt Ion:166 Resp: 1612689  
 Ion Ratio Lower Upper  
 166 100  
 165 97.5 79.0 118.4  
 167 13.3 10.3 15.5



#58  
 2,3,4,6-Tetrachlorophenol  
 Concen: 45.624 ng  
 RT: 15.63 min Scan# 2099  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

Tgt Ion:232 Resp: 420347  
 Ion Ratio Lower Upper  
 232 100  
 131 46.6 0.0 0.0#  
 130 2.6 0.0 0.0#  
 166 32.6 0.0 0.0#









#59  
Diethylphthalate  
Concen: 40.613 ng  
RT: 15.81 min Scan# 2129  
Delta R.T. 0.01 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDIC050


Tgt Ion:149 Resp: 1660716  
Ion Ratio Lower Upper  
149 100  
177 21.8 18.3 27.5  
150 12.5 9.8 14.8



#60  
4-Chlorophenyl-phenylether  
Concen: 41.725 ng  
RT: 16.04 min Scan# 2168  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51

Tgt Ion:204 Resp: 815431  
Ion Ratio Lower Upper  
204 100  
206 32.8 26.6 39.8  
141 59.6 45.2 67.8









#61  
4-Nitroaniline  
Concen: 50.379 ng  
RT: 16.08 min Scan# 2175  
Delta R.T. 0.01 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDIC050


Tgt Ion: 138 Resp: 394153  
Ion Ratio Lower Upper  
138 100  
92 52.6 16.7 56.7  
108 81.1 37.6 77.6#



#62  
Azobenzene  
Concen: 37.955 ng  
RT: 16.33 min Scan# 2217  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51

Tgt Ion: 77 Resp: 1611863  
Ion Ratio Lower Upper  
77 100  
182 24.2 6.5 46.5  
105 18.9 3.8 43.8  
51 32.7 5.5 45.5





#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.74 min Scan# 2457

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDICC050


Tgt Ion:188 Resp: 843914

Ion Ratio Lower Upper

188 100

94 9.4 8.7 13.1

80 10.8 9.3 13.9



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BM0

Ion 80.00 (79.70 to 80.70): BM0

800000

600000

400000



200000

0

17.74

17.70 17.80

Time-->



#64

4,6-Dinitro-2-methylphenol

Concen: 46.977 ng

RT: 16.13 min Scan# 2184

Delta R.T. 0.01 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion:198 Resp: 249976

Ion Ratio Lower Upper

198 100

51 48.4 28.8 68.8

105 42.4 30.5 70.5



Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BM0

Ion 105.00 (104.70 to 105.70): E

250000

200000

150000

100000


50000


0

16.13

16.05 16.10 16.15 16.20

Time-->









#65  
 n-Nitrosodiphenylamine  
 Concen: 52.232 ng  
 RT: 16.25 min Scan# 2204  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC050


Tgt Ion:169 Resp: 1391556  
 Ion Ratio Lower Upper  
 169 100  
 168 67.3 53.9 80.9  
 167 36.4 28.9 43.3



#66  
 4-Bromophenyl-phenylether  
 Concen: 48.950 ng  
 RT: 16.93 min Scan# 2319  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

Tgt Ion:248 Resp: 484177  
 Ion Ratio Lower Upper  
 248 100  
 250 96.7 77.4 116.0  
 141 72.0 45.2 67.8#









#67  
Hexachlorobenzene  
Concen: 50.363 ng  
RT: 17.05 min Scan# 2340  
Delta R.T. 0.01 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDIC050

Tgt Ion:284 Resp: 546920  
Ion Ratio Lower Upper  
284 100  
142 33.0 20.4 30.6#  
249 30.4 23.8 35.6




Abundance Ion 283.80 (283.50 to 284.50): E  
Ion 142.00 (141.70 to 142.70): E  
Ion 249.00 (248.70 to 249.70): E






#68  
Atrazine  
Concen: 45.071 ng  
RT: 17.18 min Scan# 2362  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51

Tgt Ion:200 Resp: 450381  
Ion Ratio Lower Upper  
200 100  
173 26.6 4.8 44.8  
215 47.7 26.9 66.9



Abundance Ion 200.00 (199.70 to 200.70): E  
Ion 173.00 (172.70 to 173.70): E  
Ion 215.00 (214.70 to 215.70): E





#69

Pentachlorophenol

Concen: 48.645 ng

RT: 17.39 min Scan# 2397

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050


Tgt Ion:266 Resp: 316268

Ion Ratio Lower Upper

266 100

268 64.3 52.0 78.0




264 62.4 49.0 73.4



Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E



#70

Phenanthrene

Concen: 47.707 ng

RT: 17.78 min Scan# 2464

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion:178 Resp: 2428686

Ion Ratio Lower Upper

178 100

176 19.0 15.2 22.8



179 15.4 12.1 18.1




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#71  
 Anthracene  
 Concen: 50.042 ng  
 RT: 17.87 min Scan# 2479  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC050


Tgt Ion:178 Resp: 2482265  
 Ion Ratio Lower Upper  
 178 100  
 176 18.7 14.6 22.0  
 179 15.5 12.6 19.0



#72  
 Carbazole  
 Concen: 47.579 ng  
 RT: 18.13 min Scan# 2524  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

Tgt Ion:167 Resp: 2218467  
 Ion Ratio Lower Upper  
 167 100  
 166 21.8 17.2 25.8  
 139 13.3 10.8 16.2





#73

Di-n-butylphthalate

Concen: 44.950 ng

RT: 18.66 min Scan# 2613

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDICC050


Tgt Ion:149 Resp: 2682220

Ion Ratio Lower Upper

149 100

150 9.1 7.5 11.3




104 5.1 3.7 5.5



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 45.553 ng

RT: 19.75 min Scan# 2799

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion:202 Resp: 2730136

Ion Ratio Lower Upper

202 100

101 11.5 0.0 28.7



203 17.4 0.0 37.2




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.000 ng

RT: 21.83 min Scan# 3153

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDICC050


Tgt Ion:240 Resp: 880185

Ion Ratio Lower Upper

240 100

120 11.0 8.2 12.2




236 25.2 20.0 30.0



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 54.717 ng

RT: 19.92 min Scan# 2828

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion:184 Resp: 1434294

Ion Ratio Lower Upper

184 100

185 14.1 11.3 16.9



183 11.8 8.9 13.3




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#77

Pyrene

Concen: 45.356 ng

RT: 20.11 min Scan# 2860

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050


Tgt Ion:202 Resp: 2773421

Ion Ratio Lower Upper




202 100

200 21.2 16.9 25.3

203 17.4 14.1 21.1



Abundance Ion 202.00 (201.70 to 202.70): E  
Ion 200.00 (199.70 to 200.70): E  
Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 87.927 ng

RT: 20.29 min Scan# 2891

Delta R.T. 0.01 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion:244 Resp: 3960810

Ion Ratio Lower Upper



244 100


212 7.4 4.9 7.3#

122 10.5 6.2 9.4#



Abundance Ion 244.00 (243.70 to 244.70): E  
Ion 212.00 (211.70 to 212.70): E  
Ion 122.00 (121.70 to 122.70): E









#79  
Butylbenzylphthalate  
Concen: 44.332 ng  
RT: 20.96 min Scan# 3004  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDIC050


Tgt Ion:149 Resp: 1220804  
Ion Ratio Lower Upper  
149 100  
91 69.2 50.1 75.1  
206 20.9 16.2 24.2



#80  
Benzo(a)anthracene  
Concen: 48.939 ng  
RT: 21.82 min Scan# 3151  
Delta R.T. 0.00 min  
Lab File: BM015238.D  
Acq: 22 May 2018 13:51

Tgt Ion:228 Resp: 2798956  
Ion Ratio Lower Upper  
228 100  
226 27.2 21.7 32.5  
229 19.7 16.0 24.0









#81  
 3,3'-Dichlorobenzidine  
 Concen: 49.966 ng  
 RT: 21.74 min Scan# 3137  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC050


Tgt Ion:252 Resp: 1068069  
 Ion Ratio Lower Upper  
 252 100  
 254 64.4 51.9 77.9  
 126 11.6 9.8 14.8



#82  
 Chrysene  
 Concen: 47.503 ng  
 RT: 21.87 min Scan# 3160  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

Tgt Ion:228 Resp: 2635484  
 Ion Ratio Lower Upper  
 228 100  
 226 30.0 24.1 36.1  
 229 19.4 15.5 23.3









#83  
 Bis(2-ethylhexyl)phthalate  
 Concen: 44.271 ng  
 RT: 21.70 min Scan# 3131  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC050

Tgt Ion:149 Resp: 1778296  
 Ion Ratio Lower Upper  
 149 100  
 167 27.4 22.4 33.6  
 279 4.6 3.4 5.0




Abundance Ion 149.00 (148.70 to 149.70): E  
 Ion 167.00 (166.70 to 167.70): E  
 Ion 279.00 (278.70 to 279.70): E






#84  
 Di-n-octyl phthalate  
 Concen: 54.204 ng  
 RT: 22.63 min Scan# 3289  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

Tgt Ion:149 Resp: 3014166  
 Ion Ratio Lower Upper  
 149 100  
 167 1.6 1.2 1.8  
 43 11.1 7.3 10.9#



Abundance Ion 149.00 (148.70 to 149.70): E  
 Ion 167.00 (166.70 to 167.70): E  
 Ion 43.00 (42.70 to 43.70): BM0









#85  
 Indeno(1,2,3-cd)pyrene  
 Concen: 75.402 ng  
 RT: 26.80 min Scan# 3998  
 Delta R.T. 0.01 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDIC050


|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 276   | Resp: | 3288894 |
| Ion      | Ratio | Lower | Upper   |
| 276      | 100   |       |         |
| 138      | 25.1  | 12.0  | 18.0#   |
| 277      | 25.3  | 20.0  | 30.0    |



#86  
 Perylene-d12  
 Concen: 20.000 ng  
 RT: 24.27 min Scan# 3568  
 Delta R.T. 0.00 min  
 Lab File: BM015238.D  
 Acq: 22 May 2018 13:51

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 264   | Resp: | 901546 |
| Ion      | Ratio | Lower | Upper  |
| 264      | 100   |       |        |
| 260      | 23.4  | 18.6  | 27.8   |
| 265      | 21.8  | 17.3  | 25.9   |





#87

Benzo(b)fluoranthene

Concen: 46.506 ng

RT: 23.53 min Scan# 3442

Delta R.T. 0.01 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050


Tgt Ion:252 Resp: 2909241

Ion Ratio Lower Upper

252 100

253 21.9 18.0 27.0




125 10.0 9.9 14.9



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 45.995 ng

RT: 23.58 min Scan# 3450

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion:252 Resp: 2735285

Ion Ratio Lower Upper

252 100

253 22.3 17.2 25.8



125 10.1 8.1 12.1




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#89

Benzo(a)pyrene

Concen: 49.831 ng

RT: 24.17 min Scan# 3550

Delta R.T. 0.00 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050


Tgt Ion:252 Resp: 2747102

Ion Ratio Lower Upper

252 100

253 21.7 17.6 26.4




125 11.3 7.4 11.2#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 61.983 ng

RT: 26.81 min Scan# 3999

Delta R.T. 0.01 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51


Tgt Ion:278 Resp: 2792437

Ion Ratio Lower Upper

278 100

139 16.3 13.4 20.0



279 24.0 19.0 28.4




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 64.147 ng

RT: 27.58 min Scan# 4130

Delta R.T. 0.01 min

Lab File: BM015238.D

Acq: 22 May 2018 13:51

Instrument :

BNA\_M

ClientSampleId :

SSTDIC050

Tgt Ion: 276 Resp: 2701931

Ion Ratio Lower Upper

276 100

277 23.7 18.5 27.7

138 21.1 19.0 28.6

