

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052218\
 Data File : BM015254.D
 Acq On : 22 May 2018 23:40
 Operator : SJ/JU
 Sample : J2133-07
 Misc : LOD 5 PPM
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 LOD-MDL-WATER-01-QT2-2018

Quant Time: May 23 05:37:19 2018
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM052218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 22 15:47:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.40	152	191817	20.00	ng	0.00
21) Naphthalene-d8	11.24	136	763447	20.00	ng	0.00
38) Acenaphthene-d10	15.02	164	394492	20.00	ng	0.00
63) Phenanthrene-d10	17.73	188	824436	20.00	ng	0.00
75) Chrysene-d12	21.83	240	840616	20.00	ng	0.00
86) Perylene-d12	24.27	264	864604	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.89	112	1601254	136.58	ng	0.00
7) Phenol-d6	7.55	99	2023065	132.51	ng	0.00
23) Nitrobenzene-d5	9.59	82	1213107	87.05	ng	0.00
41) 2,4,6-Tribromophenol	16.49	330	701575	141.82	ng	0.00
44) 2-Fluorobiphenyl	13.65	172	2661292	83.86	ng	0.00
78) Terphenyl-d14	20.29	244	3568964	93.75	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.62	88	24946	5.150	ng	# 100
3) Pyridine	4.09	79	44669	3.600	ng	# 85
4) n-Nitrosodimethylamine	3.99	42	26674	3.973	ng	# 60
32) Benzoic acid	10.47	122	26552	3.887	ng	90
33) 4-Chloroaniline	11.40	127	70855	4.474	ng	# 92
35) Caprolactam	12.16	113	10653	3.366	ng	91
52) 3-Nitroaniline	14.92	138	25816	3.733	ng	# 79
55) 4-Nitrophenol	15.22	139	16002	2.853	ng	# 56
64) 4,6-Dinitro-2-methylphenol	16.13	198	13884	3.051	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052218\
Data File : BM015254.D
Acq On : 22 May 2018 23:40
Operator : SJ/JU
Sample : J2133-07
Misc : LOD 5 PPM
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
LOD-MDL-WATER-01-QT2-2018

Quant Time: May 23 05:37:19 2018
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM052218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 22 15:47:28 2018
Response via : Initial Calibration

