

Data Path : Z:\HPCHEM1\BNA_M\Data\BM052218\
 Data File : BM015262.D
 Acq On : 23 May 2018 04:34
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: May 23 05:57:02 2018
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM052218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 22 15:47:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	120	0.00
2	1,4-Dioxane	40.000	38.802	3.0	114	0.00
3	Pyridine	40.000	40.742	-1.9	117	0.00
4	n-Nitrosodimethylamine	40.000	42.318	-5.8	122	0.00
5 S	2-Fluorophenol	80.000	80.588	-0.7	117	0.00
6	Aniline	40.000	41.204	-3.0	121	0.00
7 S	Phenol-d6	80.000	81.971	-2.5	120	0.00
8	2-Chlorophenol	40.000	40.888	-2.2	120	0.00
9	Benzaldehyde	40.000	40.860	-2.1	123	0.00
10 C	Phenol	40.000	40.644	-1.6	119	0.00
11	bis(2-Chloroethyl)ether	40.000	41.320	-3.3	122	0.00
12	1,3-Dichlorobenzene	40.000	40.401	-1.0	120	0.00
13 C	1,4-Dichlorobenzene	40.000	40.566	-1.4	120	0.00
14	1,2-Dichlorobenzene	40.000	40.918	-2.3	122	0.00
15	Benzyl Alcohol	40.000	42.296	-5.7	123	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.763	-4.4	126	0.00
17	2-Methylphenol	40.000	41.557	-3.9	122	0.00
18	Hexachloroethane	40.000	41.901	-4.8	123	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.459	-6.1	125	0.00
20	3+4-Methylphenols	40.000	41.555	-3.9	122	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	124	0.00
22	Acetophenone	40.000	40.507	-1.3	124	0.00
23 S	Nitrobenzene-d5	80.000	82.626	-3.3	124	0.00
24	Nitrobenzene	40.000	40.794	-2.0	123	0.00
25	Isophorone	40.000	41.816	-4.5	124	0.00
26 C	2-Nitrophenol	40.000	40.476	-1.2	123	0.00
27	2,4-Dimethylphenol	40.000	41.110	-2.8	124	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.363	-3.4	126	0.00
29 C	2,4-Dichlorophenol	40.000	41.612	-4.0	124	0.00
30	1,2,4-Trichlorobenzene	40.000	40.705	-1.8	124	0.00
31	Naphthalene	40.000	40.245	-0.6	124	0.00
32	Benzoic acid	40.000	39.774	0.6	123	0.03
33	4-Chloroaniline	40.000	40.668	-1.7	122	0.00
34 C	Hexachlorobutadiene	40.000	41.344	-3.4	126	0.00
35	Caprolactam	40.000	38.849	2.9	116	0.02
36 C	4-Chloro-3-methylphenol	40.000	41.054	-2.6	123	0.00
37	2-Methylnaphthalene	40.000	40.513	-1.3	124	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	124	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	41.145	-2.9	125	0.00
40 P	Hexachlorocyclopentadiene	40.000	41.713	-4.3	132	0.00
41 S	2,4,6-Tribromophenol	80.000	83.153	-3.9	121	0.00
42 C	2,4,6-Trichlorophenol	40.000	42.253	-5.6	123	0.00
43	2,4,5-Trichlorophenol	40.000	41.493	-3.7	122	0.00
44 S	2-Fluorobiphenyl	80.000	81.834	-2.3	124	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM052218\
 Data File : BM015262.D
 Acq On : 23 May 2018 04:34
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: May 23 05:57:02 2018
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM052218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 22 15:47:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	40.684	-1.7	124	0.00
46	2-Chloronaphthalene	40.000	40.775	-1.9	123	0.00
47	2-Nitroaniline	40.000	40.650	-1.6	122	0.00
48	Acenaphthylene	40.000	41.181	-3.0	123	0.00
49	Dimethylphthalate	40.000	39.987	0.0	121	0.00
50	2,6-Dinitrotoluene	40.000	41.453	-3.6	120	0.00
51 C	Acenaphthene	40.000	40.286	-0.7	123	0.00
52	3-Nitroaniline	40.000	39.802	0.5	119	0.00
53 P	2,4-Dinitrophenol	40.000	37.648	5.9	119	0.00
54	Dibenzofuran	40.000	39.545	1.1	121	0.00
55 P	4-Nitrophenol	40.000	38.682	3.3	115	0.00
56	2,4-Dinitrotoluene	40.000	39.380	1.5	118	0.00
57	Fluorene	40.000	39.813	0.5	120	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	40.727	-1.8	120	0.00
59	Diethylphthalate	40.000	40.319	-0.8	121	0.00
60	4-Chlorophenyl-phenylether	40.000	40.138	-0.3	123	0.00
61	4-Nitroaniline	40.000	38.701	3.2	116	0.00
62	Azobenzene	40.000	42.145	-5.4	123	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	118	0.00
64	4,6-Dinitro-2-methylphenol	40.000	40.510	-1.3	117	0.00
65 c	n-Nitrosodiphenylamine	40.000	41.935	-4.8	120	0.00
66	4-Bromophenyl-phenylether	40.000	42.637	-6.6	122	0.00
67	Hexachlorobenzene	40.000	41.294	-3.2	120	0.00
68	Atrazine	40.000	42.346	-5.9	118	0.00
69 C	Pentachlorophenol	40.000	41.329	-3.3	119	0.00
70	Phenanthrene	40.000	40.389	-1.0	118	0.00
71	Anthracene	40.000	41.073	-2.7	117	0.00
72	Carbazole	40.000	40.470	-1.2	115	0.00
73	Di-n-butylphthalate	40.000	42.475	-6.2	117	0.00
74 C	Fluoranthene	40.000	39.459	1.4	112	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	106	0.00
76	Benzidine	40.000	44.732	-11.8	108	0.00
77	Pyrene	40.000	42.665	-6.7	112	0.00
78 S	Terphenyl-d14	80.000	85.400	-6.8	112	0.00
79	Butylbenzylphthalate	40.000	44.172	-10.4	109	0.00
80	Benzo(a)anthracene	40.000	40.893	-2.2	106	0.00
81	3,3'-Dichlorobenzidine	40.000	41.573	-3.9	104	0.00
82	Chrysene	40.000	40.522	-1.3	105	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	44.292	-10.7	109	0.00
84 c	Di-n-octyl phthalate	40.000	41.639	-4.1	107	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	40.166	-0.4	103	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Benzo(b)fluoranthene	40.000	39.638	0.9	100	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM052218\
Data File : BM015262.D
Acq On : 23 May 2018 04:34
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: May 23 05:57:02 2018
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM052218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 22 15:47:28 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	42.027	-5.1	107	0.00
89 C	Benzo(a)pyrene	40.000	41.213	-3.0	104	0.00
90	Dibenzo(a,h)anthracene	40.000	40.808	-2.0	103	0.00
91	Benzo(g,h,i)perylene	40.000	40.601	-1.5	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0