

Data Path : Z:\HPCHEM1\BNA_M\Data\BM052218\
 Data File : BM015262.D
 Acq On : 23 May 2018 04:34
 Operator : SJ/JU
 Sample : SSTDC0040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDC0040

Quant Time: May 23 05:57:02 2018
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM052218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 22 15:47:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2	1,4-Dioxane	0.505	0.490	3.0	114	0.00
3	Pyridine	1.294	1.318	-1.9	117	0.00
4	n-Nitrosodimethylamine	0.700	0.741	-5.9	122	0.00
5 S	2-Fluorophenol	1.222	1.231	-0.7	117	0.00
6	Aniline	1.961	2.020	-3.0	121	0.00
7 S	Phenol-d6	1.592	1.631	-2.4	120	0.00
8	2-Chlorophenol	1.380	1.410	-2.2	120	0.00
9	Benzaldehyde	1.007	1.029	-2.2	123	0.00
10 C	Phenol	1.617	1.643	-1.6	119	0.00
11	bis(2-Chloroethyl)ether	1.282	1.325	-3.4	122	0.00
12	1,3-Dichlorobenzene	1.573	1.589	-1.0	120	0.00
13 C	1,4-Dichlorobenzene	1.590	1.613	-1.4	120	0.00
14	1,2-Dichlorobenzene	1.510	1.544	-2.3	122	0.00
15	Benzyl Alcohol	1.151	1.217	-5.7	123	0.00
16	2,2'-oxybis(1-Chloropropane	2.023	2.112	-4.4	126	0.00
17	2-Methylphenol	1.075	1.116	-3.8	122	0.00
18	Hexachloroethane	0.565	0.592	-4.8	123	0.00
19 P	n-Nitroso-di-n-propylamine	1.015	1.077	-6.1	125	0.00
20	3+4-Methylphenols	1.427	1.482	-3.9	122	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	124	0.00
22	Acetophenone	0.480	0.486	-1.3	124	0.00
23 S	Nitrobenzene-d5	0.365	0.377	-3.3	124	0.00
24	Nitrobenzene	0.380	0.388	-2.1	123	0.00
25	Isophorone	0.627	0.656	-4.6	124	0.00
26 C	2-Nitrophenol	0.171	0.173	-1.2	123	0.00
27	2,4-Dimethylphenol	0.309	0.318	-2.9	124	0.00
28	bis(2-Chloroethoxy)methane	0.411	0.425	-3.4	126	0.00
29 C	2,4-Dichlorophenol	0.293	0.305	-4.1	124	0.00
30	1,2,4-Trichlorobenzene	0.345	0.351	-1.7	124	0.00
31	Naphthalene	1.040	1.046	-0.6	124	0.00
32	Benzoic acid	0.179	0.178	0.6	123	0.03
33	4-Chloroaniline	0.415	0.422	-1.7	122	0.00
34 C	Hexachlorobutadiene	0.205	0.212	-3.4	126	0.00
35	Caprolactam	0.083	0.081	2.4	116	0.02
36 C	4-Chloro-3-methylphenol	0.299	0.307	-2.7	123	0.00
37	2-Methylnaphthalene	0.720	0.730	-1.4	124	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	124	0.00
39	1,2,4,5-Tetrachlorobenzene	0.704	0.725	-3.0	125	0.00
40 P	Hexachlorocyclopentadiene	0.407	0.449	-10.3	132	0.00
41 S	2,4,6-Tribromophenol	0.251	0.261	-4.0	121	0.00
42 C	2,4,6-Trichlorophenol	0.423	0.447	-5.7	123	0.00
43	2,4,5-Trichlorophenol	0.457	0.474	-3.7	122	0.00
44 S	2-Fluorobiphenyl	1.609	1.646	-2.3	124	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM052218\
 Data File : BM015262.D
 Acq On : 23 May 2018 04:34
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: May 23 05:57:02 2018
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM052218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 22 15:47:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.750	1.780	-1.7	124	0.00
46	2-Chloronaphthalene	1.347	1.373	-1.9	123	0.00
47	2-Nitroaniline	0.384	0.390	-1.6	122	0.00
48	Acenaphthylene	2.000	2.060	-3.0	123	0.00
49	Dimethylphthalate	1.596	1.596	0.0	121	0.00
50	2,6-Dinitrotoluene	0.333	0.345	-3.6	120	0.00
51 C	Acenaphthene	1.246	1.255	-0.7	123	0.00
52	3-Nitroaniline	0.351	0.349	0.6	119	0.00
53 P	2,4-Dinitrophenol	0.170	0.168	1.2	119	0.00
54	Dibenzofuran	1.965	1.942	1.2	121	0.00
55 P	4-Nitrophenol	0.284	0.275	3.2	115	0.00
56	2,4-Dinitrotoluene	0.456	0.449	1.5	118	0.00
57	Fluorene	1.553	1.545	0.5	120	0.00
58	2,3,4,6-Tetrachlorophenol	0.392	0.399	-1.8	120	0.00
59	Diethylphthalate	1.573	1.586	-0.8	121	0.00
60	4-Chlorophenyl-phenylether	0.791	0.794	-0.4	123	0.00
61	4-Nitroaniline	0.367	0.355	3.3	116	0.00
62	Azobenzene	1.495	1.575	-5.4	123	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
64	4,6-Dinitro-2-methylphenol	0.110	0.112	-1.8	117	0.00
65 c	n-Nitrosodiphenylamine	0.643	0.674	-4.8	120	0.00
66	4-Bromophenyl-phenylether	0.223	0.238	-6.7	122	0.00
67	Hexachlorobenzene	0.257	0.266	-3.5	120	0.00
68	Atrazine	0.204	0.216	-5.9	118	0.00
69 C	Pentachlorophenol	0.143	0.148	-3.5	119	0.00
70	Phenanthrene	1.139	1.150	-1.0	118	0.00
71	Anthracene	1.136	1.166	-2.6	117	0.00
72	Carbazole	1.005	1.017	-1.2	115	0.00
73	Di-n-butylphthalate	1.184	1.257	-6.2	117	0.00
74 C	Fluoranthene	1.249	1.232	1.4	112	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
76	Benzidine	0.626	0.700	-11.8	108	0.00
77	Pyrene	1.262	1.346	-6.7	112	0.00
78 S	Terphenyl-d14	0.906	0.967	-6.7	112	0.00
79	Butylbenzylphthalate	0.515	0.569	-10.5	109	0.00
80	Benzo(a)anthracene	1.260	1.288	-2.2	106	0.00
81	3,3'-Dichlorobenzidine	0.467	0.485	-3.9	104	0.00
82	Chrysene	1.187	1.203	-1.3	105	0.00
83	Bis(2-ethylhexyl)phthalate	0.750	0.830	-10.7	109	0.00
84 c	Di-n-octyl phthalate	1.311	1.365	-4.1	107	0.00
85	Indeno(1,2,3-cd)pyrene	1.436	1.442	-0.4	103	0.00
86 I	Perylene-d12	1.000	1.000	0.0	104	0.00
87	Benzo(b)fluoranthene	1.265	1.254	0.9	100	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM052218\
Data File : BM015262.D
Acq On : 23 May 2018 04:34
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: May 23 05:57:02 2018
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM052218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 22 15:47:28 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.197	1.258	-5.1	107	0.00
89 C	Benzo(a)pyrene	1.180	1.216	-3.1	104	0.00
90	Dibenzo(a,h)anthracene	1.204	1.229	-2.1	103	0.00
91	Benzo(g,h,i)perylene	1.173	1.190	-1.4	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0