


Data Path : Z:\HPCHEM1\BNA\_M\Data\BM052218\  
 Data File : BM015261.D  
 Acq On : 23 May 2018 03:57  
 Operator : SJ/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_M\METHODS\8270-BM052218.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Tue May 22 15:47:28 2018



AutoFind: Scans 2521, 2522, 2523; Background Corrected with Scan 2515

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 45.0      | 101112  | PASS             |
| 68          | 69           | 0.00         | 2            | 2.0       | 1959    | PASS             |
| 69          | 198          | 0.00         | 100          | 43.8      | 98480   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.5       | 523     | PASS             |
| 127         | 198          | 10           | 80           | 50.8      | 114045  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.6       | 1358    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 224640  | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 15200   | PASS             |
| 275         | 198          | 10           | 60           | 28.8      | 64776   | PASS             |
| 365         | 198          | 1            | 100          | 4.1       | 9174    | PASS             |
| 441         | 443          | 0.01         | 100          | 81.4      | 27397   | PASS             |
| 442         | 198          | 50           | 100          | 77.5      | 173992  | PASS             |
| 443         | 442          | 15           | 24           | 19.4      | 33672   | PASS             |