




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM052224\
Data File : BM045737.D
Acq On : 22 May 2024 10:02
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 23 04:19:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM052224.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 23 04:12:56 2024
Response via : Initial Calibration






TIC: BM045737.D\data.ms

(70) Pentachlorophenol (C)**16.621min (+ 0.000) 0.00 ng****response 542727**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.70	64.58
264.00	63.60	61.67
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM052224\
Data File : BM045737.D
Acq On : 22 May 2024 10:02
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 23 04:19:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM052224.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 23 04:12:56 2024
Response via : Initial Calibration



TIC: BM045737.D\data.ms

(77) Benzidine**19.227min (+ 0.000) 0.00 ng****response 3070049**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.90	13.86
183.00	11.30	11.41
0.00	0.00	0.00