

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052317\
 Data File : BM010135.D
 Acq On : 23 May 2017 11:07
 Operator : SJ/MA
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Quant Time: May 23 13:08:10 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM051917.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 20 04:25:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.96	152	150200	20.00	ng	-0.02
21) Naphthalene-d8	10.77	136	531746	20.00	ng	-0.02
38) Acenaphthene-d10	14.59	164	359572	20.00	ng	-0.02
63) Phenanthrene-d10	17.34	188	908800	20.00	ng	-0.01
75) Chrysene-d12	21.50	240	1338775	20.00	ng	-0.01
86) Perylene-d12	23.86	264	1478966	20.00	ng	-0.03

System Monitoring Compounds

5) 2-Fluorophenol	5.54	112	645163	77.57	ng	-0.02
7) Phenol-d6	7.15	99	860722	80.42	ng	-0.02
23) Nitrobenzene-d5	9.15	82	932619	94.62	ng	-0.02
41) 2,4,6-Tribromophenol	16.09	330	471494	86.34	ng	-0.01
44) 2-Fluorobiphenyl	13.22	172	2363617	84.51	ng	-0.02
78) Terphenyl-d14	19.94	244	4883261	82.94	ng	-0.01

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.43	88	155255	40.63	ng	# 100
3) Pyridine	3.85	79	420747	41.13	ng	91
4) n-Nitrosodimethylamine	3.77	42	177837	48.06	ng	# 83
6) Aniline	7.31	93	528178	39.83	ng	94
8) 2-Chlorophenol	7.53	128	354229	39.14	ng	95
9) Benzaldehyde	7.12	77	276746	41.54	ng	88
10) Phenol	7.17	94	455652	40.40	ng	89
11) bis(2-Chloroethyl)ether	7.39	93	328151	39.03	ng	98
12) 1,3-Dichlorobenzene	7.84	146	452296	39.33	ng	# 88
13) 1,4-Dichlorobenzene	7.99	146	465600	39.41	ng	94
14) 1,2-Dichlorobenzene	8.31	146	446557	39.38	ng	94
15) Benzyl Alcohol	8.22	79	324791	40.11	ng	# 89
16) 2,2'-oxybis(1-Chloropropan	8.47	45	290233	35.11	ng	81
17) 2-Methylphenol	8.42	107	318485	39.68	ng	# 85
18) Hexachloroethane	9.03	117	158808	41.47	ng	# 74
19) n-Nitroso-di-n-propylamine	8.77	70	325663	43.11	ng	92
20) 3+4-Methylphenols	8.75	107	434310	40.07	ng	87
22) Acetophenone	8.80	105	579470	42.51	ng	# 98
24) Nitrobenzene	9.19	77	477504	47.03	ng	93
25) Isophorone	9.70	82	738440	43.36	ng	# 94
26) 2-Nitrophenol	9.89	139	185412	43.32	ng	# 65
27) 2,4-Dimethylphenol	9.95	122	319934	41.40	ng	# 85
28) bis(2-Chloroethoxy)methane	10.18	93	446178	41.28	ng	99
29) 2,4-Dichlorophenol	10.42	162	394248	44.31	ng	99
30) 1,2,4-Trichlorobenzene	10.62	180	493086	42.83	ng	98
31) Naphthalene	10.82	128	1140153	40.08	ng	100
32) Benzoic acid	10.09	122	222259	42.20	ng	# 81
33) 4-Chloroaniline	10.96	127	446600	40.75	ng	# 84
34) Hexachlorobutadiene	11.06	225	351240	45.74	ng	99
35) Caprolactam	11.76	113	105839	41.46	ng	# 84
36) 4-Chloro-3-methylphenol	12.07	107	416178	43.19	ng	85
37) 2-Methylnaphthalene	12.42	142	862958	42.14	ng	# 93
39) 1,2,4,5-Tetrachlorobenzene	12.77	216	587748	42.51	ng	# 100
40) Hexachlorocyclopentadiene	12.73	237	347409	43.51	ng	99


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052317\
 Data File : BM010135.D
 Acq On : 23 May 2017 11:07
 Operator : SJ/MA
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Quant Time: May 23 13:08:10 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-BM051917.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 20 04:25:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.03	196	360083	44.21	ng	97
43) 2,4,5-Trichlorophenol	13.10	196	399818	44.43	ng #	92
45) 1,1'-Biphenyl	13.43	154	1221434	41.31	ng	99
46) 2-Chloronaphthalene	13.47	162	992952	41.48	ng	96
47) 2-Nitroaniline	13.71	65	279129	45.34	ng #	81
48) Acenaphthylene	14.32	152	1464983	41.38	ng	99
49) Dimethylphthalate	14.06	163	1332803	41.55	ng #	99
50) 2,6-Dinitrotoluene	14.19	165	258711	42.48	ng	89
51) Acenaphthene	14.66	154	904742	41.11	ng	97
52) 3-Nitroaniline	14.54	138	228795	40.42	ng #	66
53) 2,4-Dinitrophenol	14.72	184	153507	41.27	ng #	86
54) Dibenzofuran	14.99	168	1429093	41.30	ng	96
55) 4-Nitrophenol	14.85	139	181022	39.74	ng #	28
56) 2,4-Dinitrotoluene	14.97	165	366862	42.58	ng #	84
57) Fluorene	15.64	166	1260553	41.89	ng	100
58) 2,3,4,6-Tetrachlorophenol	15.22	232	341628	44.57	ng #	100
59) Diethylphthalate	15.40	149	1254203	41.36	ng	96
60) 4-Chlorophenyl-phenylether	15.63	204	712124	42.15	ng	99
61) 4-Nitroaniline	15.70	138	222898	38.79	ng #	24
62) Azobenzene	15.92	77	1055284	47.33	ng	90
64) 4,6-Dinitro-2-methylphenol	15.72	198	239525	43.14	ng	80
65) n-Nitrosodiphenylamine	15.85	169	1100780	40.41	ng	100
66) 4-Bromophenyl-phenylether	16.52	248	478901	41.16	ng	95
67) Hexachlorobenzene	16.62	284	547477	38.73	ng	96
68) Atrazine	16.79	200	422830	40.94	ng	99
69) Pentachlorophenol	16.98	266	324660	41.82	ng	98
70) Phenanthrene	17.38	178	2054401	40.40	ng	99
71) Anthracene	17.47	178	2088791	41.39	ng	99
72) Carbazole	17.76	167	1806973	40.45	ng	98
73) Di-n-butylphthalate	18.27	149	2134365	40.29	ng #	98
74) Fluoranthene	19.38	202	2697160	40.85	ng	98
76) Benzidine	19.59	184	1168088	47.30	ng	99
77) Pyrene	19.74	202	2842134	39.60	ng	100
79) Butylbenzylphthalate	20.62	149	1032637	38.86	ng #	86
80) Benzo(a)anthracene	21.48	228	3020705	40.94	ng	99
81) 3,3'-Dichlorobenzidine	21.42	252	1208504	40.72	ng #	96
82) Chrysene	21.53	228	2857635	40.84	ng	100
83) Bis(2-ethylhexyl)phthalate	21.36	149	1557909	39.21	ng #	98
84) Di-n-octyl phthalate	22.27	149	2851637	40.09	ng #	97
85) Indeno(1,2,3-cd)pyrene	26.28	276	4375759	42.64	ng #	100
87) Benzo(b)fluoranthene	23.14	252	3604022	41.64	ng #	92
88) Benzo(k)fluoranthene	23.19	252	3350438	39.51	ng #	95
89) Benzo(a)pyrene	23.76	252	3344858	40.55	ng #	97
90) Dibenzo(a,h)anthracene	26.29	278	3821907	41.44	ng #	92
91) Benzo(g,h,i)perylene	27.04	276	3748561	41.48	ng #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 7.96 min Scan# 828

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:152 Resp: 150200

Ion Ratio Lower Upper

152 100

150 162.2 119.5 179.3




115 63.2 43.0 64.4



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 40.63 ng

RT: 3.43 min Scan# 59

Delta R.T. -0.01 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion: 88 Resp: 155255

Ion Ratio Lower Upper

88 100

58 61.8 0.0 0.0#



43 27.4 0.0 0.0#




Abundance Ion 88.00 (87.70 to 88.70): BM0

Ion 58.00 (57.70 to 58.70): BM0

Ion 43.00 (42.70 to 43.70): BM0





#3

Pyridine

Concen: 41.13 ng

RT: 3.85 min Scan# 130

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion: 79 Resp: 420747

Ion Ratio Lower Upper

79 100

52 53.4 51.1 76.7




51 32.7 26.1 39.1



Abundance Ion 79.00 (78.70 to 79.70): BM010135.D (-83) (-)

Ion 52.00 (51.70 to 52.70): BM010135.D (-83) (-)

Ion 51.00 (50.70 to 51.70): BM010135.D (-83) (-)



#4

n-Nitrosodimethylamine

Concen: 48.06 ng

RT: 3.77 min Scan# 116

Delta R.T. -0.01 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion: 42 Resp: 177837

Ion Ratio Lower Upper

42 100

74 127.2 119.3 178.9



44 5.0 5.4 8.2#




Abundance Ion 42.00 (41.70 to 42.70): BM010135.D (-101) (-)

Ion 74.00 (73.70 to 74.70): BM010135.D (-101) (-)

Ion 44.00 (43.70 to 44.70): BM010135.D (-101) (-)





#5

2-Fluorophenol

Concen: 77.57 ng

RT: 5.54 min Scan# 417

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion: 112 Resp: 645163

Ion Ratio Lower Upper

112 100

64 59.4 38.6 57.8#




63 36.2 19.3 28.9#



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): B

Ion 63.00 (62.70 to 63.70): B



#6

Aniline

Concen: 39.83 ng

RT: 7.31 min Scan# 718

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion: 93 Resp: 528178

Ion Ratio Lower Upper

93 100

66 43.0 30.8 46.2



65 21.1 16.0 24.0






Abundance Ion 93.00 (92.70 to 93.70): B

Ion 66.00 (65.70 to 66.70): B

Ion 65.00 (64.70 to 65.70): B





#7

Phenol-d6

Concen: 80.42 ng

RT: 7.15 min Scan# 690

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040





Tgt Ion: 99 Resp: 860722

Ion Ratio Lower Upper

99 100

42 19.7 11.0 16.4#

71 45.9 23.4 35.2#



#8

2-Chlorophenol

Concen: 39.14 ng

RT: 7.53 min Scan# 756

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion: 128 Resp: 354229


Ion Ratio Lower Upper

128 100

130 30.8 12.0 52.0

64 49.0 24.9 64.9





#9

Benzaldehyde

Concen: 41.54 ng

RT: 7.12 min Scan# 686

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040





Tgt Ion: 77 Resp: 276746

Ion Ratio Lower Upper

77 100

105 85.6 78.8 118.8

106 82.9 73.7 113.7



#10

Phenol

Concen: 40.40 ng

RT: 7.17 min Scan# 695

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07




Tgt Ion: 94 Resp: 455652


Ion Ratio Lower Upper

94 100

65 32.2 18.8 58.8

66 51.3 39.9 79.9









#11
 bis(2-Chloroethyl)ether
 Concen: 39.03 ng
 RT: 7.39 min Scan# 732
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
93	100		
63	69.9	49.5	89.5
95	31.3	13.5	53.5



#12
 1,3-Dichlorobenzene
 Concen: 39.33 ng
 RT: 7.84 min Scan# 808
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07

Tgt Ion	Ratio	Lower	Upper
146	100		
148	67.5	50.9	76.3
75	40.8	21.4	32.2









#13
 1,4-Dichlorobenzene
 Concen: 39.41 ng
 RT: 7.99 min Scan# 834
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion:146 Resp: 465600
 Ion Ratio Lower Upper
 146 100
 148 67.4 51.9 77.9
 111 42.9 29.2 43.8



#14
 1,2-Dichlorobenzene
 Concen: 39.38 ng
 RT: 8.31 min Scan# 888
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07

Tgt Ion:146 Resp: 446557
 Ion Ratio Lower Upper
 146 100
 148 67.9 52.3 78.5
 111 44.2 30.6 45.8









#15
Benzyl Alcohol
Concen: 40.11 ng
RT: 8.22 min Scan# 873
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 324791
Ion Ratio Lower Upper
79 100
108 63.5 65.0 97.4#
77 66.8 53.0 79.6



#16
2,2'-oxybis(1-Chloropropane)
Concen: 35.11 ng
RT: 8.47 min Scan# 916
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07

Tgt Ion: 45 Resp: 290233
Ion Ratio Lower Upper
45 100
77 23.6 0.0 35.2
79 18.8 0.0 32.0









#17
2-Methylphenol
Concen: 39.68 ng
RT: 8.42 min Scan# 906
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:	107	Resp:	318485
Ion	Ratio	Lower	Upper
107	100		
108	106.7	89.8	134.8
77	59.4	32.6	48.8#
79	57.3	32.8	49.2#



#18
Hexachloroethane
Concen: 41.47 ng
RT: 9.03 min Scan# 1011
Delta R.T. -0.01 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07

Tgt Ion:	117	Resp:	158808
Ion	Ratio	Lower	Upper
117	100		
119	94.9	79.2	118.8
201	133.3	69.8	104.6#





#19

n-Nitroso-di-n-propylamine

Concen: 43.11 ng

RT: 8.77 min Scan# 966

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion: 70 Resp: 325663





Ion Ratio Lower Upper

70 100

42 47.3 44.5 66.7

101 9.9 7.4 11.2

130 18.2 15.3 22.9



#20

3+4-Methylphenols

Concen: 40.07 ng

RT: 8.75 min Scan# 963

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Tgt Ion: 107 Resp: 434310


Ion Ratio Lower Upper


107 100

108 87.1 73.2 113.2

77 45.3 10.7 50.7

79 39.0 9.6 49.6









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.77 min Scan# 1306
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:	136	Resp:	531746
Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.7	13.1
54	7.1	4.6	6.8#
68	4.5	3.7	5.5



#22
Acetophenone
Concen: 42.51 ng
RT: 8.80 min Scan# 971
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07

Tgt Ion:	105	Resp:	579470
Ion	Ratio	Lower	Upper
105	100		
71	0.2	0.6	1.0#
51	28.8	21.6	32.4
120	20.3	16.1	24.1






#23
 Nitrobenzene-d5
 Concen: 94.62 ng
 RT: 9.15 min Scan# 1030
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040




Tgt Ion	Ratio	Lower	Upper
82	100		
128	35.7	32.8	49.2
54	49.2	40.6	60.8



Abundance Ion 82.00 (81.70 to 82.70): BM010069.D (-1023) (-)


Ion 128.00 (127.70 to 128.70): BM010069.D (-1023) (-)

Ion 54.00 (53.70 to 54.70): BM010069.D (-1023) (-)



#24
 Nitrobenzene
 Concen: 47.03 ng
 RT: 9.19 min Scan# 1037
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07



Tgt Ion	Ratio	Lower	Upper
77	100		
123	35.5	32.6	49.0
65	14.3	10.9	16.3




Abundance Ion 77.00 (76.70 to 77.70): BM010069.D (-1026) (-)

Ion 123.00 (122.70 to 123.70): BM010069.D (-1026) (-)

Ion 65.00 (64.70 to 65.70): BM010069.D (-1026) (-)









#25
Isophorone
Concen: 43.36 ng
RT: 9.70 min Scan# 1124
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 738440
Ion Ratio Lower Upper
82 100
95 9.2 5.8 8.6#
138 17.6 16.3 24.5



#26
2-Nitrophenol
Concen: 43.32 ng
RT: 9.89 min Scan# 1157
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07

Tgt Ion: 139 Resp: 185412
Ion Ratio Lower Upper
139 100
109 43.2 20.1 30.1#
65 60.1 31.5 47.3#









#27
2,4-Dimethylphenol
Concen: 41.40 ng
RT: 9.95 min Scan# 1166
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
122	100		
107	129.7	84.7	127.1#
121	59.1	46.6	69.8



#28
bis(2-Chloroethoxy)methane
Concen: 41.28 ng
RT: 10.18 min Scan# 1206
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07

Tgt Ion	Ratio	Lower	Upper
93	100		
95	32.6	25.5	38.3
123	11.2	8.6	13.0









#29
 2,4-Dichlorophenol
 Concen: 44.31 ng
 RT: 10.42 min Scan# 1246
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
162	100		
164	63.4	42.8	82.8
98	33.5	14.5	54.5



#30
 1,2,4-Trichlorobenzene
 Concen: 42.83 ng
 RT: 10.62 min Scan# 1280
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07

Tgt Ion	Ratio	Lower	Upper
180	100		
182	95.5	77.6	116.4
145	28.1	23.4	35.2









#31
Naphthalene
Concen: 40.08 ng
RT: 10.82 min Scan# 1314
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:128 Resp: 1140153
Ion Ratio Lower Upper
128 100
129 11.5 8.9 13.3
127 13.0 10.4 15.6



#32
Benzoic acid
Concen: 42.20 ng
RT: 10.09 min Scan# 1190
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07

Tgt Ion:122 Resp: 222259
Ion Ratio Lower Upper
122 100
105 136.7 99.9 139.9
77 115.8 73.7 113.7#









#33
4-Chloroaniline
Concen: 40.75 ng
RT: 10.96 min Scan# 1338
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
127	100		
129	31.7	25.3	37.9
65	39.2	17.6	26.4
92	25.1	13.1	19.7



#34
Hexachlorobutadiene
Concen: 45.74 ng
RT: 11.06 min Scan# 1356
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07

Tgt Ion	Ratio	Lower	Upper
225	100		
223	59.2	47.9	71.9
227	62.1	50.1	75.1





#35

Caprolactam

Concen: 41.46 ng

RT: 11.76 min Scan# 1475

Delta R.T. -0.05 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:113 Resp: 105839

Ion Ratio Lower Upper

113 100

55 134.6 145.9 185.9#




56 113.4 101.0 141.0



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BM0

Ion 56.00 (55.70 to 56.70): BM0



#36

4-Chloro-3-methylphenol

Concen: 43.19 ng

RT: 12.07 min Scan# 1527

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion:107 Resp: 416178

Ion Ratio Lower Upper

107 100

144 24.0 23.3 34.9



142 74.5 72.6 109.0




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E





#37

2-Methylnaphthalene

Concen: 42.14 ng

RT: 12.42 min Scan# 1586

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040





Tgt Ion:142 Resp: 862958

Ion Ratio Lower Upper

142 100

141 87.6 71.9 107.9

115 43.5 25.4 38.0#



#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 14.59 min Scan# 1956

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07




Tgt Ion:164 Resp: 359572


Ion Ratio Lower Upper

164 100

162 101.4 82.4 123.6

160 47.1 35.8 53.6









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 42.51 ng
 RT: 12.77 min Scan# 1646
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07




Instrument :
 BNA_M
 ClientSampleId :
 SSTCCCC040


Tgt Ion:	216	Resp:	587748
Ion	Ratio	Lower	Upper
216	100		
214	79.1	0.0	0.0#
179	21.2	0.0	0.0#
108	17.5	0.0	0.0#



#40
 Hexachlorocyclopentadiene
 Concen: 43.51 ng
 RT: 12.73 min Scan# 1640
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07

Tgt Ion:	237	Resp:	347409
Ion	Ratio	Lower	Upper
237	100		
235	61.4	42.0	82.0
272	12.7	0.0	33.4





#41

2,4,6-Tribromophenol

Concen: 86.34 ng

RT: 16.09 min Scan# 2210

Delta R.T. -0.01 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:330 Resp: 471494

Ion Ratio Lower Upper

330 100

332 96.5 0.0 0.0#




141 33.1 0.0 0.0#



Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#42

2,4,6-Trichlorophenol

Concen: 44.21 ng

RT: 13.03 min Scan# 1690

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion:196 Resp: 360083

Ion Ratio Lower Upper

196 100

198 99.0 76.3 114.5



200 32.7 25.6 38.4




Abundance Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E









#43
2,4,5-Trichlorophenol
Concen: 44.43 ng
RT: 13.10 min Scan# 1703
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion	Resp	Lower	Upper
196	399818		
198	94.2	79.4	119.0
97	45.0	26.8	40.2
132	20.5	18.6	28.0



#44
2-Fluorobiphenyl
Concen: 84.51 ng
RT: 13.22 min Scan# 1722
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07

Tgt Ion	Resp	Lower	Upper
172	2363617		
171	35.3	28.5	42.7
170	22.7	18.8	28.2









#45
 1,1'-Biphenyl
 Concen: 41.31 ng
 RT: 13.43 min Scan# 1758
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion:154 Resp: 1221434
 Ion Ratio Lower Upper
 154 100
 153 39.8 20.7 60.7
 76 10.5 0.0 30.9



#46
 2-Chloronaphthalene
 Concen: 41.48 ng
 RT: 13.47 min Scan# 1766
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07

Tgt Ion:162 Resp: 992952
 Ion Ratio Lower Upper
 162 100
 127 37.2 26.9 40.3
 164 32.4 26.8 40.2









#47
2-Nitroaniline
Concen: 45.34 ng
RT: 13.71 min Scan# 1806
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 65 Resp: 279129
Ion Ratio Lower Upper
65 100
92 64.3 59.1 88.7
138 92.1 93.9 140.9#



#48
Acenaphthylene
Concen: 41.38 ng
RT: 14.32 min Scan# 1910
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07

Tgt Ion: 152 Resp: 1464983
Ion Ratio Lower Upper
152 100
151 20.3 15.9 23.9
153 13.4 10.6 16.0





#49

Dimethylphthalate

Concen: 41.55 ng

RT: 14.06 min Scan# 1865

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:163 Resp: 1332803

Ion Ratio Lower Upper

163 100

194 4.8 2.7 4.1#




164 10.1 8.2 12.2



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 42.48 ng

RT: 14.19 min Scan# 1888

Delta R.T. -0.01 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion:165 Resp: 258711

Ion Ratio Lower Upper

165 100

63 63.9 44.1 66.1



89 62.2 44.3 66.5




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BM0

Ion 89.00 (88.70 to 89.70): BM0





#51

Acenaphthene

Concen: 41.11 ng

RT: 14.66 min Scan# 1967

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:154 Resp: 904742

Ion Ratio Lower Upper

154 100

153 116.7 90.0 135.0




152 54.2 42.6 63.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 40.42 ng

RT: 14.54 min Scan# 1947

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion:138 Resp: 228795

Ion Ratio Lower Upper

138 100

108 13.1 7.3 10.9#



92 131.1 76.6 114.8#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BM









#53
2,4-Dinitrophenol
Concen: 41.27 ng
RT: 14.72 min Scan# 1978
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:184 Resp: 153507
Ion Ratio Lower Upper
184 100
63 66.3 69.2 103.8#
154 64.4 53.3 79.9



#54
Dibenzofuran
Concen: 41.30 ng
RT: 14.99 min Scan# 2024
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07

Tgt Ion:168 Resp: 1429093
Ion Ratio Lower Upper
168 100
139 37.1 27.3 40.9
169 13.5 10.6 16.0









#55
4-Nitrophenol
Concen: 39.74 ng
RT: 14.85 min Scan# 1999
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:139 Resp: 181022
Ion Ratio Lower Upper
139 100
109 118.0 37.7 77.7#
65 122.9 49.9 89.9#



#56
2,4-Dinitrotoluene
Concen: 42.58 ng
RT: 14.97 min Scan# 2020
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07

Tgt Ion:165 Resp: 366862
Ion Ratio Lower Upper
165 100
63 49.9 52.4 78.6#
89 77.3 72.4 108.6
182 3.3 2.0 3.0#





#57

Fluorene

Concen: 41.89 ng

RT: 15.64 min Scan# 2134

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:166 Resp: 1260553

Ion Ratio Lower Upper

166 100

165 99.1 79.0 118.4




167 13.3 10.3 15.5



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#58

2,3,4,6-Tetrachlorophenol

Concen: 44.57 ng

RT: 15.22 min Scan# 2062

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Tgt Ion:232 Resp: 341628


Ion Ratio Lower Upper

232 100

131 36.4 0.0 0.0#

130 2.3 0.0 0.0#

166 28.3 0.0 0.0#






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E





#59

Diethylphthalate

Concen: 41.36 ng

RT: 15.40 min Scan# 2094

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 1254203

Ion Ratio Lower Upper

149 100

177 25.3 18.3 27.5




150 13.3 9.8 14.8



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#60

4-Chlorophenyl-phenylether

Concen: 42.15 ng

RT: 15.63 min Scan# 2132

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion:204 Resp: 712124

Ion Ratio Lower Upper

204 100

206 32.4 26.6 39.8



141 56.5 45.2 67.8




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E









#61
4-Nitroaniline
Concen: 38.79 ng
RT: 15.70 min Scan# 2145
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 138 Resp: 222898
Ion Ratio Lower Upper
138 100
92 63.3 16.7 56.7#
108 129.0 37.6 77.6#



#62
Azobenzene
Concen: 47.33 ng
RT: 15.92 min Scan# 2182
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07

Tgt Ion: 77 Resp: 1055284
Ion Ratio Lower Upper
77 100
182 27.6 6.5 46.5
105 18.1 3.8 43.8
51 33.4 5.5 45.5









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.34 min Scan# 2423
Delta R.T. -0.01 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
188	100		
94	5.3	8.7	13.1#
80	6.5	9.3	13.9#



#64
4,6-Dinitro-2-methylphenol
Concen: 43.14 ng
RT: 15.72 min Scan# 2148
Delta R.T. -0.02 min
Lab File: BM010135.D
Acq: 23 May 2017 11:07

Tgt Ion	Ratio	Lower	Upper
198	100		
51	37.7	28.8	68.8
105	34.3	30.5	70.5









#65
 n-Nitrosodiphenylamine
 Concen: 40.41 ng
 RT: 15.85 min Scan# 2170
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
169	100		
168	67.5	53.9	80.9
167	35.7	28.9	43.3



#66
 4-Bromophenyl-phenylether
 Concen: 41.16 ng
 RT: 16.52 min Scan# 2284
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07

Tgt Ion	Ratio	Lower	Upper
248	100		
250	98.6	77.4	116.0
141	64.4	45.2	67.8





#67

Hexachlorobenzene

Concen: 38.73 ng

RT: 16.62 min Scan# 2300

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:284 Resp: 547477

Ion Ratio Lower Upper

284 100

142 28.3 20.4 30.6




249 31.4 23.8 35.6



Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#68

Atrazine

Concen: 40.94 ng

RT: 16.79 min Scan# 2330

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion:200 Resp: 422830

Ion Ratio Lower Upper

200 100

173 24.4 4.8 44.8



215 47.4 26.9 66.9




Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E





#69

Pentachlorophenol

Concen: 41.82 ng

RT: 16.98 min Scan# 2362

Delta R.T. -0.01 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:266 Resp: 324660

Ion Ratio Lower Upper

266 100

268 65.1 52.0 78.0




264 63.7 49.0 73.4



Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E



#70

Phenanthrene

Concen: 40.40 ng

RT: 17.38 min Scan# 2430

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion:178 Resp: 2054401

Ion Ratio Lower Upper

178 100

176 18.9 15.2 22.8



179 14.7 12.1 18.1




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 41.39 ng
 RT: 17.47 min Scan# 2445
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion:178 Resp: 2088791
 Ion Ratio Lower Upper
 178 100
 176 18.9 14.6 22.0
 179 15.0 12.6 19.0



#72
 Carbazole
 Concen: 40.45 ng
 RT: 17.76 min Scan# 2494
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07

Tgt Ion:167 Resp: 1806973
 Ion Ratio Lower Upper
 167 100
 166 21.8 17.2 25.8
 139 12.1 10.8 16.2





#73

Di-n-butylphthalate

Concen: 40.29 ng

RT: 18.27 min Scan# 2582

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 2134365

Ion Ratio Lower Upper




149 100

150 9.5 7.5 11.3

104 6.4 3.7 5.5#



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 40.85 ng

RT: 19.38 min Scan# 2770

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion:202 Resp: 2697160

Ion Ratio Lower Upper



202 100


101 6.5 0.0 28.7

203 17.1 0.0 37.2



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 21.50 min Scan# 3130

Delta R.T. -0.01 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:240 Resp: 1338775

Ion Ratio Lower Upper

240 100

120 5.5 8.2 12.2#




236 25.4 20.0 30.0



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 47.30 ng

RT: 19.59 min Scan# 2805

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion:184 Resp: 1168088

Ion Ratio Lower Upper

184 100

185 14.1 11.3 16.9



183 12.1 8.9 13.3




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#77

Pyrene

Concen: 39.60 ng

RT: 19.74 min Scan# 2832

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:202 Resp: 2842134

Ion Ratio Lower Upper

202 100

200 21.3 16.9 25.3




203 17.4 14.1 21.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 82.94 ng

RT: 19.94 min Scan# 2865

Delta R.T. -0.01 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion:244 Resp: 4883261

Ion Ratio Lower Upper

244 100

212 7.0 4.9 7.3



122 6.1 6.2 9.4#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#79

Butylbenzylphthalate

Concen: 38.86 ng

RT: 20.62 min Scan# 2981

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 1032637

Ion Ratio Lower Upper

149 100

91 74.1 50.1 75.1




206 26.1 16.2 24.2#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BM0

Ion 206.00 (205.70 to 206.70): E



#80

Benzo(a)anthracene

Concen: 40.94 ng

RT: 21.48 min Scan# 3127

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion:228 Resp: 3020705

Ion Ratio Lower Upper

228 100

226 26.5 21.7 32.5



229 19.5 16.0 24.0




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#81
 3,3'-Dichlorobenzidine
 Concen: 40.72 ng
 RT: 21.42 min Scan# 3117
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion:252 Resp: 1208504
 Ion Ratio Lower Upper
 252 100
 254 63.3 51.9 77.9
 126 7.6 9.8 14.8#



#82
 Chrysene
 Concen: 40.84 ng
 RT: 21.53 min Scan# 3136
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07

Tgt Ion:228 Resp: 2857635
 Ion Ratio Lower Upper
 228 100
 226 30.2 24.1 36.1
 229 19.5 15.5 23.3









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 39.21 ng
 RT: 21.36 min Scan# 3107
 Delta R.T. -0.02 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion:149 Resp: 1557909
 Ion Ratio Lower Upper
 149 100
 167 29.0 22.4 33.6
 279 6.1 3.4 5.0#



#84
 Di-n-octyl phthalate
 Concen: 40.09 ng
 RT: 22.27 min Scan# 3261
 Delta R.T. -0.03 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07

Tgt Ion:149 Resp: 2851637
 Ion Ratio Lower Upper
 149 100
 167 1.9 1.2 1.8#
 43 8.0 7.3 10.9





#85

Indeno(1,2,3-cd)pyrene

Concen: 42.64 ng

RT: 26.28 min Scan# 3943

Delta R.T. -0.05 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:276 Resp: 4375759

Ion Ratio Lower Upper

276 100

138 13.8 0.0 0.0#




277 25.3 0.0 0.0#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#86

Perylene-d12

Concen: 20.00 ng

RT: 23.86 min Scan# 3531

Delta R.T. -0.03 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion:264 Resp: 1478966

Ion Ratio Lower Upper

264 100

260 24.2 18.6 27.8



265 22.6 17.3 25.9




Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E





#87

Benzo(b)fluoranthene

Concen: 41.64 ng

RT: 23.14 min Scan# 3409

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:252 Resp: 3604022

Ion Ratio Lower Upper




252 100

253 21.3 18.0 27.0

125 5.1 9.9 14.9#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 39.51 ng

RT: 23.19 min Scan# 3417

Delta R.T. -0.02 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion:252 Resp: 3350438

Ion Ratio Lower Upper



252 100


253 22.1 17.2 25.8

125 5.2 8.1 12.1#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#89

Benzo(a)pyrene

Concen: 40.55 ng

RT: 23.76 min Scan# 3514

Delta R.T. -0.03 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040


Tgt Ion:252 Resp: 3344858

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4




125 5.8 7.4 11.2#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 41.44 ng

RT: 26.29 min Scan# 3945

Delta R.T. -0.05 min

Lab File: BM010135.D

Acq: 23 May 2017 11:07


Tgt Ion:278 Resp: 3821907

Ion Ratio Lower Upper

278 100

139 9.0 13.4 20.0#



279 24.2 19.0 28.4




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91
 Benzo(g,h,i)perylene
 Concen: 41.48 ng
 RT: 27.04 min Scan# 4072
 Delta R.T. -0.05 min
 Lab File: BM010135.D
 Acq: 23 May 2017 11:07

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Tgt Ion: 276 Resp: 3748561
 Ion Ratio Lower Upper
 276 100
 277 23.7 18.5 27.7
 138 11.6 19.0 28.6#

