

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM052523\
 Data File : BM040044.D
 Acq On : 26 May 2023 19:38
 Operator : CG/JU
 Sample : 02851-08
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 H0FK1

Quant Time: May 27 00:12:11 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM052523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 25 15:01:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.012	152	8322	0.400	ng/ul	0.00
4) Naphthalene-d8	10.827	136	29131	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.646	164	14353	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.391	188	24193	0.400	ng/ul	0.00
17) Chrysene-d12	21.567	240	10174	0.400	ng/ul	0.00
23) Perylene-d12	24.014	264	10169	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.380	96	38758	4.000	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.406	152	9205	0.258	ng/ul	0.00
18) Fluoranthene-d10	19.412	212	12573	0.410	ng/ul	0.00
Target Compounds						
					Qvalue	
15) Phenanthrene	17.429	178	1870	0.027	ng/ul#	87
19) Fluoranthene	19.440	202	3388	0.074	ng/ul#	90
20) Pyrene	19.802	202	3471	0.075	ng/ul#	87
21) Benzo(a)anthracene	21.550	228	972	0.028	ng/ul#	82
22) Chrysene	21.602	228	1364	0.037	ng/ul#	87
24) Benzo(b)fluoranthene	23.265	252	1900	0.047	ng/ul#	1
27) Indeno(1,2,3-cd)pyrene	26.561	276	1020	0.021	ng/ul#	81
29) Benzo(g,h,i)perylene	27.339	276	2485	0.062	ng/ul#	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM052523\
Data File : BM040044.D
Acq On : 26 May 2023 19:38
Operator : CG/JU
Sample : 02851-08
Misc :
ALS Vial : 50 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
H0FK1

Quant Time: May 27 00:12:11 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM052523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 25 15:01:16 2023
Response via : Initial Calibration

