

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052815\
 Data File : BM001431.D
 Acq On : 28 May 2015 15:35
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM052815

Quant Time: May 28 16:12:30 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM052815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 28 15:41:25 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2 S	2-Fluorophenol	0.768	0.744	3.1	91	0.00
3 S	Phenol-d6	0.806	0.764	5.2	92	-0.03
4	bis(2-Chloroethyl)ether	0.843	0.810	3.9	90	0.00
5 P	n-Nitroso-di-n-propylamine	0.547	0.515	5.9	93	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
7 S	Nitrobenzene-d5	0.268	0.256	4.5	92	0.00
8	bis(2-Chloroethoxy)methane	0.264	0.265	-0.4	91	0.00
9 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
10 S	2,4,6-Tribromophenol	0.154	0.145	5.8	94	0.00
11 S	2-Fluorobiphenyl	1.388	1.371	1.2	94	0.00
12	2,6-Dinitrotoluene	0.216	0.191	11.6	93	0.00
13	2,4-Dinitrotoluene	0.223	0.217	2.7	92	0.00
14	4-Chlorophenyl-phenylether	0.767	0.746	2.7	91	0.00
15 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
16	4-Bromophenyl-phenylether	0.197	0.192	2.5	96	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
18 S	Terphenyl-d14	0.622	0.607	2.4	97	0.00
19	Benzo(a)anthracene	1.250	1.156	7.5	95	0.00
20	3,3'-Dichlorobenzidine	0.299	0.274	8.4	95	0.00
21 I	Perylene-d12	1.000	1.000	0.0	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0